



Saleae, Inc.
408 N Canal St STE A
South San Francisco, CA 94080
Ph: 510-984-2463 Fax: 415-684-7727
support@saleae.com
www.saleae.com

Statement of Calibration for the Saleae Logic, Logic16, Logic 4, Logic 8, Logic Pro 8 and Logic Pro 16

There is no calibration certificate for any of the Saleae products listed below because periodic calibration of these devices is not required.

Devices:

- Saleae Logic USB Logic Analyzer
- Saleae Logic16 USB Logic Analyzer
- Saleae Logic 4 USB Logic Analyzer
- Saleae Logic 8 USB Logic Analyzer
- Saleae Logic Pro 8 USB Logic Analyzer
- Saleae Logic Pro 16 USB Logic Analyzer

The Saleae devices which support analog, the Saleae Logic 4, Logic 8, Logic Pro 8, and Logic Pro 16 all undergo a onetime factory calibration. These calibration files are automatically distributed through the internet. The status of the device calibration can be found by following the instructions here:

<https://support.saleae.com/hc/en-us/articles/208666836>

The one-time factory calibration is not NIST traceable. No additional calibration is necessary and it is not possible to re-calibrate the device.

Kind regards,
Mark

--

Mark Garrison
Co-founder and Chief Technical Officer
Saleae Inc
+1 (510) 984-2463