Certificate

Report Number:	EA_Cert_20260924_144641_36000315	**Date of Validation:**	24.09.2026 14:46
Job Number:	000	**Date of Next Validation Due:**	24.09.2027
Company:	-	**Validating Technician:**	End Assembly
Address:	-	**Contact:**	-
	-	**Email Address:**	-
	-	**Telephone:**	-

Machine Configuration

Model:	Raycon EX2A	**Aperture Size:**	300/150
Serial:	36000315	**Reject Type:**	Pusher / nozzle
Device Name:	Raycon 1	**X-Ray Voltage:**	50 kV
Line Name:		**X-Ray Current:**	2.0 mA

Product Details

Name:	Reference Product
Length:	160 mm
Width:	160 mm
Height:	10 mm
Temperature:	-
Belt Speed:	0.4 m/s
Pack Rate:	-

Validation Results

Test Sample Information	Test Sample Identification	Test Sample Density	Testcard	Probability of Detection	Confirmed Rejections	Result
Stainless Steel,0.5 mm	77194596	8.0 g/cm^3	5x (same)	100 %	-	Passed

For any questions or to schedule your next validation please contact Sesotec at +4985543080 or at webcontact@sesotec.com

Note: The test is performed with the metal sample at the leading edge, middle, and trailing edge of the product (if applicable). The validation was performed with certified test bodies / test cards. Foreign body detection results can be impacted by the product signals, the results listed above are based on the specific product that was used during this Validation process.