

# REDLINER: AN ACTIVITY MONITOR FOR MANUAL WHEELCHAIR USERS

Kenton Hamaluik, MSc, EIT<sup>1</sup>, Zohreh Salimi, MSc<sup>1</sup>, Martin Ferguson-Pell, PhD, C.Phys. 1,2

<sup>1</sup>Faculty of Rehabilitation Medicine, <sup>2</sup>Peter Lougheed Leadership College

## Background

- Wheelchair users have a strong need to monitor activity
- Manual wheelchair users are highly susceptible to repetitive stress injuries [2]
- Propulsion forces over 80% of maximum capacity result in damage [1]
- Typical existing activity monitors don't work with manual wheelchair users
- SmartWheels are the gold standard, but are a very expensive clinical tool
- SmartWheels require modifying a user's chair (replacing the wheels)
- There is a need for an affordable consumer-grade tool for activity monitoring

## Objective

The objective of this work was to create an inexpensive activity monitor for manual wheelchair users capable of measuring the following data without modifying the user's chair:


- Number of propulsion strokes
- Average travel velocity
- Amount of time spent active
- Estimated distance travelled
- Number of "redline events"<sup>1</sup>
- <sup>1</sup> redline events are instances of when the user's propulsion force exceeds 80% of the maximal propulsion force they can generate

#### Method


 Redliner uses a novel non-intrusive sensor which attaches to the side of a user's wheel

#### Results


As you can see in Figure 1, derp.



**Figure 1:** Velocity traces for both SmartWheel and Redliner for 10 pushes on rough gravel. Despite the rough terrain, the traces are in close agreement—allowing Redliner to reasonably estimate the velocity and distance travelled by the wheelchair.



**Figure 2:** Acceleration traces for both SmartWheel and Redliner for 10 pushes on rough gravel. The traces are in close agreement, allowing Redliner to accurately detect overexertion events.



**Figure 3:** Acceleration, detected pushes, and redlines of 10 pushes on gravel as measured by Redliner. An algorithm was developed to detect pushes from this acceleration data while also monitoring redlining events.

## Conclusions

- Redliner is a new activity monitor for manual wheelchair users
- Redliner has been validated against expensive SmartWheel devices
- Redliner is moving forward as a commercial entity to produce and sell the devices

### References

[1] I.P. Freely. A small paper. *The journal of small papers*, -1, 1997. to appear. [2] Hugh Jass. A big paper. *The journal of big papers*, MCMXCVII, 7991.

## Acknowledgements

Funding for this work was provided for by the University of Alberta, Telus, the Government of Canada, and Redliner Inc.