

1. Install drivers and operating software from supplied StellarNet USB-drive or website

=>for 64bit Windows 7/ 8 / Vista => see SWDriver64 readme.txt

=>for 32bit Windows 7 / Vista / other versions:

- a) Attach USB cable to spectrometer. b) Attach UP5V to spectrometer (if older unit)
- c) Attach USB cable to available USB port on PC (or USB hub) => **Device wizard appears**

WinXP => allow to search for a suitable driver on StellarNet USB-drive

Win 7 /other => specify path to USB-drive \SWDrivers12\ **spectro-type \ OS-version**

spectro-type = \SWDrivers-USB2-Spectrometers or \SWDrivers-USB1-GREEN-Wave

OS-version = \WinXP-Vista-7 or \Win9x or \Win7-64bit <64bit also works this way

Use Device Manager to verify “StellarNet Spectrometer” is listed under USBDEV device.
Windows 7 users may see “unknown device” appear and will need to click “Update driver” then specify proper driver location as shown above. To start Device Manager, Right click on MyComputer then select Manage, then select Device Manager on list.

If Device Manager says “spectrometer - start” you need to select “Update driver”.

If you have the time, it is smart to repeat the cable installation for each USB port on your PC however **Windows 7 users do NOT need to do this.**

2. Install latest version of SpectraWiz using SWUpdate.exe from StellarNet website.

If web not available, then click SWUpdate-Install.exe in StellarNet Software folder.

Click on SpectraWiz desktop icon to start, then verify continuous spectral graph display updates appear with spectrometer attached, then exit SpectraWiz.

Win2000 only: after installation completes, goto SpectraWiz folder, click on **Fix2000.bat**.

For operation as a SpectroRadiometer goto step 3.

If you see BW-16 listed on spectro bottom label: the x-axis coefficients load automatically.

Otherwise use “Setup -> Unit Calibration” menu. See reverse side for detailed instructions.

3. Install intensity calibration files for SpectroRadiometer operation

The calibration files are installed by clicking CD:\MyCal-nnnnn.exe

where “nnnnn” = spectrometer serial # shown on label.

This self extracting zip file moves all “.cal” files (if you ordered more than one) to

C:\Program Files\StellarNet\SpectraWiz\SwCals

Start SpectraWiz and verify continuous spectral graph display updates appear.

Get going FAST -->>> WATCH the SpectraWiz software training videos on the USB-drive!

Driver and software updates are easily downloaded from the StellarNet website

a. All drivers are in SWDrivers12 password=”wdrivers”

b. SWUpdate (all Window versions) password= no password ←get the latest version now

Additional information is available in the Spectrometer Manual on StellarNet USB-drive!

For StellarNet technical support, Phone: 813-855-8687 or email: Support@StellarNet.us

IMPORTANT >> Please turn over this page and follow steps for SpectraWiz software configuration !!

Software Configuration

>>Information found on unit label

All spectrometers automatically load parameters from unit memory when SpectraWiz software is started from your windows computer.

“Older spectrometers” with parallel port connector plus USB2EPP cable (silver DB25 connector at other end of USB2 cable) require parameters to be entered the first time a unit is attached. The same procedure can be used to verify each spectrometers wavelength coefficients that define the units x-axis:

Note: the following steps are not required if you received a new spectrometer!

However, if you received a “new” **GREEN-Wave16** spectrometer see 3a below!

1. Open SpectraWiz and select the “**Setup**” menu → then select “**Unit Calibration Coefficients.**”
2. Enter 1 at the channel prompt and then enter C1, C2, C3, and C4 values which are listed on the bottom label of the spectrometer. If no C4 is listed on the label then enter 0.
3. Select: “**Setup**” → “**Interface and Port Detector.**”
 - a) Always check USB2EPP cable unless using the GREEN-Wave (12bit spectrometer **which is now discontinued**). However, if installing the new “**GREEN-Wave16**”, check USB2EPP + LT16 + BW >> it is 16-bit!
 - b) Select the digitizer type listed on bottom of the spectrometer. These are LT12/LT14/LT16. If the digitizer is not listed, do not select any.
 - c) Select detector type; default is CCD 2048 unless otherwise stated on label.
4. Now, exit SpectraWiz for the changes to take effect. Start SpectraWiz and verify continuous spectral graph display updates appear.

For further instruction -->>> WATCH the SpectraWiz Application videos !!

Special note for Windows 7 users: (my-cname below is your computer's name)

c:\users\my-cname\AppData\local\VirtualStore\Program Files\StellarNet\SpectraWiz

This path is where Windows 7 will store all SpectraWiz application log files and other data that you may save along with any folders created, if you don't select a location outside the SpectraWiz folder (which is in the Program Files path). So when you go to look for your data and don't find it, relax! It's right in this location :)

IMPORTANT >> Please turn over this page and follow steps for SpectraWiz software configuration !!