

Maintenance sheet version 10 for ablation area (tripod, one level)

Station name	QAS_U_SF
People present	SEB JS Siggil
Purpose	Maintenance / installation / take-down / quick visit / emergency
Date	27 Aug 2024 + 1 Sept

Notes: Upgrade to new windmill power design

diode already installed, so new
v 1.1 SF existing, Batt Ok

AWS to do list		Check after maintenance			
Part	Shipped (Y/N)	Replaced? (Y/N)	Number shipped/arrival	Number departure	
Radiometer	N				Clean? Y / N
Ablation hose assembly	N				No air in hose and bladder half full? Y / N
Iridium modem IMEI number	Reuse from old			13790	
Inclinometer/compass	N		Attached to the radiometer? Y / N		
Satellite antenna	N				
Wind sensor	N	N	Black box pointing toward the radiometer (i.e., along the boom)? Y / N		
Temperature / humidity assembly	N	/	Fan spins? Y / N Sounds, OK? Y / N		
HygroClip T/hum assembly	N	/	NB: Rotronics (Y / N) could have a 40 °C offset. New or old type? NB: Lufft sensor (Y / N)		
Sonic ranger on AWS	Y		Old sensor with new membrane? Y / N		
Sonic ranger on stake(s)	N		Old sensor with new membrane? Y / N NB: make sure not being buried next winter		
Thermistor string And maybe an ablobender	N				
Solar panel	Reuse ✓ from old		Clean? Y N NB: is not effective when partially covered		
GPS antenna	Inside loggerbox		At top of logger enclosure? Y / N		
Logger box ID	T1SF-0002		CR1000 or CR1000X? ID number: 10672		
Multiplexer	N		Clicking during measurement cycle? Y / N		
Logger enclosure	Y		Replaced (including everything inside)? Y / N New desiccant bags? Y / N		
Battery box	Y	Y	New box? Y N New desiccant bags? Y / N		
Stakes	N	N	New one(s) drilled in? Y / N Length of new stake(s) 8 m? Y / N Bottom plugged to reduce melt-in? Y / N		
Tripod	Reuse ✓ from old	N	New tripod? Y N New sensor boom? Y / N		

Station name	QAS_U_SF
Date	

Metadata maintenance

	Arrival	Departure
Time difference between logger clock and UTC ($\pm$ mm:ss)		
Download data and/or change CF/SD Card (wait for green light)	Y / N	Y / N
Name of logger program		
Screen dump of values in fast scan mode	Y / N	Y / N
Photos of tripod, sensors, logger box wiring etc. (more is better)	Y / N	Y / N
Mast tilt towards radiometer (- if tilting towards anemometer)	°	°
Mast tilt across boom looking from radiometer (- if clockwise)	°	°
Boom direction relative to north	° true/magn.	° true/magn.
Box on wind sensor exactly towards mast? If not, measure.	Y / N	Y / N
Radiometer exactly aligned with inclinometer? If not, measure.	Y / N	Y / N
Temperature / humidity height (bottom of casing – surface)	cm	cm
Stake assembly standing?	Y / N / sort of	Y / N / sort of
Free length of stake (if 3: outer, holding boom with sonic ranger)	cm	cm
Free length of stake 2 (if 3: middle)	cm	cm
Free length of stake 3 (if 3: outer, without sonic ranger)	cm	cm
Height of ablation bladder fluid level above the ice surface	cm	cm
Vertical difference in surface height (at station – at hose) (approx.)	cm	cm
Length of thermistor string on surface from surface marking	cm	cm
Apparent damage to sensors? If yes, which?	Y / N	Y / N

Notes:

Must-do at next visit:

2 x PTA recoverable
 from QAS U
 solar U-bolts cold-welded!
 angle grinder solution?
 lots of other recoverables
 @ site, bent aluminum
 pipes