

Station name	SCO_L
Date	12/8-21

Metadata maintenance

	Arrival	Departure
Time difference between logger clock and UTC (± mm:ss)		- 00:06
Download data and/or change CF/SD Card (wait for green light)	Y / N	Y / N
Name of logger program		V3REWired-1-5
Screen dump of values in fast scan mode	Y / N	(Y) / N
Photos of tripod, sensors, logger box wiring etc. (more is better)	(Y) / N	(Y) / N
Mast tilt towards radiometer (- if tilting towards anemometer)	- 5 °	0 °
Mast tilt across boom looking from radiometer (- if clockwise)	10 °	5 °
Boom direction relative to north	5 ° true/magn.	° true/magn.
Box on wind sensor exactly towards mast? If not, measure.	(Y) / N	(Y) / N
Radiometer exactly aligned with inclinometer? If not, measure.	(Y) / N	(Y) / N
Temperature / humidity height (bottom of casing – surface)	266 cm	270 cm
Stake assembly standing?	Y / N / sort of	Y / N / sort of
Free length of stake (if 3: outer, holding boom with sonic ranger)	650 cm	53 cm
Free length of stake 2 (if 3: middle)	- cm	- cm
Free length of stake 3 (if 3: outer, without sonic ranger)	- cm	- cm
Height of ablation bladder fluid level above the ice surface	170 cm	175 cm
Vertical difference in surface height (at station – at hose) (approx.)	30 cm	30 cm
Length of thermistor string on surface from surface marking	800 cm	- 30 cm
Apparent damage to sensors? If yes, which?	Y / (N)	Y / (N)

Notes:

Ablation hose melted out
 Problems with T on CNR
 DATA TRANSFER COMPLETED IN CNP HANZAK

Must-do at next visit:

Maintenance sheet version 10 for ablation area (tripod, one level)

Station name	SCO_L
People present	
Purpose	Maintenance / installation / take-down / quick visit / emergency
Date	

Notes: Upgrade from v2 to v3_rewired, reuse some items from v2

AWS to do list		Check after maintenance			
Part	Shipped (Y/N)	Replaced? (Y/N)	Number shipped/arrival	Number departure	
Radiometer	212155 /212893	Y			Clean? (Y) / N
Ablation hose assembly	Y (50274)	Y			No air in hose and bladder half full? Y / N
Iridium modem IMEI number	Inside loggerbox				
Inclinometer/compass	Inside CNR4	Y			Attached to the radiometer? (Y) / N
Satellite antenna	N				
Wind sensor	Y	Y			Black box pointing toward the radiometer (i.e., along the boom)? (Y) / N
Temperature / humidity assembly	Y	Y			Fan spins? (Y) / N Sounds, OK? Y / N
HygroClip T/hum assembly	N				NB: Rotronics (Y / N) could have a 40 °C offset. New or old type? NB: Luft sensor (Y / N)
Sonic ranger on AWS	Y	Y			Old sensor with new membrane? Y / (N)
Sonic ranger on stake(s)	Y	Y			Old sensor with new membrane? Y / (N) NB: make sure not being buried next winter
Thermistor string	Y (digital)	Y			
Solar panel	N	N			Clean? (Y) / N NB: is not effective when partially covered
GPS antenna	N	N			At top of logger enclosure? Y / N
Logger box ID	T3B-15/T3B-17	T35-15			CR1000 or CR1000X? ID number:
Multiplexer	N				Clicking during measurement cycle? (Y) / N
Logger enclosure	Y	Y			Replaced (including everything inside)? (Y) / N New desiccant bags? (Y) / N
Battery box	Y	Y			New box? (Y) / N New desiccant bags? (Y) / N
Stakes	Y	Y			New one(s) drilled in? (Y) / N Length of new stake(s) 8 m? Y / N Bottom muffled to reduce melt-in? Y / (N)
Tripod	N	N			New tripod? Y / N New sensor boom? Y / (N)