

Maintenance sheet version 9 for ablation area (tripod, one level)

Station name	THULE L
People present	CLAN & WIC
Purpose	Maintenance / installation / take-down / quick visit / emergency
Date	31 MAY 2023

Notes:

AWS to do list		Check after maintenance			
Part	Shipped (Y/N)	Replaced? (Y/N)	Number shipped/arrival	Number departure	
Radiometer	N	N			Clean? Y / N
Ablation hose assembly	N	N			No air in hose and bladder half full? Y / N
Iridium modem IMEI number	N	N	5153530		
Inclinometer/compass	N	N	Attached to the radiometer? Y / N		
Satellite antenna	N	N			
Wind sensor	N	N	Black box pointing toward the radiometer (i.e., along the boom)? Y / N		
Temperature / humidity assembly	N	N	Fan spins? Y / N Sounds, OK? Y / N		
HygroClip T/hum assembly	Y (round)	N (NOT SHIPPED)	NB: Rotronics (Y / N) could have a 40 °C offset. New or old type? NB: Lufft sensor (Y / N)		
Sonic ranger on AWS	Y	N (CAN'T LOWER STATION)	Old sensor with new membrane? Y / N		
Sonic ranger on stake(s)	Y	Y	Old sensor with new membrane? Y (N) NB: make sure not being buried next winter NEW!		
Thermistor string	N	N			
Solar panel	N	N	Clean? Y / N NB: is not effective when partially covered		
GPS antenna	N	N	At top of logger enclosure? Y / N		
Logger box ID	N	N	CR1000 or CR1000X? ID number:		
Multiplexer	N	N	Clicking during measurement cycle? Y / N		
Logger enclosure	N	N	Replaced (including everything inside)? Y / N New desiccant bags? Y / N		
Battery box	N	N	New box? Y / N New desiccant bags? Y / N		
Stakes	Y	Y	New one(s) drilled in? Y (N) Length of new stake(s) 8 m? Y (N) Bottom plugged to reduce melt-in? Y (N) ~7m DRILLED		
Tripod	N	N	New tripod? Y / N New sensor boom? Y / N		

Metadata maintenance

	Arrival	Departure
Time difference between logger clock and UTC ($\pm$ mm:ss)		
Download data and/or change CF/SD Card (wait for green light)	Y / N	Y / N
Name of logger program		
Screen dump of values in fast scan mode	Y / N	Y / N
Photos of tripod, sensors, logger box wiring etc. (more is better)	Y / N	Y / N
Mast tilt towards radiometer (- if tilting towards anemometer)	°	°
Mast tilt across boom looking from radiometer (- if clockwise)	°	°
Boom direction relative to north	° true/magn.	° true/magn.
Box on wind sensor exactly towards mast? If not, measure.	Y / N	Y / N
Radiometer exactly aligned with inclinometer? If not, measure.	Y / N	Y / N
Temperature / humidity height (bottom of casing – surface)	cm	cm
Stake assembly standing?	Y / N / sort of	Y / N / sort of
Free length of stake (if 3: outer, holding boom with sonic ranger)	cm	cm
Free length of stake 2 (if 3: middle)	cm	cm
Free length of stake 3 (if 3: outer, without sonic ranger)	cm	cm
Height of ablation bladder fluid level above the ice surface	cm	cm
Vertical difference in surface height (at station – at hose) (approx.)	cm	cm
Length of thermistor string on surface from surface marking	cm	cm
Apparent damage to sensors? If yes, which?	Y / N	Y / N

Notes:

- NO TIME OFFSET
- SCREEN DUMP YES
- CARBON FIBER STAKE VERTICAL CRACK
- SNOW DEPTH 0 – 25 cm AROUND STATION
- TRIPOD FROZEN IN, CANNOT TIP OVER.

Must-do at next visit: