

ONLY HARD COPY

AWS maintenance / establishment checklist version 5

Station name	ZAK - MAIN AWS
Observer(s)	MOLA + MCT
Purpose of visit	MAINTENANCE

Establishment of new AWS: Go to **Metadata** after maintenance / establishment

Metadata upon arrival


Date & time of arrival at station	15/4/2012 12:10
Time difference between logger clock and UTC	1s +1-
Adjust logger time to UTC	Y/N
Download data to PC and/or change CF Card (wait for green light)	Y/N
Download logger program to PC	Y/N
Note or photograph current values in fast scan mode	Y/N
Latitude (dd mm.mmm)	74 37.4873
Longitude (dd mm.mmm)	21 22.53352
Altitude (m, from GPS)	653
Photos of tripod, sensors, logger box wiring etc.	Y/N
Mast inclination in CNR1 direction (+ if CNR1 tilting down)	-7 °
Mast incl. across boom looking from CNR1 (- if clockwise)	-6 °
Boom direction relative to true/magnetic north	° true/magn
Young box direction along boom? If not, measure direction	Y/N
CNR1 aligned with mast and inclinometer? If not, measure	Y/N
Young height (center of propeller - surface)	185 cm
Rotronic height (bottom of casing - surface)	141 cm
SR50 height on AWS (membrane - surface)	142 cm
SR50 height on stakes (membrane - surface)	N/A cm
Free length of stake - outer, holding boom with SR50	— cm
Free length of stake - middle	— cm
Free length of stake - outer, holding boom without SR50	— cm
Length of ablation sensor tubing on the surface (to T-piece)	FULL HOSE cm
Or: depth of ablation sensor using markings on hose	— cm
Length of thermistor string on surface from surface marking	N/A cm

UTL AWS BEHIND

SEE BELOW

MELTED C.F. 2011

Remarks:



Maintenance step-by-step

Change Internal battery? Data and program will be lost!	Y / (N)
Change vent filter in logger enclosure?	Y / (N)
Change wires or aluminum of tripod?	Y / (N)
Change sonic rangers membranes or replace by new sensors	Y / (N) / New
Align CNR1 and inclinometer with mast within 0.5°	Y / (N)
Clean domes and check fastenings of CNR1	Y / N
Change Hygroclip in Rotronic	Y / (N)
Rotronic fan OK?	(Y) / N
Apparent damage to Young wind sensor?	Y / (N)
Clean solar panel	Y / (N)
Voltage solar panel OK?	(Y) / N
Refill ablation hose?	Y / (N)
Redrill stakes	(Y) / N
Refill Ablation drinking bag (With _____ L, 100/50%)	Y / (N)
Redrill old ablation hose?	Y / (N) / new
Redrill old thermistor string?	Y / N / new
Change desiccant bags in logger box	Y / N
Change desiccant bags in battery box	Y / N
Upload new logger program? (Name: _____)	Y / N

Whole
new
rotro-
nics
installed


New
install -
already
filled on
CPN

Remarks:

Metadata after maintenance / establishment

Latitude (dd mm.mmm)	SAME
Longitude (dd mm.mmm)	SAME
Photos of tripod, sensors, logger box wiring etc.	Y / N
Mast inclination in CNR1 direction (+ if CNR1 tilting down)	SAME °
Mast incl. across boom looking from CNR1 (- if clockwise)	SAME °
Boom direction relative to true/magnetic north	SAME true / magn
Young box direction along boom? If not, measure direction	(Y) / N
CNR1 aligned with mast and inclinometer? If not, measure	(Y) / N
Young height (center of propeller – surface)	SAME cm
Rotronic height (bottom of casing – surface)	SAME cm
SR50 height on AWS (membrane – surface)	SAME cm
SR50 height on stakes (membrane – surface)	48 cm
Free length of stake – outer, holding boom with SR50	166 164 cm
Free length of stake – middle	166 cm
Free length of stake – outer, holding boom without SR50	123 cm
Number of 2-m pieces used for new stakes	1 / 2 / 3 / (4) / 5
Length of ablation sensor tubing on the surface (to T-piece)	12.10 m cm
Or: depth of ablation sensor using markings on hose	cm
Length of thermistor string on surface from surface marking	0 cm
Approximate height of surface irregularities	10 cm
Time of departure	

Remarks:



Serial numbers of hardware

SR50 Snow (sonic ranger on AWS)		
	Replaced with:	1811
SR50 Abl (sonic ranger on stakes)		
3503	Replaced with:	— (membrane changed)
Rotronic assembly (Assembly no. if new assembly)		
Hygroclip S3 no.	replaced with:	60091615
MP100H no.	replaced with:	60079555
RS12T (Fan) no.	replaced with:	55210003
MKRS (Arm) no.	replaced with:	?
NT1400 pressure transducer (Retrieved / Left in ice)		
???	Replaced with:	5003 ABS
Thermistor string (Retrieved / Left in ice)		
	Replaced with:	R11
CNR1 radiation sensor		
	Replaced with:	
Young wind sensor		
82158	Replaced with:	105090
Inclinometer		
SN 0750	Replaced with:	none
Solar Panel		
210-702026139250	Replaced with:	—
Iridium modem (if present), IMEI number:		
300034012437350	Replaced with:	—
DC-DC Converter (SYN-DC-936)		
	Replaced with:	
Satellite antenna		
	Replaced with:	
GPS antenna		
	Replaced with:	
In loggerbox:		
Logger Cr1000/CR10X		
	Replaced with:	
Card reader CFM100		
	Replaced with:	
Multiplex AM16/32		
	Replaced with:	
Barometer		
	Replaced with:	

Remarks:

OLD NATURALLY MP. TEMP. SHIELD EXCHANGED w/ NEW.
HYGROCLIP 51382185. SWAPPED w/ NEW: 60687675