MONTECH CALIBRATION SERVICES
measure with confidence

MONTECH Calibration Services (Pty) Ltd
Postnet Suite 266, Private Bag X21, Bryanston, 2021 | 47 Fabriek Street, Strydom Park, 2169
director Seola Mashamaite | vat no 4450269099 | reg no 2014/182132/07

tel 011 464 5071 | email info@moncal.co.za | www.moncal.co.za

CALIBRATION CERTIFICATE

Calibration Certificate No: **HI261D14-23**

Unit Under Test (UUT)	**Thermo-Hygrometer**
Manufacturer	**ACDC Dynamics**
Model	**CTH-609**
Serial No.	**AS00111**
Asset No.	**Not Marked**
Condition of Instrument	**Good**
Customer	**Azuresky** **Unit 3,17 Makriel Road** **Wadeville, Germiston** **1401**
Date of Calibration	**16-19 January 2026**
Calibration Due Date	**19 January 2027**
Date of Issue	**19 January 2026**

Montech Calibration Services laboratory is accredited by SANAS (South African National Accreditation System). Calibrations and Measurements performed are traceable to the National and International standards thus conforms to ISO/IEC 17025 requirements. This is to ensure the unbroken chain of traceability is maintained and for the realisation of the International Systems of Units (SI).The measurement results recorded in this certificate were correct at the time of calibration and only relates to the equipment detailed on this certificate. The subsequent accuracy will depend on factors such as care, handling and frequency of use. It is recommended that recalibration be undertaken at an interval that will ensure that the instrument remains within the desired limits. This calibration certificate may not be reproduced without prior permission of Montech Calibration Services Pty Ltd and SANAS. Calibrations are performed using approved procedures as per quality manual and ISO/IEC 17025. Calibration certificates without signature are not valid. The results given relate only to the item (s) calibrated and can only be guaranteed at the time of issue. The SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM (SANAS) is a member of the International Laboratory Accreditation Cooperation (ILAC) mutual recognition Arrangement (MRA). This arrangement allows for the mutual recognition of Technical Test and calibration data by member accreditation bodies. For more information on the arrangement please contact www.ilac.org

Technical Signatory



Seola Targett

MONTECH CALIBRATION SERVICES
measure with confidence

sanas Calibration Laboratory
CAL 071 17 00 CAL 071 03 00 CAL 071 02 00 CAL 071 01 00

CALIBRATION CERTIFICATE

Unit Under Test (UUT):	**Thermo-Hygrometer**	
Model:	**CTH-609**	
Resolution:	**0.1 °C**	**1.0 % rh**
Serial Number:	**AS00111**	
Asset Number:	**Not Marked**	
Adjustment made to UUT:	**None**	

Calibration Standards	Serial Number	Calibration Due Date	Certificate Number	Measurement Uncertainty
Digital Thermo-Hygrometer-HP32	5250589 & 25031475	August 2026	NMISA-TEM-2025-9790-A	± 0,60- 1,60 % rh & 0.20 °C
Temperature and Humidity Data Logger	23020795082	October 2026	TI2510D20-19	± 2 % rh and 0.2 °C
Humidity Generator/ Source	Humidity/Temperature Chamber (Serial Number: 08-47782)			

Calibration Procedure :	The calibration was performed by means of comparison to a reference standard in a controlled environment chamber.	
Procedure Number:	CP23-Thermo-Hygrometer	
Measuring Conditions:	22.1 °C	45.9
Standard/s Used:	(5210288 & 20464287	

The calibration was performed at the Montech premises in a controlled laboratory (Address on page 1).

CALIBRATION RESULTS

HUMIDITY RESULTS

Standard Temperature °C	Standard Humidity % rh	UUT Humidity % rh	Correction % rh	Measurement Uncertainty % rh (±)
15	**33**	**34**	**-1**	**2.0**
23	**55**	**53**	**2**	**2.0**
25	**76**	**76**	**0**	**2.0**

TEMPERATURE RESULTS

Immersion Depth (mm)	Standard Temperature °C	UUT Temperature °C	Correction °C	Measurement Uncertainty °C (±)
N/A	**14.9**	**14.5**	**0.4**	**0.4**

The reported expanded uncertainty is based on a standard uncertainty multiplied by the coverage factor k= 2, providing a level of confidence of approximately 95%, the uncertainty of measurement has been estimated in accordance with the principles defined in the GUM, Guide to Uncertainty of Measurement.

Calibrated by	Checked by	Technical Signatory
Phuti Mashamaite	Susan Gumunyu	Seola Targett

--- x --- End of Certificate --- x ---