MICRON LABORATORY SERVICES

9 Beatty Avenue, Witbank 1035 Tel: 013 690 1532 Fax: 013 656 1000
Postnet Suite 139 Private Bag X 7260 Witbank 1035
Reg No: 2017/365305/07 Vat nr: 4430102030
www.micronlab.co.za

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

SANAS Accredited Calibration Laboratory
Temperature Metrology

Certificate of Calibration

Certificate no:	OBS T2.2b
Date of calibration	06 March 2026
Date of next calibration	March 2027

For:	AzureSky (Pty) Ltd
	Unit 3, 17 Makriel Road
	Wadeville, Germiston
Contact:	nonhlanhla@azuresky.it.com
Location:	Laboratory

The results of all measurements are traceable to the national measuring standards.
The values in this certificate are correct at the time of calibration and only relates to the equipment detailed on this certificate. Subsequently the accuracy will depend on such factors as the care executed in handlin and use of the device , and the frequency of use. Recalibration should be performed after the period so chosen to ensure that the instrument's accuracy remains within the desired limits.

This certificate is issued in accordance with the conditions of the accreditation granted by the South African National Accreditation System (SANAS). It is a correct record of the measurements made. This certificate may not be reproduced other than in full except with prior written approval of the issuing laboratory. Legal liability shall be limited to the cost of recalibration and or certification, but the applicant indemnifies Micron Laboratory Services (PTY) LTD against any consequential or other loss.

The South African National Accreditation System (SANAS) is a member of the International Laboratory Accreditation cooperation (ILAC) Mutual Recognition Arrangement (MRA). This arrangement allows for the mutual recognition of technical test and calibration data by the member accreditation bodies worldwide. For more information on the Arrangement please contact www.ilac.org

Technical Signatory: M Stroebel

Calibrated by: M van Jaarsveld

MICRON LABORATORY SERVICES

t +27 (0)13 690 1532 | f +27 (0)13 656 1000 | e admin@mlabs.co.za | 9 Beatty Avenue, Witbank, 1035
Postnet suite 139, Private bag X7260, Witbank, 1035

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

Certificate of Calibration

Calibrated for	AzureSky (Pty) Ltd	**Procedure Number**	P1.3.2.TP	
Certificate Number	OBS T2.2b	**Date of Calibration**	06 March 2026	
Calibration of a	Digital Thermometer and Solid Probe	**Date of Next Calibration**	March 2027	
Manufacturer and Type	Sinometer DM6801B and Type K	**Laboratory Temperature °C**	21.5	± 2°C
Serial Number	S1237965	**Laboratory Humidity %**	74.3	± 10%
Company number	AS8284	**Unit Condition**	Good	

Reference Standards

Description	Make	Model	Coy No:	Re-calibration date
Temperature Probe	NMISA	R	C1-R	30th September 2030
Precision Multi-meter	Tektronix	DMM 4050	MM2	28th February 2027
Tube Furnace	Micron	1000	TF4	N/A
Data logger	Hobo	UX100-001	MLS EL1	30th September 2027

Nominal Temperature °C	DUT Temperature °C		Uncertainty of Measurement ± °C	Immersion Depth mm
	Probe 1	Correction		
500	499	1	1.8	140
815	815	0	1.8	140
900	897	3	1.8	

The reported uncertainty of measurement is based on a standard uncertainty multiplied by a coverage factor of k = 2 ,which unless specifically stated otherwise, provides a confidence level of 95%, in accordance with the Guide to the Uncertainty in Measurement, first edition, 1993.

Observations / Disclaimers:

Remarks :

Technical Signatory: M Stroebel

Calibrated by: M van Jaarsveld

End of Certificate