NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**

NC Report Number: **AB01**

Area/Scope Assessed: **Management Requirements**

Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

At the time of the assessment it was observed that the Impartiality policy ASFM 4.2, was in place, but personnel did not sign the acknowledgement of the policy.

Signed (Representative):

Signed (Team Leader / Inspector): A.Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 8.4.2**

Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

Date:

Signed (Representative):

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS