sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**

NC Report Number: **AB20**

Area/Scope Assessed: **Management Requirements**

Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

At the time of the assessment, it was observed that Procedure ASMF8.7 does not clearly state how often will the lab conduct internal audits. Furthermore Audit plan AUI001 has a period date of November-December 2025 and the next review was scheduled for Feb 2026. Furthermore the plan indicated that an audit will be conducted in November-December 2025, but only done in Jan-Feb 2026 with no reasonable deviation. Furthermore the lab does not have an Internal Audit plan for 2026.

Signed (Representative):

Signed (Team Leader / Inspector): A. Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 8.8.2a**

Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

⇒ Planning vs execution alignment
⇒ Unmonitored deviation
⇒ Audit resource requirement

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

⇒ Draft 2026 audit plan.
⇒ Procedural clarification.
⇒ ~~External verification~~

Date: 21/04/2026

Signed (~~Representative~~):

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: AB20

Brainstorming team

Tshepo Mapanya
Thabo Nyambi
Nonhlanhla Msibi

What went wrong? ASMF 8.7 does not specify the frequency for internal audits. Furthermore the audit plan for NOV-DEC 2025 had a next review date for Feb 2026 but was not reviewed. Additionally the plan indicated that the audit was to take place in Nov-december 2025 but was only done in Jan-Feb 2026 with no reason for the deviation. There is also no internal audit plan for 2026.

Why 1? Why was the internal audit conducted almost two mths later and why was there no audit plan for 2026.

The reliance of the internal audit resources lead to the 2025 year-end audit being pushed into early 2026 due to Operational workload, with no contingency planned identified to maintain the planned interval.


Supervisor:

AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: AB20



SUPERVISOR

Additional Space for NC's

IRQ no: AB20

Extent of the non-conformity: A review of the internal audit cycle showed that the 2025 audit was completed in Jan-Feb 2026, roughly two months after the scheduled date of Nov-Dec 2025. Additionally, no formalized plan for the 2026 calender year had been authorised at the time of the assessment, indicating a lapse in the forward planning of the management system.

Root cause:

1. Planning vs execution alignment: While procedure ASMF 8.7. delegates the audit frequency for the annual plan (AU001) it lacks the "scheduling trigger" to ensure the next years plan is drafted an approved before current cycle expires.
2. Unmanaged deviation: The delay in the 2025 audit (Nov-Dec 2025 to Jan-Feb) was treated as an informal operational delay. Because the procedure did not require a formal "change note" or "deviation record" for rescheduled audits, the reason for the delay was not documented, leading to the appearance of a lack of system control.
3. Audit resource requirement: The reliance on the internal resources led to the 2025 year-end audit being pushed into early 2026 due to operational workload, with no contingency plan identified to maintain the planned intervals.

Corrective action taken

① 2026 Audit plan: The internal auditplan for 2026 (AU001) has been formally drafted, approved and authorised. The plan ensures that all ISO17025 clauses are covered.

② Procedural clarification: While the CAB maintains that the frequency is best managed via the annual plan, procedure ASMF 8.7 has been up dated to include a requirement that "the audit plan for the following year must be approved no later than 30th November of the current year" to prevent a planning vacuum.

③ External verification - To ensure the integrity of the system and proper enclosure of all current NCs, an external consultant has been contracted to perform a "pre-re-evaluation" audit". This provides impartial check of CAMs before SANAS return visit.

Supervisor:

Before

Laboratory Audit Procedure

Azuresky management shall be responsible for scheduling and performing laboratory audits. The overview of the audit procedure is illustrated below. The management team shall:

- Step 1: plan the audit,
- Step 2: communicate planned audit,
- Step 3: Execute the planned audit and record all non-conformities
- Step 4: Address all non-conformities recorded
- Step 5: Close the audit and share the outcomes of the audit with management.

laboratory audit schedule

The schedule of the laboratory audit program shall be developed by Management team and reviewed annually or when risks and opportunities that which required continuous evaluation through audits are identified. The audit program shall be documented and signed off by the Operations Manager before the start of each financial year. The audit program schedule shall be documented and shall be readily available on QMTEKLIMS server.

Planning for the laboratory audit

The Management team shall establish the laboratory audit objectives and scope. Audit objectives should be consistent with and support the management system policy and its objectives.

Audit objectives shall be based on the following considerations:

- Business risks e.g. legal, contractual and other requirements applicable to the Azuresky (Pty) Ltd;
- Results of the previous audits;
- Process, products and project performance;

Audit Scope shall be based on the Technical and Management Requirements.
The audit shall be notified to all personnel **verbally or through production meeting.**

The Management team shall discuss and appoint the Lead Auditor.

- The Lead Auditor shall generate the scope of the audit and any requirements to the Auditee at least 24 hours prior to the day of the audit or unless otherwise there is urgent need to conduct the audit with immediate effect.

After

AzurESky	INTERNAL AUDITS PROCEDURE	
	Document no: ASMF 8.7	
	Compiled by: Sifiso Nyambi - Laboratory Manager	
	Approved by: Nonhlanhla Msibi - Operations Manager	

Laboratory Audit Procedure

Azuresky management shall be responsible for scheduling and performing laboratory audits. The overview of the audit procedure is illustrated below. The management team shall:

- Step 1: plan the audit,
- Step 2: communicate planned audit,
- Step 3: Execute the planned audit and record all non-conformities
- Step 4: Address all non-conformities recorded
- Step 5: Close the audit and share the outcomes of the audit with management.

laboratory audit schedule

The schedule of the laboratory audit program shall be developed by Management team and reviewed annually or when risks and opportunities that which required continuous evaluation through audits are identified. The audit program shall be documented and signed off by the Operations Manager before the start of each financial year. The audit program schedule shall be documented and shall be readily available on QMTEKLIMS server.

The CAB maintains that the frequency is best managed via the annual plan. The audit plan for the following year must be approved no later than 30th November of the current year.

Planning for the laboratory audit

The Management team shall establish the laboratory audit objectives and scope. Audit objectives should be consistent with and support the management system policy and its objectives.

Audit objectives shall be based on the following considerations:

- Business risks e.g. legal, contractual and other requirements applicable to the Azuresky (Pty) Ltd;
- Results of the previous audits;
- Process, products and project performance;

Audit Scope shall be based on the Technical and Management Requirements.
The audit shall be notified to all personnel **verbally or through production meeting.**

The Management team shall discuss and appoint the Lead Auditor.

CONTROLLED DOC

AzurESky	**INTERNAL AUDITS PROCEDURE**
	Document no: ASMF 8.7

Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

Date	Rev.	Content Changed	Reason for change
21/04/2026	03	Addition: Audit plan for the following year must be done by the 30th November 2026.	To prevent a planning vacuum.

CONTROLLED DOC

Effective date	Rev status	Next review date
04/2026	03	04/2029

AzurESky	**Document No.** AS008MF	**Page 1 of 1**
	Effective Date	**April 2026**
	Revision Date	**April 2029**
Title:	Memorandum	
Document by:	Azuresky Pty Ltd	**Revision No**
Compiled by:	Sifiso Nyambi & Tshepo Maponya	**00**
Authorizing Officer: N Msibi	**Signature:**	

Department: Management

Memorandum no: 020

Position: Lab Manager

Date: 21.04.2026

Title: ASMF 8.7 Internal Audit procedure - schedule

Internal Audit schedule must be approved by no later than 30 November of the current year.

AzurESky	ATTENDANCE REGISTER		
	Document no:	Version:	Page:
	ASMF 17043-8.4.1	0	1
	Compiled by:	Sifiso Nyambi - Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	
Date:			
Chair:		Scribe	
Subject:	ASMF 8.7 Internal audit procedure – Laboratory audit plan must be approved no later than 30 Nov of the current year.		
Absent:			

Present	Cell Number	Designation	Signature	Email address
TMS Mapoma	076 455 0692	Lab Supervisor		tshepo@azuresky.it.com
SB Mncimbi	0721741892	Lab Manager		sifiso@azuresky.it.com
N. Msibi	0827868149	Ops Manager		nonhlanhla@azuresky.it.com

AzurESky	**INTERNAL AUDIT PLAN**		
	Document no: AUI001		
	Compiled by:	**Tshepo Maponya – Laboratory Supervisor**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

INTERNAL AUDIT PROGRAMME

4. Audit Schedule (May – December 2026)

Week / Date	Audit Focus Area	ISO 17025 Clauses	Audit Scope & Key Activities	Responsible Auditor	Outputs / Records
12 May 2026	General requirements	4 & 5	Impartiality, Risks to impartiality and confidentiality Structural requirements	Lead Auditor	Results pack
13 May 2026	Resource requirements	6	Equipment requirements and metrological traceability, personnel requirements, externally provided services	Lead Auditor	Results pack
14 May 2026	Process requirements	7	7.1; 7.9; 7.10 & 7.11	Lead Auditor	Results pack
15 May 2026	Management system	8	All clauses	Lead Auditor	Results pack
July 2026	Documentation, Quality Management &	8.1 – 8.3	Review Quality Manual, objectives, policies, and control of documents. Verify review, approval, and version control processes. Check accessibility of controlled documents.	Lead Auditor	Results pack

	INTERNAL AUDIT PLAN	
AzurESky	Document no: AUI001	
Compiled by:	**Tshepo Maponya – Laboratory Supervisor**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

	Control of Records		Evaluate retention, protection, and disposal of records.		
August 2026	Personnel Competence, Training & Impartiality	6.2, 4.1	Review competence and authorization records. Check impartiality declarations and conflict-of-interest handling. Interview staff for awareness of QMS, methods, and confidentiality requirements. Evaluate training plans and effectiveness.	Lead Auditor / HR	Audit notes, interview forms
September 2026	Facilities, Equipment & Environmental Conditions	6.3 – 6.4	Assess suitability and maintenance of lab facilities. Verify calibration, verification, and maintenance records. Review environmental control logs and calibration traceability. Observe equipment use and handling.	Technical Auditor	Equipment checklist, results pack
October 2026	Testing, Method Validation & Measurement Traceability	7.2 – 7.8	Verify method validation, uncertainty, and verification records. Observe sample handling and testing workflow. Review traceability of standards and reference materials. Evaluate data integrity and calculation verification.	Technical Auditor	Observation sheets, NCR forms
November 2026	Customer-related Activities, Reporting & Complaints	7.1 – 7.10	Review contract reviews, quotations, and job acceptance forms. Evaluate client communication records and confidentiality handling. Verify test reports for accuracy, authorisation, and data transfer checks. Review of complaint records and corrective	Lead Auditor	Results pack, corrective action log

Effective date	Rev status	Next review date
04/2026	00	11/2026

	INTERNAL AUDIT PLAN		
AzurESky 'ensure quality of test'	Document no: AUI001		
	Compiled by:	**Tshepo Maponya – Laboratory Supervisor**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

	Handling		actions taken.		
December 2026	Risk Management, Corrective Action & Management Review Preparation	8.5 – 8.9	Review risk and opportunity register and updates. Evaluate recent corrective actions for root cause and effectiveness. Verify audit follow-ups and closure of NCRs. Review readiness for Management Review (inputs and outputs).	Quality Manager / Lead Auditor	Risk register, CAPA tracking, results pack

Disclaimer: This plan will be carried out in accordance with 8.7. internal audits procedure, which serves as the main guideline for conducting audits

Authorisation of the plan

Name: NP Msibi
Designation: Operations Manager
Date: 28.04.2026
Signature:

Effective date	Rev status	Next review date
04/2026	00	11/2026