AzurESky		
AS018TF	**ROOM TEMPERATURE**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

LABORATORY RELATIVE HUMIDITY AND TEMPERATURE RECORDING

Date: _____ Location: Laboratory

				TIMES READINGS TAKEN	READINGS TAKEN BY
Temp. (^{0}C)	SPECIFICATIONS: 19-26	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25-60	ACTUAL READING			
Temp. (^{0}C)	SPECIFICATIONS: 19-26	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25-60	ACTUAL READING			
Temp. (^{0}C)	SPECIFICATIONS: 19-26	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25-60	ACTUAL READING			

Document no	Revision Status	Effective Date	Revision date
AS018TF	01	04/2025	04/2028

CONTROLLED DOCUMENT

AzurESky		
AS018TF	**ROOM TEMPERATURE**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

Shift Supervisor sign: _____

Revision table

Content changed	Reason	Date
Humidity changed from 25-45% to 25-60%	Initial humidity values were provisional, practical work revealed that the initial conditions could not be fulfilled.	12/2025

Document no	Revision Status	Effective Date	Revision date
AS018TF	01	04/2025	04/2028

CONTROLLED DOCUMENT