AzurESky	
AS024TF	TEMPERATURE RECOVERY VOLATILE MATTER

Compiled by:	Sifiso Nyambi- Laboratory Manager	
Approved by:	Nonhlanhla Msibi- Operations Manager	

LIMS No	FURNACE TEMPERATURE						REMARKS	DONE BY	SUPERVISOR SIGNATURE
	DATE	1ST READ (INITIAL)	2ND READ	3RD READ	MEAN + CORRECTION FACTOR	Furnace ID:			

Document no	Revision Status	Effective Date	Revision date
AS024TF	00	04/2025	04/2028

AzurESky	
AS024TF	TEMPERATURE RECOVERY VOLATILE MATTER

Compiled by:	Sifiso Nyambi- Laboratory Manager	
Approved by:	Nonhlanhla Msibi- Operations Manager	

LIMS No	FURNACE TEMPERATURE						REMARKS	DONE BY	SUPERVISOR SIGNATURE
	DATE	1ST READ (INITIAL)	2ND READ	3RD READ	MEAN + CORRECTION FACTOR	Furnace ID:			

CORRECTION FACTOR: ____ °C

THERMOCOUPLE No: _____

TOLERANCE-LIMITS:900 ± 5 (°C)

Document no	Revision Status	Effective Date	Revision date
AS024TF	00	04/2025	04/2028