MICRON LABORATORY SERVICES

9 Beatty Avenue, Witbank 1035 Tel: 013 690 1532 Fax: 013 656 1000
Postnet Suite 139 Private Bag X 7260 Witbank 1035
Reg No: 2017/365305/07 Vat nr: 4430102030
www.micronlab.co.za

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

SANAS Accredited Calibration Laboratory
Temperature Metrology

# Certificate of Calibration

| | |
|---|---|
| Certificate no: | AZUS TH2.1 |
| Date of calibration | 19 March 2026 |
| Date of next calibration | March 2027 |

| | |
|---|---|
| For: | Azuresky |
| | Unit 3,17 Makriel Road |
| | Wadeville, Germiston , 1401 |
| Contact: | tshepo@azuresky.it.com |
| Location: | Laboratory |

The results of all measurements are traceable to the national measuring standards.
The values in this certificate are correct at the time of calibration and only relates to the equipment detailed on this certificate. Subsequently the accuracy will depend on such factors as the care executed in handlin and use of the device , and the frequency of use. Recalibration should be performed after the period so chosen to ensure that the instrument's accuracy remains within the desired limits.

This certificate is issued in accordance with the conditions of the accreditation granted by the South African National Accreditation System (SANAS). It is a correct record of the measurements made. This certificate may not be reproduced other than in full except with prior written approval of the issuing laboratory. Legal liability shall be limited to the cost of recalibration and or certification, but the applicant indemnifies Micron Laboratory Services (PTY) LTD against any consequential or other loss.

The South African National Accreditation System (SANAS) is a member of the International Laboratory Accreditation cooperation (ILAC) Mutual Recognition Arrangement (MRA). This arrangement allows for the mutual recognition of technical test and calibration data by the member accreditation bodies worldwide. For more information on the Arrangement please contact www.ilac.org

Technical Signatory: Z Sampson

Calibrated by: L Sampson

MICRON LABORATORY SERVICES

t +27 (0)13 690 1532 | f +27 (0)13 656 1000 | e admin@mlabs.co.za | 9 Beatty Avenue, Witbank, 1035
Postnet suite 139, Private bag X7260, Witbank, 1035

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

# Certificate of Calibration

| | | | | |
|---|---|---|---|---|
| **Calibrated for** | Azuresky | **Procedure Number** | P1.5.2.DH | |
| **Certificate Number** | AZUS TH2.1 | **Date of Calibration** | 19 March 2026 | |
| **Calibration of a** | Digital Hygrometer and Probe | **Date of Next Calibration** | March 2027 | |
| **Manufacturer and Type** | Sinometer SDC103 | **Laboratory Temperature °C** | 22 | ± 2°C |
| **Serial Number** | AS0096D | **Laboratory Humidity %** | 57 | ± 10% |
| **Company number** | OBS202503-01 | **Unit Condition** | Good | |

**Reference Standards**

| Description | Make | Model | Coy No: | Re-calibration date |
|---|---|---|---|---|
| Temperature Probe | Isotherm | RTD | C3-RTD | 31st October 2027 |
| Logger | Omega | DP9601 | DL3 | 31st August 2027 |
| Water Bath | Micron | 80 | WB1 | N/A |
| Data logger | Hobo | UX100-001 | MLS EL1 | 30th September 2027 |

| Nominal Temperature °C | DUT Temperature °C | | Uncertainty of Measurement ± °C | Immersion Depth mm |
|---|---|---|---|---|
| | Probe 1 | Correction | | |
| 15 | 15.0 | 0.0 | 0.2 | 140 |
| 25 | 24.9 | 0.1 | 0.2 | 140 |
| 35 | 34.9 | 0.1 | 0.2 | 140 |

The reported uncertainty of measurement is based on a standard uncertainty multiplied by a coverage factor of k = 2 ,which unless specifically stated otherwise, provides a confidence level of 95%, in accordance with the Guide to the Uncertainty in Measurement, first edition, 1993.

Observations / Disclaimers:

Remarks :

Signed by:Zelda Sampson
Signed at:2026-03-26 13:04:36 +02:00
Reason:I approve this document

Technical Signatory: Z Sampson

Calibrated by: L Sampson

*End of Certificate*