MICRON LABORATORY SERVICES

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

9 Beatty Avenue, Witbank 1035 Tel: 013 690 1532 Fax: 013 656 1000
Postnet Suite 139 Private Bag X 7260 Witbank 1035
Reg No: 2017/365305/07 Vat nr: 4430102030
www.micronlab.co.za

CERTIFICATE OF CALIBRATION

Certificate no: OB WS1.1R
Date of calibration: 17.03.2025
Date of next calibration: 31 March 2027

Customer: AzureSky(Pty)Ltd.
Unit 3 , 17 Makriel Road
Wadeville
Germiston
Location: Laboratory
Contact: nonhlanhla@azuresky.it.com

Conditions of Issue:

This certificate is issued in accordance with the conditions of accreditation granted by the South African National Accreditation System of South Africa. Subject to the uncertainties specified, this is a correct record of measurements made.

Copyright of this certificate is owned jointly by SANAS and the issuing laboratory. This certificate may not be reproduced other than in full without written approval of SANAS and of the issuing laboratory.

The values in this certificate were correct at the time of calibration. Subsequently the accuracy will depend on such factors as, the care of handling, the use of the instrument and the frequency of use. Re-calibrations should be performed after a period which has been chosen to ensure that the instrument's accuracy remains within the desired limits.

Where statements of conformity are made in the certificate, the following rules are applied:
Within Tolerance: Results ± expanded uncertainty within limits / specification.
Out of Tolerance: Results ± expanded uncertainty exceeds limits / specification.
Indeterminate: Results ± expanded uncertainty overlaps limits / specification.

Measurement results relate only to the item calibrated.



Technical Signatory: M Stroebel



Calibrated by: J Kruger

The Recalibration Date Was Revised As Requested By The Customer ON 04.08.2026

MICRON LABORATORY SERVICES

t +27 (0)13 690 1532 | f +27 (0)13 656 1000 | e admin@mlabs.co.za | 9 Beatty Avenue, Witbank, 1035
Postnet suite 139, Private bag X7260, Witbank, 1035

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

CERTIFICATE OF CALIBRATION

Customer: AzureSky(Pty)Ltd.
Location: Laboratory
Date of Calibration: 17.03.2025
Date of Next Calibration: 31 March 2027

Certificate Number: OB WS1.1R
Weight Set: OB WS1
Procedure Reference: P1.1.MP
Contact: nonhlanhla@azuresky.it.com

Weight ID	Nominal Value (g)	Actual Value (g)	Measurement Uncertainty (g) ±	Line Number
	100	100.00013	0.00010	1
	50	50.00008	0.00005	2
	20	20.00008	0.00006	3
*	20	20.00005	0.00005	4
	10	10.00002	0.00004	5
	5	5.00006	0.00003	6
	2	2.00002	0.00003	7
*	2	2.00004	0.00003	8
	1	0.99999	0.00003	9

NINE WEIGHTS ONLY

Procedure These weights have been calibrated on the basis of weighings made against standards of known mass and assumed density of **8000** kg/m³ traceable to the international standards.

Uncertainty The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor K=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Mass Density Should weighings be made requiring buoyancy correction, A density of 8000 kg/m3 should be assumed for these weights

Traceability **Standards Ref.No.** Set 1 **Re-Calibration Date:** 31st January 2027

Magnetism The weights have been screened for magnetism line 1 to 9 have been found Non-Magnetic

Temperature 21.1 **ºC ± 2ºC** **Humidity** 62 **% ± 10%**

Remarks This Certificate Supercedes OB WS1.1
The Recalibration Date Was Revised As Requested By The Customer On 04.08.2026

Technical Signatory: M Stroebel

Calibrated By: J Kruger

End of Certificate