F 03-09

sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.:

Organisation Name: Forestry Pty Ltd

NC Report Number: NC017/2026 T

Area/Scope Assessed: Documentation (Internal audits)

Assessor / Inspector / Expert: Sifiso Mganiloi

PART 1: DETAILED OBSERVATION

The internal audit programme lacks clarity of completeness in several key areas. Audit criteria are not explicitly defined as required by ISO 17025, and there is limited documented evidence of formal reporting to management. Corrective actions arising from audits not assigned implementation timelines, and audit record retention is not clearly specified.

Signed (Representative): Mnoleka

Signed (Team Leader / Inspector): SMganiloi

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: ISO 17025 (SANAS F48-07 Clause 8.8)

Date: 11/02/2026

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

The internal audit procedure document existed but had incomplete information in several key areas due to incomplete system documentation

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supporting evidence of implementation)

The internal audit procedure (AU1001) was revised and updated to include omitted information.

Date: 12/02/2026

Signed (Representative): Mnoleka

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Corrective action is sufficient & appropriate.

Date: 12/02/2026

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: NC017/2026-T

Brainstorming team

Nosipho NHlongwane
Katumetse Maleka
Abenathi Jama
Tshepo Mapanya

Problem Statement: The internal audit procedure is incomplete in several key areas.

→ The audit criteria is not clearly defined
→ audit reports are limited
→ No specified implementation timeline for corrective actions & record retention requirements

Why 1: why was the internal audit procedure incomplete in several key areas?

The internal audit procedure document did exist, however it did not clearly define audit criteria, have audit reports & specified implementation timelines for corrective actions & record retention. This was due to poor incomplete system documentation.


Supervisor:

AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: NC017/2026 – T



SUPERVISOR

Before

AzurESky "ensure quality of test"	**INTERNAL AUDITS PROCEDURE**
	Document no: ASMF 8.7

Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

Introduction

This procedure is developed as a guideline for Azuresky (Pty) Ltd to follow when conducting internal laboratory audits. The objective is to assess for compliance to requirements stipulated as per ISO 17025:2017; ISO 17034:2016; ISO 17043:2023 and other relevant SANAS codes of practice.

Supporting Clauses

Scope

This procedure details the activities performed when conducting internal laboratory audit. The procedure covers the process to be followed by Azuresky (Pty) Ltd management.

Purpose

The purpose of this procedure is to describe the requirements for planning, executing audits, establishing records and reporting results of the audits.

Normative/Informative References

Normative

[1] ISO 17034:2016: General requirements for the competence of reference material producers

[2] ISO 17043:2023: General requirements for the competence of proficiency testing providers

[3] ISO 17025:2017: General requirements for the competence of testing and calibration laboratories

[4] ISO 19011:2018: Guidance for auditing management systems

CONTROLLED DOC

Effective date	Rev status	Next review date
07/2025	00	07/2028

- The audit objectives,
- The audit scope,
- Identification of the auditee,
- Identification of audit team and Auditee's participants,
- The dates and locations where the audit activities were conducted
- The audit findings and,
- The audit conclusions.

conduct follow up audit and audit close out

- The Auditee shall within the specified time frame provide the Lead Auditor with evidence of the corrective action implemented.
- The Lead Auditor and the audit team members shall assess the suitability of the evidence of the corrective action submitted by the auditee.
- In the event that the provided corrective actions/evidence are suitable then, the non-conformances shall be closed.
- In the event that the implemented corrective actions provided are not suitable, appropriate action shall be taken to address the gap.
- Should the assessment indicate that the corrective actions are not effectively implemented, then the non-conformance(s) will remain open.
- The Lead Auditor shall conduct the close out sign off the non-conformance indicating the status of the corrective action (Closed/Rejected).
- The copies of non-conformances shall be retained or stored both hardcopy and softcopy.
- The Lead Auditor shall share the audit results at the management review meeting.

Audit records

- All records of the audit evidence shall be readily available both hardcopy and softcopy.

CONTROLLED DOC

K.M Before After

AzurESky "ensure quality of test"	**INTERNAL AUDIT PLAN**		
	Document no: AUI001 PROCEDURE		
	Compiled by:	**Tshepo Maponya – Laboratory Supervisor**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

INTERNAL AUDIT PROGRAMME

Period Covered: November – December 2025
Standard: ISO/IEC 17025:2017
Approved by: Nonhlanhla Msibi
Next Review: February 2026

1. Objective

To verify that the laboratory's management system and technical operations conform to ISO/IEC 17025:2017, internal policies, and regulatory requirements; and to ensure that the system is effectively implemented, maintained, and continually improved.

2. Scope

The audit covers all areas of the laboratory including:
- Management and documentation system
- Personnel competence and impartiality
- Facilities and environmental conditions
- Equipment and calibration
- Testing and method validation
- Sampling, data integrity, and reporting
- Risk management, corrective action, and improvement

ADDED (to address "Audit criteria not addressed"):
Audit Criteria:
Audits shall be conducted against the following criteria:
- ISO/IEC 17025:2017 requirements
- Laboratory Quality Manual and management system procedures
- Approved technical methods and work instructions
- Applicable regulatory, customer, and accreditation body requirements
- Risk management and corrective action procedures

Effective date	Rev status	Next review date
02/2026	01	02/2029

6. Reporting and Closure

Audit reports issued within 5 working days of completion. Corrective actions assigned to responsible personnel with deadlines. Effectiveness verified during the final week's audit or next review. Summary findings presented at the next Management Review meeting.

ADDED (to address "No audit report submitted / results not demonstrated"):

- Audit results shall be formally communicated through an approved Audit Report.
- The Audit Report shall be distributed to relevant management and affected personnel within 12 working days after completion of the audit.
- Audit reports shall be submitted as objective evidence during internal reviews and external assessments.

ADDED (to address "No implementation date addressed"):

- Each nonconformity shall include an agreed implementation date for corrective action.
- Implementation dates shall be recorded on the NCR and corrective action tracking log.
- Delays to implementation shall be justified, documented, and approved by the Laboratory Manager or Operations Manager.

7. Audit Records

Records include:

- Annual audit programme
- Completed checklists and reports
- NCR forms and CAPA tracking sheet
- Attendance and meeting minutes
- Verification of corrective actions

Effective date	Rev status	Next review date
02/2026	01	02/2029

ADDED (to address "No audit records referred to in procedure"):

- Audit records shall be maintained as controlled quality records in accordance with the document and record control procedure.
- Audit records shall provide objective evidence of audit planning, execution, findings, corrective actions, and effectiveness verification.
- Audit records shall be retained for a minimum period defined in the record control procedure and made available for internal and external audits.

8. Evidence of Systematic Coverage

This programme ensures full coverage of both management system (Clauses 8.1–8.9) and technical competence (Clauses 4–7) within the two-month period. It demonstrates that internal audits are planned, systematic, conducted by competent and impartial auditors, and used for continual improvement.

Annex A: Internal Audit Checklist Template

Clause	Requirement	Objective Evidence	Findings / Comments	NC / OFI

This template is used during each audit activity to record evidence and findings against the relevant ISO 17025 clauses.