F 03-09

sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.:

Organisation Name: Azuresky Pty Ltd

NC Report Number: NC07 2026 P

Area/Scope Assessed: Documentation review

Assessor / Inspector / Expert: Ifiso Nyambi

PART 1: DETAILED OBSERVATION

The requirements of the standard, customer, regulatory authorities & organization providing regulation was not addressed/documented in the QMS of the CAB.

Signed (Representative):

Signed (Team Leader / Inspector):

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: ISO 17025 (SANAS 5.4. F 48-07)

Date: 02/09/2026

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

The QMS was still being developed therfore the document that states the Requirement of standards, customers, regulatory authorities was not drafted leading to an incomplete system documentation.

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supporting evidence of implementation)

The drafting of the Requirement of standards, customers, reg auth etc (ASMF5.4) was done

Date: 09/02/2026

Signed (Representative):

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

ASMF5.4. addresses the gap.

Date: 10/02/2026

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: NCO 7/2026 T

Brainstorming team
Keitumetse Maleka; Abenathi Jama; Lilo Photo; Nosipho Hlongwane & Tshepo Maponya

problem Statement : The requirements of the standard, customer, regulatory authorities & organization providing regulation was not addressed/documented in the QMS of the CAB.

Why 1 : why was there no documentation of the Requirements of the Standard, customer, regulatory authorities & organization ?

Due to the lab still starting up in that period of time, the QMS was still being developed, leading to an incomplete system documentation.


Supervisor:

AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: NCO.7/.2026..T.....



.....

SUPERVISOR

Asfer

	Requirements of standards, customers, reg auth etc		
AzurESky	Document no: ASMF5.4		
	Compiled by:	Sifiso Nyambi – Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	

Procedure: Identification and Review of Applicable Requirements

1. Purpose

To ensure that Azuresky (Pty) Ltd identifies, understands, and complies with:
a) the requirements of applicable standards,
b) customer requirements,
c) regulatory and statutory requirements, and
d) requirements of organisations providing recognition (e.g. SANAS), before accepting and performing laboratory activities.

2. Scope

This procedure applies to all laboratory activities, including testing, proficiency testing scheme activities, reference material production, and externally provided services.

3. Sources of Requirements

3.1 Standard Requirements

The laboratory operates a management system aligned with the following standards:
- ISO/IEC 17025:2017
- ISO/IEC 17034:2016
- ISO/IEC 17043:2023

Applicable test methods include, but are not limited to:
- ISO 1171 (Ash)
- ISO 562 (Volatile Matter)
- ISO 1928 (Calorific Value)
- ISO 13909-4 (Sample Preparation)
- SANS 5925 (MIAS)

The laboratory ensures current editions of all applicable standards are available and controlled.

Effective date	Rev status	Next review date
02/2026	00	02/2029