sanas South African National Accreditation System	**GENERAL, STRUCTURAL, RESOURCE AND MANAGEMENT REQUIREMENTS OF ISO 17034:2016**	Organisation's SANAS No/s.

Date/s of evaluation	
Assessor/s & Observers	
Laboratory	**Azuresky (Pty) Ltd**
Area / field of operation	**Reference material producer of Coal**
Laboratory Representative	**Nonhlanhla Msibi**

This report covers the following:

Document Review only		Implementation on Site Visit only		Document Review and Site Visit		Other	

Compliance = C, Non-compliance = NC

REQUIREMENTS & COMMENTS. ***Compliance = C, Non-compliance = NC***

NB1: *References to ISO 17034 are in italics. The order of assessment need not follow the order of the checklist. Assessors are expected to know & have the standard, this worksheet is designed as guidance to prompt detailed recording of the process.*

REFER TO ISO 17034:2016 FOR DETAIL AND FOR CLARIFICATION NOTES.

NOTE 1: *For **CAB's comments**: The CAB must provide information on how requirements have been addressed, documented and/or implemented. Make reference to policies / procedures, incl. clause numbers.*

NOTE 2: *For **Assessor's Comments**: The Assessor must provide information on the CAB's conformity with the requirements*

©SANAS

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
4	**GENERAL REQUIREMENTS**			
4 . 1	**Contractual matters:**			
4.1.1	Are all requests, tenders and contracts for the production of a RM reviewed to ensure:			
a)	The requirements for the RM defined, documented and understood?	• 4.1.2. Provides the requirements/criteria (Checklist) of acceptance for contracts , tenders, and customer requests • 4.7.1. Document outlines the requirements under section 3.2. Initial review.		
b)	The producer has access to the capability and resources?	• 4.7.1. Document outlines the capability assessment under section 3.3.		
NOTE i)	Capability includes, for example access to equipment, knowledge, information resources, personnel with skills and expertise ii) a contract could be written or verbal)			
4.1.2	Does the review include any work to be subcontracted?	• Checklist 4.1.2 ensure adherence to the requirement		
4.1.3	Are records available of the reviews, including any changes and any sub-contracted work?	• Records to be stored for two accreditation cycles.		
4. 2	**Impartiality**			

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
4.2.1	Is there a structure and is it managed so as to safeguard impartiality?	• Maintain clear job descriptions and reporting lines to prevent undue influence. (attached on website) • Include impartiality as part of the onboarding for new employees. • There's impartiality policy ASMF 4.2 implemented. • There's conflict of interest declaration ASMF 4.2.1 completed by personnel		
	NOTE Impartiality implies, decisions based on objective criteria, and not on bias, prejudice or preference for improper reasons.			
4.2.2	Does the producer:-			
a)	Have arrangements that ensure that their staff are free from any undue pressures, including commercial, financial, and any other pressure that may adversely affect their work?	Within the impartiality policy document under Commitment to remaining autonomous, it reads as follows: To ensure independence and objectivity and protect impartiality risks within us as an organization and from any outside influence, the following needs to be enforced: • Continuous mandatory disclosure of conflicts of interests (ASMF 4.2.1) • Separate commercial functions from technical decision-making, ensuring that financial pressures do not affect scientific integrity. • Disclosure of incentives from clients and registering them in the incentive register form (This is to protect personnel from any undue pressure in the organisations' activities.)		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
b)	Identify risks to impartiality on an on-going basis, including risks from the activities, relationships of the producer and personnel? NOTE not all relationships present a risk to impartiality	The lab has assessed the risks and consequences associated with impartiality as outlined by doc ASRI 4.2-001- Risk assessment to impartiality		
c)	Demonstrate, how and identified risk, is eliminated or minimized?	There are existing controls within the ASRI 4.2-001- Risk assessment to impartiality document which highlight how each of the risks identified will be addressed should the need arise.		
d)	Have top management committed to impartiality? NOTE impartiality can be influenced by ownership, governance, management, personnel, shared resources, finances and contracts (other than for the sale of RMs).	Top management have signed the Quality Policy where top management commits to confidentiality and impartiality.		
4.3	**Confidentiality**			
4.3.1	• Does the RMP treat all information in an appropriate manner, including confidential information?	**NDA signed by all personnel see doc 4.3.1 B NON DICLOSURE CONTRACT**		
	• Is information received from other individuals and bodies treated as confidential unless the information is in the public domain, or its disclosure agreed to?	**Covered in section 5.5. of the 4.3. Confidentiality procedure.**		
4.3.2	When the RMP is required by law or authorized to release confidential information, is the individual or body notified? *(unless prohibited by law)*	**Covered in section 5.5. of the 4.3. Confidentiality procedure.**		
6.1.2	Do all personnel, including sub-contractors comply with the policies and procedures set by the RMP for the management of confidential information?	• All Azuresky personnel signed NDA's • Visitors or contractors or sub-contractors register their attendance in the visitor's register book which has included a statement for them to manage all information obtained at Azuresky confidentially.		
5.	**STRUCTURAL REQUIREMENTS**			

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
5.1	Is the RMP a legal entity, or a defined part of a legal entity, that can be held responsible for their activities? (List the legal name and registration number/s of the RMP)	• The company is a legally registered entity, details of registration can be found under ASMF5 section 1		
5.2	Is the RMP organized and operated in such a way as to meet the requirements of ISO 17034?	• Azuresky RM producer carries the activities of the RM production consistent with the requirements stated in ISO/IEC 17034:2016 see **Management system document ASDP001**		
5.3	Does the RMP:-			
a)	• Have a description of its legal status?	• The company is a proprietary limited see ASMF5 section 1		
	• Define the organizational and management structure?	• Organogram outlining the line of reporting (hierarchy) see ASMF 5.1 • Job descriptions outlining scope of authority as well as duties/responsibilities.		
	• Define its place in any parent organization?	• The company is a private entity see ASMF5 section 1		
	• Define the relationship between management, technical operations, support services and subcontractors?	• Organogram outlining the line of reporting (hierarchy) see ASMF 5.1 • Job descriptions outlining scope of authority as well as duties/responsibilities.		
b)	Define the parts of the organization covered by the management system for RM production?	• The company has designed all PT and RMP activities to be handled by all personnel on the organogram (small company)		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
c)	Specify the responsibility, authority and inter-relationships of personnel whose work could affect the quality of the RM produced?	• **It is all personnel** • Organogram outlining the line of reporting (hierarchy) see ASMF 5.1 • Job descriptions outlining scope of authority as well as duties/responsibilities.		
d)	• Have managerial personnel with the authority and resources required?	**Yes**		
	• Able to identify departures from the management system and procedures?	• Non-conforming work and corrective action procedure documented, see doc 7.17. Management of non-conforming work • Improvement Action Request doc 8.7.3. used to initiate actions to prevent deviations, and identify NC work, to prevent future reoccurrence.		
	• Able to initiate actions to prevent or minimize these departures?	• 8.5 Actions to address risks and opportunities were identified, control measures defined and implemented. See attachment of SANAS website Azuresky uploaded documents • There are also other forms used in the root cause analysis, internal audits. Forms such as AS001MF, AS003MF , AS004MF		
e)	• Have technical management with responsibility for technical operations?	• Organogram outlining the line of reporting (hierarchy) see ASMF 5.1 • Job descriptions outlining scope of authority as well as duties/responsibilities.		
	• Resources to ensure the required quality?	Technical equipment, LIMS for recording and traceability of results, etc are available		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
f)	• Have personnel with defined responsibility and authority to ensure that the requirements of ISO 17034 are implemented and followed?	• Organogram outlining the line of reporting (hierarchy) see ASMF 5.1 • Job descriptions outlining scope of authority as well as duties/responsibilities.		
	• And having access to highest management responsible for RM production?	• Azuresky is small company this is not a hindrance at all. See organogram structure and structural requirements ASMF 5		
g)	Have adequate insurance or provisions to cover liabilities?	• Only insurance for company vehicle is in place. • Insurance for equipment is not in place due to unresolved dispute to obtain COC from landlord to update additional electrical fixtures on it.		
5.4	Has the RMP ensured that:			
a)	Internal & external communication mechanisms have been established?	• Any changes or improvements to the management system is communicated to personnel via memorandums and/or WhatsApp group. 5.6. Azuresky memorandum template • 8.2. Quality Policy Statement also outlines the review of management system through auditing the system periodically to provide confidence in its effectiveness. **Lastly, management review meeting to held in Jan 2026 will address this requirement.**		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
b)	The effectiveness of the management system is communicated?	• Any changes or improvements to the management system is communicated to personnel via memorandums and/or WhatsApp group. 5.6. Azuresky memorandum template • 8.2. Quality Policy Statement also outlines the review of management system through auditing the system periodically to provide confidence in its effectiveness. **Lastly, management review meeting to held in Jan 2026 will address this requirement.**		
c)	The importance of meeting customer requirements is communicated to the RMP personnel?	• Any changes or improvements to the management system is communicated to personnel via memorandums and/or WhatsApp group. 5.6. Azuresky memorandum template		
6.	**RESOURCE REQUIREMENTS**			
6.1	**Personnel**			
6.1.1	• Are all personnel involved in RM production supervised and competent?	• Yes competency declaration of all personnel from operator to top management uploaded on SANAS website		
	• Do the personnel work in accordance with the RMP's management system?	• Yes, 8.7.3 Improvement action request is implemented to record and investigate deviations. • 8.5 Actions to address risks and opprortunities was identified, control measures identified this in order enable personnel to effectively manage the system		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
6.1.3	• Has the RMP ensured the competence of personnel involved in the production of RM's?	• Yes competency declaration of all personnel from operator to top management uploaded on SANAS website • ASDP001 document clause 2.1 states that All competence and authorizations must be reviewed annually or when there are any role changes. Records must be kept for 2 accreditation cycle.		
	• Are there sufficient personnel having the necessary, education, training, technical knowledge and experience for their assigned functions?	• Yes this is requirement is documented per post in ASDP001 document clause 2.1		
6.1.4	• Has the RMP identified the training needs of personnel?	• Training matrix has been developed and implemented		*(Record the details of records examined – Names, positions, authorizations, education, training and job descriptions)*
	• Has training been provided?	• Testing methods related to RM production were conducted, outstanding is other training needs listed on training matrix.		
	• Is the training program relevant to present and anticipated tasks of the RMP?	• Yes, as codes of practices are reviewed, all procedures, policies or checklists, etc will be reviewed, authorised, implemented and training conducted.		
6.1.5	Are there Job Descriptions for personnel involved in the RM Production?	• Yes all uploaded on SANAS website		
6.1.6	Has the RMP authorized as competent personnel involved in the RM Production?			
	• Has the RMP maintained records for staff:	• See uploaded documents on SANAS website		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Authorizations?	• Yes all uploaded on SANAS website		
	• Competence?	• Yes all uploaded on SANAS website		
	• Educational and professional qualifications?	• Yes all uploaded on SANAS website		
	• Do these records provide sufficient evidence that the individuals are trained and competent?	• Yes subject to SANAS auditor assessment		
	• Are the records available, and dated?	• Yes all uploaded on SANAS website		
6.2	**Subcontracting:**			
6.2.1	Where the RMP makes use of subcontractors for production, sampling, processing, handling, homogeneity and stability testing, characterization, storage or distribution, have procedures been implemented that ensure that the subcontractors experience and technical competence sufficient for their assigned tasks? (Provide details of the procedure for sub-contracting and its implementation)	• No subcontracting was/is applied at the company.		
6.2.2	Are sub-contractors selected on the basis of their ability to meet the requirements of the RMP?	• Document ASMF6.2. defines the procedure for selecting, evaluating, approving, and monitoring of subcontractors.		
6.2.3	Have any of the following processes being sub-contracted?			

©SANAS

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
i)	production planning	**No,** ASMF 7.2.1 production planning **NB: Only one production plan of CRM is presented here, others will be assessed onsite.**		
ii)	selection of sub-contractors	**No**		
iii)	assignment of property values and uncertainties	**No,** ASMF 7.13		
iv)	authorization of property values and uncertainties	CRM 002 UNCERTAINTIES MAY **No,** **NB: only uncertainty budget of AS002 CRM attached only, others will be presented onsite**		
v)	authorization of RM documents	**No** AS002 CRM certificate		
	(Note - none of the aforementioned tasks may be sub-contracted)			
6.2.4	Has the RMP established and maintained procedures to ***assess*** all tasks performed by sub-contractors to ensure they comply with the clauses of this standard?	• No subcontracting was/is applied at the company.		
6.2.5	Is evidence available of the competence of sub-contractors, including evaluations and audits of the sub-contractor/s?	• None, however a checklist has been developed		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
6.2.6	In cases where it is not possible to confirm the competence of subcontractors by documentary evidence, has the RMP either evaluated the competence or supervised the activities of the sub-contractor/s? (Provide details of the evidence, documentary and otherwise confirming the competence of the sub-contractors)	• Not applicable		
6.2.7	Does the RMP have results and descriptions of procedures available for technical data evaluation?	• ASMF 7.11 Data integrity risk management procedure		
6.2.8	Does the RMP have personnel with sufficient knowledge of the sub-contractor's tasks, in order to evaluate their activity?	• Not applicable		
6.3	**Provision of equipment, service and supplies:**			
6.3.1	Are procedures available for the selection of equipment, services and supplies that can affect the quality of the RM's?	• **ASMF 6.3 Provision of Equipment, Service and supplies**		
6.3.2	Is only equipment, services and supplies that comply with specifications used?	• **Yes, test methods (Ash/ Cv/ etc) are specific as per requirements**		
6.3.3	Is equipment & consumables affecting quality of results not used until inspected/calibrated/verified?	•		
6.3.4	Has the RMP kept records of all equipment, services and supplies, including the selection criteria used, acceptance and commissioning data?	• **Yes , all is available**		
6.4	**Facilities and environmental conditions:**			

©SANAS

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
6.4.1	Has the RPM ensured that all the lab facilities, calibration and test areas (where applicable), material handling, storage, processing and packaging areas, energy sources, lighting, humidity, temperature, pressure, and ventilation are such as to facilitate proper material handling, storage, processing and packaging, and performance of calibration and testing (where applicable)? (provide details of the facilities and environmental conditions)	• In each Working Instruction, the organization has included environmental requirements needed to perform the method of interest together with its tolerance ranges and monitoring frequencies. (e.g. 19-26 °C for the analysis room, 16-20 °C for the storage facility, 25-45 % RH on both), tolerances and monitoring frequency. • See conditions stated in section 5 of ASMF 7.2.3 Environmental control which addressed the parameter, it's control range and the area in the process to be controlled.		
6.4.2	In cases where the environmental conditions could have an effect on the RM are these environmental conditions monitored, recorded and controlled?	• The laboratory has in place aircons in the analysis room and storage facilities that control the temperature to the desired ranges. Hygrometers are present to always read the room temperature and help monitor the relative humidity as well. Deviations outside the acceptable ranges trigger investigation and possible data flagging.		
	Is the equipment used to monitor the environmental conditions calibrated? (provide details of the environmental condition monitoring and control equipment and its calibration status)	• Yes and calibration certificates are available.		
6.4.3	Are other factors, other than temperature and humidity, such as incompatible activities, vibration, aerosols, airborne dust, microbiological contamination, magnetic fields, light and electromagnetic and ionizing radiation considered, and are RM Processing, Testing and Calibration Areas protected from these conditions?	• **Yes,** See conditions stated in section 5 of ASMF 7.2.3 Environmental control which addressed the parameter, it's control range and the area in the process to be controlled. They are controlled.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
6.4.4	Is access control implemented where necessary to maintain the quality of the RM's?	• All authorised personnel as represented on the organogram have access to the laboratory and have received the training to perform their duties and produce quality RM's		
7.15	**Distribution Service**			
7.15.1	• Has the distribution process been specified?	• ASMF 7.15 Distribution and customer support has described two modes which are courier service and inhouse distribution of samples		
	• Does the distribution service include precautions to avoid deterioration of the RM?	• Yes, their described as avoid contamination, direct sunlight, moisture ingress, etc during packing. Tamper proof containers shall be used. ASMF 7.15 Distribution and customer support		
	• Does the distribution service include appropriate documentation to allow customs clearance?	**The business has not yet made applications to the relevant authorities to import products outside the country It is only trading in RSA.**		
7.15.2	Has the RMP kept 'up-to-date' records of all RM sales and distribution?	• Yes, this is the record kept currently:  CRM INVENTORY.xlsx		

©SANAS

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.15.3	Does the RMP provide 'reasonable' guidance & support to its users?	• **RM certificate is provided with the sale items** • **There's complaints and appeals procedure shared publicly on www.azuresky.it.com** • **For any enquiry, customer is able to contact representative through email or telephonically.**		
7.15.4	Has the RMP notified its users of changes in its RM property or uncertainty values?	• Not yet occurred. However requirement is noted and documented in ASMF 7.15 Distribution and customer support		
7.15.5	Has the RMP made use of a distributor for the resale of RM, if yes:-			
	• Has the distributor been provided with all necessary information concerning the RM?	**There's no contractual agreement in place, should the need arises in future, procedure** ASMF 7.15 Distribution and customer support **will be updated to describe how to meet this requirement.**		
	• Are the relevant clauses of the standard adhered to by the distributor?	**N/A**		
7.16 & 8.5 Control of quality and technical records: **Control of records:**				

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.5.1	Have procedures been established that define the controls for the identification, storage, protection, retrieval, retention time and disposition of records?	• ASCR001 record matrix document describe the procedure to protect, access from the management system		
7.16.1	Has the RMP established and maintained procedures for the identification, collection, indexing, access, storage, maintenance and disposal of quality and technical records?	• ASMF 7.16 Control of technical records procedure was developed and implemented. Also the requirements are described in ASDP001 management system document clause 4.2		
7.16.2	Has the RMP kept records of information that may be necessary to resolve a future dispute?	• Yes, A record matrix form ASCR 001, lists the records produced by the laboratory, storage, protection, back-up, archive, retrieval and disposal.		
7.16.3	• Are records legible?	• Yes		
	• Are records stored so that they are readily retrievable, and damage, deterioration and loss prevented?	• Yes, A record matrix form ASCR 001, lists the records produced by the laboratory, storage, protection, back-up, archive, retrieval and disposal.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Are the retention times of records established and documented?	• Yes, ASMF 7.16 Control of technical records procedure was developed and implemented section 7 listed the following: • PT & CRM Production Data: At least 10 years after expiry date • Calibration/Equipment logs: 5-10 Years • Complaints and recalls : 3 Years • Training and competence: Duration of employment +5 years • CoAs: >10 Years		
8.5.2	Are the retention times consistent with contractual and legal obligation?	• Yes, ASDP001 section 4.2 states 10 years retention period.		
7.16.4	• Are mistakes in records crossed out and not erased?	• Yes see section 7 of ASMF 7.16 Control of technical records procedure		
	• Are alterations initialled and dated?	• Yes see section 7 of ASMF 7.16 Control of technical records procedure		
	• Are similar arrangements in place for amendments to electronic records?	• Section 8 of ASMF 7.16 Control of technical records procedure states that "For digital records, maintain audit trails."		
7.16.5	Are records held securely, and where applicable in confidence?	• Yes		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.5.2	Is access to records consistent with confidentiality arrangements?	• Yes sensitive folders are not available to all personnel in the laboratory.		
7.16.6	Are procedures in place to prevent unauthorised access to, and amendment of electronically held data? (examples of electronic data protection include physical, passwords, malware / virus protection)	• Yes, a monthly data back-up is done through external hard drive. • LIMS is protected, everything recorded can't be erased. • Limited access to personnel on LIMS is provided to personnel through allocated passwords of personnel.		
7.16.7	Are measurements, calculations, and derived data, including uncertainty budgets, calibration records and preparation reports kept for a defined period, longer that valid lifetime of the RM?	• Yes, ASMF 7.16 Control of technical records procedure was developed and implemented section 7 listed the following: • PT & CRM Production Data: At least 10 years after expiry date • Calibration/Equipment logs: 5-10 Years • Complaints and recalls : 3 Years • Training and competence: Duration of employment +5 years • CoAs: >10 Years		
7.16.8	Are the results of each calibration or measurement reported in accordance with ISO/IEC 17025?	• Yes, corrective factors applied, etc		
7.17	**Management of non-conforming work**			

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.17.1	Has the RMP implemented procedures whenever its production procedures or requirements are found not to comply?	• ASMF 7.17 management of non-conforming work procedure was developed and implemented as well as ASMF 8.7.3 Improvement Action request was implemented to adequately investigate and analyse the non-conforming work.		
7.17.2	Do these procedures address:			
a)	Responsibilities and authority for the management of non-conforming work?	• Yes responsibilities are described in section 3 of the document		
b)	• The actions to be taken when non-conforming work is identified?	• Yes, section 5 of the procedure states initial steps to be taken		
	• Does this include root cause analysis, and ensure effective implementation?	• Yes, ASMF 8.7.3 Improvement Action request form guides personnel to determine the root cause analysis using analysis tools.		
c)	• Evaluation of the significance of non-conforming work?	• Risk analysis should be assessed, and this requirement is stated in section 7 of ASMF 7.17 management of non-conforming work		
	• Identification and implementation of corrective action?	• Yes, ASMF 8.7.3 Improvement Action request form provides this section.		
d)	The halting of work, and withholding of certificates and other appropriate documentation?	• Yes, section 5 of the procedure states initial steps to be taken		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
e)	Remedial actions including customer notifications within specified time frames?	• Yes, section 6 stated that "Immediate action taken (containment),If necessary ensure notification to customers within 24 hours of recording an NCW"		
f)	• Where necessary the notification of users, and the recall of RM's and certificates and	• Requirements is stated in section 8 of ASMF 7.17 management of non-conforming work		
	• Documentation within an appropriate period?	• Requirements is stated in section 8 of ASMF 7.17 management of non-conforming work		
g)	The responsibility for the resumption of work?	• Section 9 of ASMF 7.17 management of non-conforming work states that "Document closure with review sign-off from operations/lab manager and resume work."		
h)	Where necessary, and Internal Audit to be held to evaluate the corrective action taken?	• Section 9 of ASMF 7.17 management of non-conforming work states that "A planned Internal Audit must include the checking or evaluate the corrective action taken, this it ensure effectiveness of the implementation"		
7.17.3	Has any decision to recall RM's been taken in a timely manner?	• The incident has not occurred.		
7.18	**Complaints:**			

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.18.1	Does the RMP have a documented process (procedure) for the receipt, evaluation and decision making on complaints?	• ASMF 7.18 Complaints handling procedure		
7.18.2	Is the description for the handling process of complaints made available to any interested party on request?	• Yes procedure will be shared with the interested party and it is also displayed on public platform www.azuresky.it.com		
7.18.3	Has the RMP on receipt of a complaint determined if it related to the services it provides, and if so, have they dealt with it?	• No complaint or appeal received currently.		
7.18.4	Is the RMP responsible for all decisions, at all levels, related to the complaint?	**Yes**, the following three instructions are mentioned inn the complaints handling procedure: • Take the final decision after careful review of the evidence gathered and where necessary, consult with Operations Manager. • Communicate the official outcome to the complainant and if necessary, offer further assistance which may include replacement of a batch of certified reference material if appropriate. • Risk analysis shall be undertaken and logged in the risk register.		
7.18.5	Have the investigations and decisions on complaints resulted in any discriminatory actions?	• No complaint or appeal received currently.		
7.18.6	Does the process for handling complaints include at least the following:			

©SANAS

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
a)	A description on the process for receiving, validating, investigating and decisions?	• Yes, all four points are covered.		
b)	Tracking and recording of complaints, including actions to resolve the complaints?	• Yes, this statement it is made in the procedure: "Record, non-conformity in the ASMF 8.7.3 Improvement Action request document." 7.10 non-conforming work register is implemented		
c)	Ensuring that appropriate action is taken?	• Yes, it is stated in the procedure that "Take the final decision after careful review of the evidence gathered and where necessary,"		
7.18.7	Is the RMP responsible for receiving, gathering and verifying information to validate the complaint?	• Yes, it is stated in the procedure that "Gather all evidence needed for the investigation which will include but not be limited to test records, shipping logs and communication history"		
7.18.8	Has the RMP:-			
	• Acknowledged receipt of the complaint (where possible)?	• No complaint or appeal received currently. • But requirement it is stated in the procedure.		
	• Provided the complainant with progress reports and the outcome?	• **It** is stated in the procedure that "Should corrective/preventive action take time to resolve as per described action on ASMF 8.7.3 Improvement Action request, provide progress report to the complainant and the final outcome.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.18.9	Is the decision reviewed, approved and communicated to the complainant by individual(s) not involved in the original RM activities in question?	• Yes, this requirements has been included in the procedure		
7.18.10	Wherever possible has the complainant been provided with formal notice at the end of the process?	• The procedure provides this instruction that complainant shall be provided with progressive report and final outcome		
8. MANAGEMENT SYSTEM REQUIREMENTS				
8.2 Quality Policy				
8.2.1	Has the RMP defined and documented its quality policy?	• Yes, it is document ASMSDoc 8.2-01 (uploaded on SANAS website)		
	Does the policy include a commitment to ensure and maintain the quality of all aspects of the RM production, storage and distribution?	• **Yes**		
8.2.2	Is the RMP policies related to quality, and quality policy statement issued under the authority of 'top management'?	• **Yes**		
8.2.3	Does the quality policy include the following:			
a)	To produce RM's which conform to this International standard?	• **Yes**		
b)	To conduct testing and calibration in compliance with ISO/IEC 17025?	• **Yes**		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
c)	To require personnel to familiarize themselves, and implement the policies and procedures in their work?	• Not all personnel defined on the organogram require exposure to all management system documentation. • Ie. Operators and analysts have received competence training for sample preparation and all testing procedures, quality policy and quality manual only. • Top management (Laboratory supervisor, Laboratory Manager and Operations Manager) all developed the whole management system documentation.		
d)	• to continuously improve the effectiveness of the management system, and	• The CAB was established in April 2025, the first management review will be held in January 2026; thereafter, annually. • KPI mid-term review		
	• be committed to good professional practice, and	• Yes		
	• the quality of the RM's produced?	• Document ASMSDoc 8.2-01 (uploaded on SANAS website), makes commitment by Top management to produce quality RM as per ISO 17034:2016		
8.2.4	Have the overall objectives been reviewed during the management review?	Management review is scheduled Jan 2025, however mid-term objectives have been rated to provide direction of the business • KPI mid-term review		
8.3	**General management of system documentation**			
8.3	• Has the RMP adequately documented its systems, programmes, procedures Instructions and findings?	Yes, see the attached documents on website		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Is the RMP documentation communicated, available and understood and implemented by the RMP personnel?			
8.4	**Control of management system documents**			
8.4.1	Does the RMP control the documents required to fulfil the requirements of this standard?	• ASMF 8.4 Document control was implemented and references ISO 17025, ISO 17034 and ISO 17043		
8.4.2	Does the RMP ensure: -			
a)	Documents are approved for adequacy prior to issue?	• Yes, Operations Manager signed the documents. ASMF 8.4 Document control section 4.1		
b)	Documents are periodically reviewed and updated?	• Yes, There is dates of review stated on document (ie.5 years or during audits or when codes of practice change) ASMF 8.4 Document control section 4.6		
c)	Changes and the revision status are identified?	• Yes identified, ASMF 8.4 Document control section 4.2		
d)	Relevant versions are available at the point of use?	• Yes this requirements id defined ASMF 8.4 Document control section 4.3		
e)	Documents are uniquely identified, and (where necessary) their distribution controlled?	• Yes, requirement stated in ASMF 8.4 Document control section 4.9 • Each document shall be given a unique reference code. The prefix letters of the code refer to a particular kind of document		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
f)	• The unintended use of obsolete documents prevented?	• Yes requirement stated in ASMF 8.4 Document control section 4.5		
	• Any obsolete document retained are suitably identified?	Withdrawn and obsolete documents are held in the Withdrawn/Obsolete Document file and retained for a specific period with a minimum period of 3 years.		
	(Note - Documents can include external documents, text books, procedures, specifications, tables, charts, including software, on various media – hard copy, electronic, photographic or written)			
8.6	**Management reviews**			
8.6.1	• Is there a predetermined schedule for management reviews?	**As per ASDP001 Management System section 4.7** Management review meetings shall be conducted annually in January month		
	• Is there a procedure for management reviews?	**No, As per ASDP001 Management System section 4.7** The inputs as described in ISO 17043:2023 clause 8.9.2; ISO 17034:2016 clause 8.6.1 and ISO 17025:2017 clause 8.9.2 shall be discussed. A record of the minutes of the meeting shall be developed, reviewed within two weeks of the meeting and signed by the Operations Manager and the Laboratory Manager.		
	• Is the purpose to ensure continuing suitability and effectiveness, and introduce necessary changes and improvements?	Yes, this requirement is stated in ASDP001 Management system clause 4.7		
	• Does the management review address: (as a minimum)			

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
a)	The suitability of policies and procedures?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
b)	Reports from managerial and supervisory personnel?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
c)	Internal audit outcomes?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
d)	Corrective actions?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
e)	The results of risk identification?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
f)	Assessments by external bodies?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
g)	Changes in the scale and type of work?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
h)	Customer feedback?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
i)	Customer complaints and improvements?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
j)	Resources, staff training, technical issues related to subcontractors/distributors?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
k)	The quality objectives, as per the RMP quality policy statement/s?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
8.6.2	• Are the findings and actions from the management reviews recorded?	See the attached template of the minutes designed to be held in January 2026 ASTR 8.6 Management Review meeting minutes		
	• Are the actions discharged in an appropriate and agreed timescale?	This will be recorded and tracked next year		
	(Include details of the procedure, attendance, agenda, minutes, actions arising, date of the last MR)			
8.7	**Internal audit (IA)**			
8.7.1	• Has the RMP conducted IA's periodically?	• Audit of the system is planned for Nov-Dec 2026. This will be the very first comprehensive audit program, given that the company has been business for ~7 months. Please see the audit programme doc AUI001-INTERNAL AUDIT PLAN FOR 2026 • This audit programme is comprehensive and thus covers 8.81-8.8.2 (a-e)		
	• Is there a predetermined schedule for IA's?	• Yes, AUI001 Internal audit plan 2026		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Is there a procedure for IA's?	• Yes, ASMF 8.7 Internal Audit procedure		
	• Does the IA program address the management system, technical and production activities?	• **Yes,** Please see the audit programme doc AUI001-INTERNAL AUDIT PLAN FOR 2026		
	• Are the IA's conducted by trained and qualified personnel?	• Yes, ASMF 8.7 Internal Audit procedure that "Azuresky management shall be responsible for scheduling and performing laboratory audits."		
	• Are the auditors independent of the activity being audited?	Yes, ASDP001 section 4.6 states that "The criteria for an independent person to conduct an audit, a competent experienced person in the field will be consulted to conduct the internal audit however this will be dependent on the availability of funds. Operations Manager will conduct the audit in case of financial constraint."		
8.7.2	When audit findings cast doubt on the effectiveness of operations, integrity of the RM, or correctness of documentation, are customers who may be affected notified?	• The statement has been included as a requirement to be adhered to as per ASDP001 section 4.6		
8.7.3	• Are audit findings and corrective actions recorded?	• Yes; ASMF 8.7.3 Improvement action request is used to state the objective evidence of finding and corrective action.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Are the actions discharged in appropriate manner and within agreed timescales?	• ASMF 8.7 Internal Audit procedure section "conducting internal audit" **require that** Major NC shall be addressed within 24 hours; Minor NC shall be addressed within 72 hours		
8.7.4	Is there follow up to verify & record the implementation and effectiveness of corrective action?	• **Yes,** ASMF 8.7 Internal Audit procedure section "conduct follow-up audit and close audit"		
8.8	**Actions to address risks and opportunities**			
8.8.1	Has the RMP considered risks and opportunities to: -			
a)	Assure that the management systems intended results can be achieved?	Yes, the whole system was assessed to identify the risks from sample preparation until to packing and shipping both PT and RM with the rationale to enhance intended results or achieve intended results or achieve improvement or prevent undesirable effects 8.5 Actions to address the risks and opportunities		
b)	Enhance desirable effects?	Yes, the whole system was assessed to identify the risks from sample preparation until to packing and shipping both PT and RM with the rationale to enhance intended results or achieve intended results or achieve improvement or prevent undesirable effects 8.5 Actions to address the risks and opportunities		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
c)	Prevent, or reduce undesirable effects?	Yes, the whole system was assessed to identify the risks from sample preparation until to packing and shipping both PT and RM with the rationale to enhance intended results or achieve intended results or achieve improvement or prevent undesirable effects 8.5 Actions to address the risks and opportunities		
d)	Achieve improvement?	Yes, the whole system was assessed to identify the risks from sample preparation until to packing and shipping both PT and RM with the rationale to enhance intended results or achieve intended results or achieve improvement or prevent undesirable effects 8.5 Actions to address the risks and opportunities		
8.8.2	Has the organization taken action to: -			
a)	Address the risks and opportunities?	• Yes, causes, existing controls of causes, consequences, existing controls of consequences all were identified. 8.5 Actions to address the risks and opportunities		
b)	Integrate and implement actions into the management system processes?	• Yes (management of safety of personnel) for example, the personnel are now using respirators rather than dusk mask during sample preparation to mitigate against high dust levels of sample prep area.		
c)	Evaluate the effectiveness of these actions?	• **Yes,** 8.5 Actions to address the risks and opportunities evaluation of effectiveness is there.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.8.3	Are the actions taken in proportion to potential impact?	• Yes, causes, existing controls of causes, consequences, existing controls of consequences all were identified. 8.5 Actions to address the risks and opportunities		
8.9	**Corrective actions (CA)**			
8.9.1	**General**			
	• Has the RMP established policies and procedures for implementing corrective actions?	• ASMF 8.7 Internal Audit procedure section "conducting internal audit" **require that** Major NC shall be addressed within 24 hours; Minor NC shall be addressed within 72 hours • **The Quality Policy statement has recorded that "** Identify risks of the production line, assess and document opportunities or mitigation. Ensure actions are taken proportionate to the impact of product quality.		
	• Are appropriate authorities designated for implementing corrective actions?	• Azuresky is a growing company, management team (Operations Manager, Laboratory Manager and Laboratory Supervisor) shall implement corrective actions.		
	• Is corrective action taken when non-conforming work or RM's are identified?	• Yes; ASMF 8.7.3 Improvement action request is used to state the objective evidence of finding and corrective action. • Customers who may be affected by the non-conforming work or RM shall be notified.		
8.9.2	**Cause Analysis**			

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Does Is the corrective action process begin with an investigation to identify the root cause?	• Yes; ASMF 8.7.3 Improvement action request is used to state the objective evidence of finding and corrective action**. The document requires the investigation to identify the root cause.**		
	• Does the investigation cover in-house production, and subcontractors work?	• Yes, it is imperative to identify the root cause of the NC which may be in-house or with subcontractor. This is to effectively apply appropriate corrective action and reduce the impact to the management system.		
8.9.3	**Selection and implementation of corrective action**			
8.9.3.1	Does the RMP, from potential corrective actions, identify and implement the action most likely to eliminate and prevent recurrence of the problem?	• Yes; ASMF 8.7.3 Improvement action request is used to state the objective evidence of finding and corrective action**. The document requires the investigation to identify the root cause. Through the best analysis tools CAB will be able to select and implement action most likely to eliminate and prevent recurrence of the problem.**		
8.9.3.2	Is the corrective action taken appropriate to magnitude and risks encountered?	• **Through the best analysis tools CAB will be able to select and implement action most likely to eliminate and prevent recurrence of the problem.**		
8.9.3.3	Has the RMP documented and amended operations procedures where required?	• Yes, operations improved to prevent mixing of samples (an NC occurred and recorded)		
8.9.4	**Monitoring of corrective actions**			
	After implementation has the RMP monitored the results to ensure the effectiveness of the CA?	• An internal audit will test if corrective actions are still in implementation by personnel and their effectiveness.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.9.5	**Additional audits**			
	Has the RPM arrangements to conduct additional audits of appropriate areas whenever non-conformities or departures place doubt in their policies, procedures and compliance with this international standard.	• Azuresky, is applying to get the business accredited to ISO 17034:2016, this process test, assess and validate that management meet the requirements of ISO 17034:2016		
8.10	**Improvement**			
8.10.1	Does the RMP continuously improve the effectiveness of the management system, through the use of the quality policy and objectives, audit, corrective & preventative actions and the management review?	• Yes all of these are in place.		
8.10.2	• Does the RMP identify improvements and potential sources of non-conformities?	• Yes, NC are reported and procedures are in place to be implemented to effect management system compliant to ISO 17034:2016		
	• When improvements are required, have action plans been developed, implemented and monitored?	• Yes, currently the business is growing, and identification of improvement areas is the order of the day and it is one of the objectives to get the business accredited to ISO 17025, ISO 17034 and ISO 17043.		
8.10.3	Has the RMP monitored the results after implementation of improvements?	• An implementation of the internal audit schedule will test this.		
8.11	**Feedback from customers**			
	• Does the RMP seek feedback from its customers (both positive and negative)?	• Yes customer feedback form was developed and implemented. It seeks both negative and positive feedback from customers.		

©SANAS

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	• Is the feedback used and analysed to improve the management system, RM production and customer service?	Yes it's in the best interest of the business to use the feedback and increased efficiency, improved customer satisfaction, and better regulatory compliance. One of our clients shared an NC recorded of our PT report during their audit, the investigation of the NC improved the Pt report content.		

CLAUSE	ISO/IEC 17034:2016 REQUIREMENTS	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
Additional / General Comments *This space may also be used to expand on comments in specific sections*				
Team Leader's Signature:			**Date:**	

©SANAS