sanas South African National Accreditation System	**MANAGEMENT REQUIREMENTS OF ISO/IEC 17025:2017**	Organisation's SANAS No/s.

Date/s of evaluation	
Assessor/s & Observers	
Laboratory	**Azuresky (Pty) LTD**
Area / field of operation	**Analytical laboratory of coal**
Laboratory Representative	**Nonhlanhla Msibi**

This report covers the following:

Document Review only		Implementation on Site Visit only		Document Review and Site Visit		Other	

Compliance = C, Non-compliance = NC

REQUIREMENTS & COMMENTS. ***Compliance = C, Non-compliance = NC***

NB1: *References to ISO/IEC 17025:2017 are in italics & indicated as Std. The order of assessment need not follow the order of the checklist. Assessors are expected to know & have the standard, this worksheet is designed as guidance to prompt detailed recording of the process.*

REFER TO ISO/IEC 17025:2017 FOR DETAIL AND FOR CLARIFICATION NOTES.

NOTE 1: *For **CAB's comments**: The CAB must provide information on how requirements have been addressed, documented and/or implemented. Make reference to policies / procedures, incl. clause numbers.*

NOTE 2: *For **Assessor's Comments**: The Assessor must provide information on the CAB's conformity with the requirements*

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
4	**GENERAL REQUIREMENTS**			
4 . 1	**Impartiality:** *How are the following addressed/implemented?*			
4.1.1	Laboratory activities undertaken impartially, structured and managed to safeguard impartiality.	• Maintain clear job descriptions and reporting lines to prevent undue influence. • Include impartiality as part of the onboarding for new employees. • There's impartiality policy ASMF 4.2  4.2. IMPARTIALITY POLICY.pdf implemented. • There's conflict of interest declaration ASMF 4.2.1 completed  4.2. 1 Conflict of Interest Declaration.p by personnel		
4.1.2	Management's commitment to impartiality.	• Impartiality policy in place, that has a mission statement on how management commits to safeguard impartiality. See ASMF4.2. • Quality Policy has stated that Top Management is committed to impartiality. Impartiality policy has been developed.  8.2. Quality Policy Statement.pdf		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
4.1.3	Responsibility of the laboratory to its impartiality and to ensure personnel are free form commercial, financial or other pressures that may compromise impartiality.	Within the impartiality policy document under Commitment to remaining autonomous, it reads as follows: To ensure independence and objectivity and protect impartiality risks within us as an organization and from any outside influence, the following needs to be enforced: • Continuous mandatory disclosure of conflicts of interests • Separate commercial functions from technical decision-making, ensuring that financial pressures do not affect scientific integrity. • Disclosure of incentives from clients and registering them in the incentive register form (This is to protect personnel from any undue pressure in the organisations' activities.)		
4.1.4	Identification of risks to impartiality on an ongoing basis, including those that arise from its relationships.	The lab has assessed the risks and consequences associated with impartiality as outlined by doc ASRI 4.2-001- Risk assessment to impartiality Risk assessment to impartiality.pdf		
4.1.5	If a risk is identified, the demonstration by the laboratory how it eliminates or minimise such risk.	There are existing controls within the ASRI 4.2-001- Risk assessment to impartiality document which highlight how each of the risks identified will be addressed should the need arise.		
4. 2	**Confidentiality**			

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS How are the following addressed / implemented	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
4.2.1	Responsibility through a legally enforceable commitment for all information.	**NDA signed by all personnel see doc 4.3.1 B NON DICLOSURE CONTRACT** 4.3.1 B NON DICLOSURE CONTRAC		
	The laboratory inform customer in writing of any information it intends to place in the public domain, except for information that the customer makes publicly available.	**Covered in section 5.5. of the 4.3. Confidentiality procedure.** 4.3. CONFIDENTIALITY.pdf		
4.2.2	Laboratory inform the customer if it is required by law to release confidential information, unless prohibited by law.	**Covered in section 5.5. of the 4.3. Confidentiality procedure.**		
4.2.3	Confidentiality of information about the customer from sources other than the customer shall be treated as such, unless agreed by the source.	**Covered in section 5.5. of the 4.3. Confidentiality procedure.**		
4.2.4	All personnel i.e. committee members, contractors, personnel of external bodies or individuals acting on the laboratory's behalf, must maintain confidentiality.	• All Azuresky personnel signed NDA's • Visitors or contractors or sub-contractors register their attendance in the visitor's register book which has included a statement for them to manage all information obtained at Azuresky confidentially.		
5	**STRUCTURAL REQUIREMENTS**			
5.1	Legal responsibility	• The company is a legally registered entity, details of registration can be found under ASMF5 section 1		

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
5.2	Identify management that has the overall responsibility for the laboratory.	• Responsibilities have been outlined under section 2 of the structural requirements, see doc 5 Structural Requirements • There are also job descriptions outlining the responsibilities/duties of managerial roles		
5.3	Define and document the range of laboratory activities.	5.3. Statement of conformity		
	Does the laboratory claim conformity with the standard for the range of laboratory activities, excluding externally provided lab activities on an ongoing basis.	The lab has outlined the scope of activities to which it wants accreditation for, and has also listed outsourced functions see doc 5.3. Statement of conformity		
5.4	Requirements of the standard, customer, regulatory authorities and organization providing recognition.			
	Permanent facilities, site away from permanent facilities, associated temporary or mobile facility or customer's facility.	The lab has permanent facilities in which it performs all tests (excluding subcontracted, as outlined by doc 5.3).		
5.5	Define structure, its place in the parent organisation and relationship between management, technical and support services.			
	Responsibility, authority and interrelationship of personnel who manage, perform or verify work affecting the results of laboratory activities.	• Organogram outlining the line of reporting (hierarchy) • Job descriptions outlining scope of authority as well as duties/responsibilities. • Impartiality policy in place to ensure conformity.		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	Document procedures to the extent necessary to ensure consistent application of lab activities.	• The lab does have procedures in place to conduct day-to-day activities		
5.6	Personnel with the authority and resources to carry out their duties:			
	Implement, maintain and improve the management system.	• Lab Manager, Lab supervisor, and Operations Manager responsible for documentation of QMS, improving it, and implementation,		
	Identify deviation from the management system.	• Non-conforming work and corrective action procedure documented, see doc 7.17. Management of non-conforming work		
	Initiation of actions to prevent or minimize deviations.	• Improvement Action Request doc 8.7.3. used to initiate actions to prevent deviations, and identify NC work, to prevent future reoccurrence. • There are also other forms used in the root cause analysis, internal audits. Forms such as AS001MF, AS003MF , AS004MF		
	Reporting to management the performance of the management system or any improvements.	• Management review meeting to be conducted at least once per year, as outlined in 8.6Management Review Procedure		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
5.7	Communication by laboratory management regarding the effectiveness of the management system.	• Any changes or improvements to the management system is communicated to personnel vis memorandums and/or WhatsApp group. 5.6. Azuresky memorandum template • 8.2. Quality Policy Statement also outlines the review of management system through auditing the system periodically to provide confidence in its effectiveness.		
	Laboratory management ensures the integrity of the system when changes are implemented.	• Any changes or improvements to the management system is communicated to personnel vis memorandums and/or WhatsApp group. 5.6. Azuresky memorandum template		
6	**RESOURCE REQUIREMENTS**			
6.6	**Externally provided products and services**			
6.6.1	Suitable externally provided products and services that affect the laboratory are used.	• ASMF6.3. document is a checklist on what to look out for when buying, hiring, or outsourcing something that affects the quality of results or validity of test results. • A list of externally provided services as well as the suppliers ASMF 4.1 List of outsourced services		
6.6.2	Procedure and records for:			

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	Defining, reviewing and approving lab requirements for products and services	• Document ASMF6.2. defines the procedure for selecting, evaluating, approving, and monitoring of subcontractors.		
	Defining criteria for evaluation, selection, monitoring of performance and re-evaluation.	• Document ASMF6.2. also outlines each of these points.		
	Externally provided products and services conform to lab's established requirements.	• Checklist in place, see doc 6.2 ASMF		
	Taking action from evaluations, monitoring of performance and re-evaluations.	• An Early warning notification ASMF 6.2B is implemented to warn of slip ups by the contractor, this is to take action to prevent the ultimate negative impact. • Apply 8.7.3 Improvement action request		
6.6.3	Communication by the laboratory its requirements to external providers:			
	Products and services to be provided;	• Document ASMF6.2. mandates communication of this requirement. See page 2.		
	Acceptance criteria;	• Checklist in place, see doc 6.2 ASMF, which will be approved by the operations manager after review. • Document ASMF6.2. mandates communication of this requirement. See page 2.		
	Competence including required qualification;	• Document ASMF6.2. mandates communication of this requirement. See page 2.		

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
	Activities (either lab or customer) intents to perform at the external providers premises.	• Document ASMF6.2. mandates communication of this requirement. See page 2.		
7	PROCESS REQUIREMENTS *How are the following addressed/implemented?*			
7.1	Review of request, tenders and contracts			
7.1.1	Procedure for the review of request, tenders and contracts.			
a)	Requirements adequately defined, documented and understood.	• 4.1.2. Provides the requirements/criteria (Checklist) of acceptance for contracts , tenders, and customer requests • 4.7.1. Document outlines the requirements under section 3.2. Initial review.		
b)	Capability and resources to meet the requirements.	• 4.7.1. Document outlines the capability assessment under section 3.3.		
c)	Advice customer and gain approval for activities performed by external provider.	• 4.7.1. Document explains that any unclear or conflicting specifications shall be clarified in writing with the customer under section 3.4. • Contract review register 4.1.4 document to keep track of progress		
d)	Selected methods are capable of meeting customer requirements.	• Covered in the review of tenders and contracts docs 4.7.1, 4.1.2, and 4.1.4.		

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.1.2	Customer informed as to the method chosen.	• Customer to be informed in writing as stipulated in 4.7.1.		
7.1.3	When the statement of conformity is requested, clearly define the decision rule.	• Statement of conformity will be shown is CRM certificate, see CRM 002 certificate for example.		
7.1.4	Differences resolved before work commences.	• Customer to be informed in writing as stipulated in 4.7.1. • There is also a deviations or comments section together with approval/acceptance by client before work commences under 4.1.4.		
7.1.5	Customer informed of any deviations from the contract.	• Covered as stated in 7.1.4. above		
7.1.6	Amendment after work commenced, review repeated, communicated to affected personnel.	• Azuresky has not yet entered into a legal agreement to provide service or receive service.		
7.1.7	Cooperation with customer in clarifying customer requests and monitoring the labs performance.	• Azuresky allows clients to come and review their own work/activities performed by the lab (i.e. audit the lab). • These requirements can also be included within the contracts that AzureSky gets into with its clients.		
7.1.8	Records maintained: reviews; significant changes; pertinent discussions with customer.	• Records to be stored for two accreditation cycles.		
7.9	Complaints			

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.9.1	Documented process: receive, evaluate and make decisions.	• Covered by document 7.18. Complaint Handling Procedure.		
7.9.2	Description of the handing process – available to any interested party on request.	• Yes, the procedure is available upon request.		
	Confirm whether the complaint relates to lab activities.	• Covered under 7.18. document, under procedure steps.		
7.9.3	Process for handling complaints: process for receiving, validating, investigation and decision on action; tracking and recording; appropriate action taken.	• Covered under 7.18. document, under procedure steps.		
7.9.4	Responsibility of the laboratory for gathering and verifying all information.	• Covered under 7.18. document, under procedure steps.		
7.9.5	Whenever possible, acknowledge receipt of the complaint, provide progress reports and the outcome.	• Covered under 7.18. document, under procedure steps.		
7.9.6	Outcome reviewed by individuals not involved in original laboratory activities.	• Covered under 7.18. document, under procedure steps.		
7.9.7	Whenever possible, give formal notice at the end of complaint.	• Covered under 7.18. document, under procedure steps.		
7.10	Nonconforming work			

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.10.1	Procedure for activities that do not conform to procedures or requirements.	• 7.17.Management of Non-conforming work		
a)	Responsibility and authorities.	• Addressed under 7.17. document section 3.		
b)	Action based upon risk levels established by the laboratory.	• Risks matrix present doc 8.5. Risk Matrix		
c)	Evaluation of the significance of the NC.	• Risks matrix present doc 8.5. Risk Matrix		
d)	Decision taken on acceptability of the non-conforming work.	• Risks matrix present doc 8.5. Risk Matrix		
e)	Customer is notified and work re-called, where necessary.	• 7.17.Management of Non-conforming work under the procedure section.		
f)	Defined responsibility for authorizing the resumption of work.	• Responsibilities defined, and final decision lies with the operations manager as stipulated in 7.17.Management of Non-conforming work		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.10.2	Records of nonconforming work and actions.	• NC work register see 7.10. • Use SANAS F 03-09/ AS 003MF document • Root cause analysis AS001MF • AS002MF list of possible root causes • AS004MF Brainstorming form		
7.10.3	Corrective action promptly followed, where evaluation indicates possible recurrence or doubt about compliance of the laboratory's operations with its own policies/procedures.	• Make a follow up as per SANAS F03-09		
7.11	**Control of data and information management**			
7.11.1	Access to the data and information needed.	• ISO Standards available • Access to equipment and facilities such as QMWARELIMS 12		
7.11.2	LIMS: Validated for functionality. Changes such as laboratory software configuration or modification to commercial off-the shelf software – Authorised, documented and validated before implementation.	• Yes , ASMF 7.11 Data integrity risk management and preventive maintenance procedure states that "The functioning of the Laboratory Information Management System shall be validated for proper functioning of interface before implementation. Formulas, calculations and all required entries/input data is able or activated to collect, process, record, store and report." Evidence is available for perusal at the laboratory.		
7.11.3	Laboratory information management system (LIMS)			

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
a)	Protection from unauthorised access	• Access to the laboratory server (data drive) is restricted to management personnel only. All other staff members have limited user permissions preventing unauthorized access, modification, or deletion of data. Password protection and access control are in place, and user rights are reviewed periodically.		
b)	Safeguarding from tempering and loss	• All files on the server are backed up to an external hard drive on a monthly basis. Access to data and information is limited to management to prevent tampering. Backup integrity is verified regularly to ensure data recoverability.		
c)	Operated in environment that complies with specifications.	• The server and computer systems are located in a fixed, secure area. The environmental conditions of the analysis room and storage areas (temperature, humidity, and dust levels) are monitored and maintained within defined limits to prevent data corruption and equipment malfunction.		

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
d)	Maintained to ensure the integrity of the data and information	• The information management system and related hardware are maintained according to the manufacturer's specifications. Regular system maintenance, software updates, antivirus protection, and integrity checks are conducted to ensure data accuracy and reliability. Any identified risks to data integrity are recorded and mitigated through preventive maintenance schedules		
e)	Record system failures and appropriate immediate and corrective actions.	• Any server, computer, or software malfunction is immediately reported to management. The incident is logged in the non-conformance register, including details of the failure, immediate containment actions taken, and corrective/preventive measures implemented. The system is not returned to operational use until functionality and data integrity are verified and approved by management.		
7.11.4	LIMS managed and maintained off-site or through an external provider, the provider /operator must comply with the std.	• LIMS managed and maintained by an external provider. Also provides support as required.		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
7.11.5	Instructions, manuals and reference data relevant to LIMS are readily available to personnel.	• There are no user manuals for LIMS, instead training is provided, and competency assessments are done to ensure personnel can operate it • The procedures on how to conduct each of the tests analysis are available upon request/		
7.11.6	Calculations and data transfers subject to appropriate and systematic checks.	• LIMS Validation is done bi-annually, these calculations are verified independently by a second analyst		
8	**MANAGEMENT SYSTEM REQUIREMENTS**			
8.1	**General**			
8.1.1	Establish, document, implement and maintain a management system			
	Option A	• The CAB has adopted Option A and established a documented management system consistent with ISO/IEC 17025:2017. Policies, objectives, procedures and records form the backbone of this system • ASDP001 Quality Manual, 8.2 Quality Statement, Scope of work		
	Option B	**N/A**		
8.2	**Management system documentation**			
8.2.1	Establish, document and maintain policies and objectives.	• The CAB maintains documented quality and operational policies, including competence, impartiality and quality objectives aligned with its accreditation scope.		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.2.2	Address competence, impartiality and consistent operation of the laboratory in policies and objectives.	• The policy explicitly commits to ensuring competent personnel, maintaining impartiality in all laboratory operations, and sustaining consistent processes across tests. • Quality Statement(8.2),		
8.2.3	Provide evidence of commitment to the development and implementation of the management system.	• Senior management signs off on the quality policy and allocates resources (staff, infrastructure, training) to support continuous system improvement. • Work instructions, checklists were developed, authorised and implemented. Competence records are available. Equipment maintenance records are available, etc. • The CAB was established in April 2025, the first management review will be held in January 2026; thereafter, annually.		
8.2.4	All documentation, processes, systems, records; included, referenced or linked to the management system.	• All laboratory procedures, work instructions, forms, systems (including LIMS), and records are either part of, or referenced in, the management system. Document index and links ensure traceability(see doc ASDP001 Quality Manual)		
8.2.5	Access to the parts of the management system documents and related information by all personnel.	• Personnel have access (electronic and physical) to relevant portions of the management system. Controlled shared folders or intranet access ensure up-to-date versions are available at points of use.		
8.3	**Control of management system documents**			

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.3.1	Control of all internal and external documents.	• The CAB maintains document control over both internally generated and externally sourced documents (e.g. standards, method references). External documents are evaluated, approved, versioned, and integrated into the system.		
8.3.2	Laboratory ensures:	• Documents are uniquely identified and approved by authorized personnel (e.g. Operations Manager). Periodic reviews take place as and when necessary. Changes and revisions are clearly marked on revision tables integrated within technical documents and SOP's. Current versions are distributed to points of use, and obsolete versions are stored in folders marked "Obsolete" to prevent unintended use. This applies to all sections a-f below see doc 7.16 Control of quality and technical records		
a)	Approval of documents by authorised personnel.			
b)	Periodic review and updating of documents.			
c)	Changes and current revision status identified.			
d)	Relevant revisions of documents are available at points of use and distribution is controlled.			

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
e)	Unique identification.			
f)	Obsolete documents are assured against unintended use and retained suitably marked.			
8.4	**Control of Records**			
8.4.1	Establish and retain records.	• All records (technical, quality, training, audit, corrective) are created, stored, and retained in accordance with defined retention times and policies, see doc 7.16 Control of quality and technical records , section 7, Retention period.		
8.4.2	Implementation of controls for: identification, storage, protection, back up, archive, retrieve, retention time and disposal of records. For the period consistent with contractual obligations.	• Records are uniquely identified (e.g. numbering, metadata). Storage is secure (locked cabinets, restricted electronic access). Backups of electronic records are done quarterly. Archives are to be done at the end of each year, retrievable, and disposed of in line with contractual/legislative requirements • Storage, backup, archiving of technical documents will be in accordance to doc 7.16 Control of quality and technical records		
8.5	**Action to address risks and opportunities**			

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.5.1	Consider risks and opportunity associated with the laboratory activities to: give assurance that the system achieves its intended results; Enhance opportunities to achieve the purpose and objectives of the laboratory; Prevent, or reduce, undesired impacts and potential failures and archive improvement.	• The CAB conducts risk assessments (technical, procedural, resource) to anticipate impacts on result validity, client service, and system performance, see doc 8.5. Actions to address risks and opportunities		
8.5.2	Plan actions to address risks and opportunities, how to integrate and implement the action into the system and evaluate effectiveness of actions.	• Mitigation and opportunity plans are incorporated into the system with controls to risks and consequences and each having an assigned personnel to implement and review effectiveness of corrective actions, also covered in doc 8.5. Actions to address risks and opportunities		
8.5.3	Action taken proportional to the potential impact on validity of lab results.	• Actions (e.g. calibrations, verifications) are scaled to cover criticality of each process, equipment calibrations performed annually since its covered by metrologically audited service provider and verified weekly to address criticality, inhouse calibration as and when necessary(CV).		
8.6	**Improvement**			
8.6.1 Note Std	Identify, select and implement opportunities for improvement.	• Continuous improvement opportunities are identified via audits, corrective actions, performance indicators, customer feedback, and management reviews, then implemented appropriately.		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.6.2 Note Std	Getting feedback from customer. Analysing and using feedback to improve the management system, laboratory activities and customer service.	• Client feedback is formally collected and reviewed periodically (biannually); actionable changes are fed back into system updates and process modifications. • There is also a survey form AS006MF which is used for reviews by customers.		
8.7	**Corrective action**			
8.7.1	When a non-conformity occurs:			
a)	React to a non-conformity (take action to control and correct, address the consequences)	• When a nonconformity (NC) is identified; whether through internal audit findings, customer complaints, proficiency testing failures, equipment breakdowns, or deviations from approved procedures; the responsible personnel immediately record it using a Nonconformity Register (NCR) form. Immediate containment and control actions are taken to prevent further impact (e.g., isolating affected samples, suspending analysis, or withholding reports). Any potential effects on previously released data are evaluated, and clients are notified if validity may be affected. • Documentation used: AS007MF, 8.7.3.		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
b)	Evaluate the need to eliminate the cause.	• Root cause analysis (RCA) is conducted using structured techniques such as 5 Whys, Ishikawa (fishbone) diagrams, or other systematic problem-solving tools. The investigation considers technical, procedural, equipment, and human factors to determine the underlying cause and assess whether additional actions are required to prevent recurrence. • Forms used: AS001MF, AS002MF, AS003MF, SANAS F 03-09,		
c)	Implement action	• Based on the RCA findings, corrective actions are implemented. These actions may include procedure revision, staff retraining, equipment recalibration or maintenance, or process redesign. • Responsibilities, target dates, and verification steps are documented in the corrective action log.		
d)	Review effectiveness of corrective action taken.	• Effectiveness is evaluated after implementation to confirm that the corrective action eliminated the root cause and did not introduce new risks. Follow-up verification may include internal audit checks, review of subsequent results, or trend analysis of similar incidents. • Effectiveness reviews are signed off by the Operations Manager or designated responsible person.		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
e)	Update risks and opportunities, if necessary.	• When applicable, the laboratory updates its Risk and Opportunity Register to reflect lessons learned from the nonconformity. Changes in risk level or new preventive measures are recorded, ensuring integration into the broader risk management framework. • Docs: 8.5. Actions to address risks and opportunities		
f)	Make changes to the management system, if necessary.	• If a nonconformity reveals systemic weaknesses, the Operations Manager ensures that related procedures, SOPs, or forms are revised accordingly. Revisions are controlled through the document control process (Clause 8.3), reviewed, approved, and communicated to all affected personnel. Where appropriate, changes are reviewed during management review meetings to ensure ongoing system suitability and effectiveness.		
8.7.2	Appropriate corrective actions.	• Corrective actions are carefully chosen to address root causes rather than symptoms, and documented with responsibilities and timelines, implementation outlined in doc SANAS F 03-09		
8.7.3	Retain records (nature of NC, cause(s), subsequent actions taken and results of any corrective action)			
8.8	**Internal audits**			

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS *How are the following addressed / implemented*	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
8.8.1	At planned intervals conduct internal audits. Verify that systems conform to lab's requirements and the standard and it is implemented.	• Audit of the system is planned for Nov-Dec 2026. This will be the very first comprehensive audit program, given that the company has been business for ~7 months. Please see the audit programme doc AUI001-INTERNAL AUDIT PLAN FOR 2026 • This audit programme is comprehensive and thus covers 8.81-8.8.2 (a-e)		
8.8.2 a)	Plan, establish, implement and maintain audit programme.			
b)	Define for each audit: criteria and scope	• ASMF 8.7 Internal Audit procedure, was implemented and require that Lead auditor (who may be independent expect or management team member) shall define the criteria and scope of internal audit.		
c)	Results of the audit reported to relevant management.	• ASMF 8.7 Internal Audit procedure, States that the audit findings shall be shared with all personnel in the organisation. • ASDP001 management system document require that audit finding be also shared at management review meeting.		
d)	Implement corrective actions without undue delay.	• ASMF 8.7 Internal Audit procedure section "conducting internal audit" **require that** Major NC shall be addressed within 24 hours; Minor NC shall be addressed within 72 hours		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS How are the following addressed / implemented	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
e)	Audit records.	• ASMF 8.7 Internal Audit procedure states that the Lead Auditor shall write and share the audit report within 12 working days. • ASMF 8.7.3 Improvement action request is used to state the objective evidence of finding and corrective action.		
8.9	**Management reviews**			
8.9.1	Review laboratory's management system at planned intervals to ensure their continuing suitability, adequacy and effectiveness.	• Given that the lab was recently established, there has not yet been a management review meeting yet. However, the lab has a procedure in place on how it will go about doing it's management reviews and what sections of the standard to look at. The details of the review will be geared towards all the information outlined in 8.9.2. (a-f) and as stipulated by internal management review procedure doc ASMF8.6		
8.9.2	Review of: changes in internal and external issues relevant to the laboratory; fulfilment of objectives; suitability of policies and procedures; status of actions from previous management reviews; outcome of recent internal audits; corrective actions; assessments by external bodies; changes in volume & type of work; customer and personnel feedback, complaints, improvements; resources, results of risk id, outcomes of assurance of the validity of results, monitoring activities and training.	• Given that the lab was recently established, there has not yet been a management review meeting yet. However, the lab has a procedure in place on how it will go about doing it's management reviews and what sections of the standard to look at. The details of the review will be geared towards all the information outlined in 8.9.2. (a-f) and as stipulated by internal management review procedure doc ASMF8.6		
8.9.3	Record all decisions and actions: the effectiveness of the management system, improvement of the lab activities, provision of requires resources and any need for change.	• To be done during the first management review		

©SANAS

CLAUSE	ISO/IEC 17025:2017 REQUIREMENTS How are the following addressed / implemented	CAB's COMMENTS	C/ NC	ASSESSOR's COMMENTS
Additional / General Comments *This space may also be used to expand on comments in specific sections*				

©SANAS

Additional / General Comments *This space may also be used to expand on comments in specific sections*

Team Leader's Signature:		**Date:**	

©SANAS