AzureSky		
AS023TF	**ROOM TEMPERATURE**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

LABORATORY RELATIVE HUMIDITY AND TEMPERATURE RECORDING

Date: _____ Location: Storage Room

				TIMES READINGS TAKEN	READINGS TAKEN BY
Temp. (^{0}C)	SPECIFICATIONS: 16^{0}C-20^{0}C	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25%-45%	ACTUAL READING			
Temp. (^{0}C)	SPECIFICATIONS: 16^{0}C-20^{0}C	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25%-45%	ACTUAL READING			
Temp. (^{0}C)	SPECIFICATIONS: 16^{0}C-20^{0}C	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25%-45%	ACTUAL READING			

AzureSky		
AS023TF	**ROOM TEMPERATURE**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

*N/B: READINGS TO BE **RECORDED ON MONDAY AND FRDAY EVERY WEEK**, HOWEVER, PERSONNEL ARE TO CHECK VISUALLY THE ENVIRONMENTAL CONDITIONS FOR COMPLIANCE AT LEAST THREE TIMES PER SHIFT AT INTERVALS. IF OUT OF SPEC REPORT TO SUPERVISOR*

Shift Supervisor sign:_____

Revision table

Content changed	Reason	Date
Humidity changed from 40-60% to 25-45%	Initial humidity values were provisional, practical work revealed that the initial conditions could not be fulfilled.	29/07/2025

Document no	Revision Status	Effective Date	Revision date
AS023TF	00	04/2025	04/2028