RISK ASSESSMENT TO ASSES RISK AND OPPORTUNITIES

Prepared By: Top Management team (S. Nyambi, T. Maponya & N Msibi)	Date compited: 19.09.2025	Template review date: 19.08.2029	Rev: 00	Doc number: ASRI 8.5-001

Identified Risk	Causes of the risk	Existing controls	Responsible person	What are the consequences	Existing controls	Responsible person	Tasks to improve control effectiveness
1. Uncalibrated equipment	1. Not adehering to the shcedule of calibration	1. Plan the calibration ahead of time, and ensure that it is adhered to.	Operations Manager	Non-conformance to ISO standards, since there will not be metrological traceability.	Decomission the instrument and put an "out of order" sign so that no one uses it when the calibration date has expired	S Nyambi	
	2. Lack of funds to calibrate	2. Sub-contract parameters for which the equipment's calibration certificate has expired.	Operations Manager	Affect the validity of results (ISO 17025 Clause 7,7)			
		3. In the case of drying ovens, the alternative would be to air dry the samples.	Laboratory Supervisor				
2. Unlabelled samples collected/received by the lab	Collection samples from the site without labels	Carry tags and cable ties to the site, label all bags before loading them onto the truck.	N Msibi	Sample mix-up (cross contamination)		S Nyambi	
				Lack of traceability to the original source of sample.			
3. The use of unverified equipment	Unavailability of samples to verify equipment	There is always surplus of coal samples to aid in equipment verification	N Msibi	Affects the validity of results	The supervision is in control	N Msibi	
	Negligence	There is supervision in control to curb negligence		Affects the effectiveness of stage crushing	The supervision is in control		
				The difference with bringing a sample that is greater than -212µm will cause the the results of the analysis to flactuate especially for CV, Iso volatile and moisture			
4. Employees getting sick from the lack of ventilation/extraction fan	No extraction fan	Household fan will be purchased	N Msibi	Employees health deteriorates	Mask	N Msibi	
	Masks do not properly fit	Management commits to purchase the respirators		High exposure to dust	Mask		
5, Backpain due to heavy lifting	Heavy lifting	Ergonomics course to be provided to employees	N Msibi	Leads to long-term backpain problems	Continuous awareness training	N Msibi	
	Lack of steel trolley						

The identified list of risks were considered risks to:
- Assure that the management systems intended results can be achieved
- Enhance desirable effects
- Prevent, or reduce undesirable effects
- Achieve improvement