F 03-09

sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.:

Organisation Name: Azurestiv Pty Ltd

NC Report Number: NC015/2026

Area/Scope Assessed: Documentation (Corrective action)

Assessor / Inspector / Expert: IFSO Nyambi

PART 1: DETAILED OBSERVATION

The laboratory procedure for control of Non-conforming work doesn't adequately define how the effectiveness of corrective actions is reviewed & verified. Timelines for effectiveness review are not specified, particularly where immediate verification is not practical, and no distinction is made b/w minor & major NCs; Procedure doesn't clearly require changes to the management system where necessary to prevent recurrence. In addition, corrective action records, including root causes, actions taken, and results of effectiveness verifications are not explicitly required to be retained. This limits the lab's ability to demonstrate that corrective actions are effective & sustained.

Signed (Representative): Kmadeha

Signed (Team Leader / Inspector): [signature]

MAJOR: ☒ MINOR: ☐

Reference No. of relevant Guide / Standard / Requirement: ISO 17025 (SANAS F48-07 Clause 8.7)

Date: 11/02/2026

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

Due to ~~the~~ incomplete system documentation & processes not defined, the procedure for non-conforming work failed to address how corrective action effectiveness are revised/verified; Detail requirements for changes in management system as well as record retention

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supporting evidence of implementation)

The management of non-conforming work (ASMF 7.17) was updated to include the omitted requirements

Date: 12/02/2026

Signed (Representative): Kmadeha

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Addressed.

Date: 12/02/26

Signed: Assessor / Inspector/Expert: [signature]

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:


©SANAS

AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: NC015/2026 - T



.....

SUPERVISOR

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: NC015/2026 - T

Brainstorming team
Keitumetse Maleka; Nosipho N Hlongwane; Abonathi Jama & Tshepo Maponya.

Problem Statement: Lab procedure for non-conforming work doesn't: adequately define how the effectiveness of corrective actions is reviewed & verified; detail required changes to management system where necessary & explicitly state the retention of records

Why 1: why did the lab's procedure for NC work not specify the revision & verification of corrective action effectiveness. detail the requirement to change the management system where necessary nor retain records of NCs?

Although the procedure for NC work existed, it failed to address how corrective action effectiveness are revised & verified. The reason for the 'omission' stems from incomplete system documentation/process not defined.

Why 2: why were there processes not defined?

The lab was working on establishing documents which meant that it can only identify gaps within its QMS ~~was~~ through audits, observations & production.

Supervisor:

8. Corrective Actions

Once root cause is confirmed:

- Define the correction and corrective action
- Assign responsible personnel and target dates
- Validate effectiveness after implementation
- Document updates to SOPs, training, or process controls if needed
- Where necessary:
  - Ensure notification of customers and recall CRM's/RM's and certificates and documentation within an appropriate period. This also means that any tested samples which can have a significant impact on the operations of customers shall be relayed to the client.
  - PT reports already sent to participants are recalled or disregarded.

9. Follow-Up and Closure

- Verify the corrective action eliminated or mitigated the issue
- Retest or recharacterize RM/CRM/sample batch if necessary
- Document closure with review sign-off from operations/lab manager and resume work.
- Update NCW register and archive all records securely
- A planned Internal Audit must include the checking or evaluate the corrective action taken, this it ensures effectiveness of the implementation

CONTROLLED DOC

10. Record Retention

- NCW records must be stored for **minimum of 10 years**, linked to batch records.
- Include them in management review for trend analysis.

CONTROLLED DOC

- Planned date for effectiveness review
- Results of effectiveness review
- Date of closure

7. Evaluation and Risk Assessment

Assess whether the non-conformance:
- Affects previously released PT or RMs/CRMs
- Affects previously released reports/work/samples tested
- Requires rework, retesting, or full recall
- Risks invalidating assigned property values or traceability
- Has occurred before (recurring issue)

Use a risk matrix to evaluate severity and urgency.

8. Corrective Actions

Once root cause is confirmed:
- Define the correction and corrective action
- Assign responsible personnel and target dates
- Validate effectiveness after implementation
- Document updates to SOPs, training, or process controls if needed

ADDED:
- Where necessary, changes to the management system (procedures, work instructions, risk registers, or controls) shall be implemented to prevent recurrence of the non-conformance.
- Appropriate corrective actions shall be proportional to the risk and impact of the non-conforming work.
- All corrective actions shall be traceable to the identified root cause.

- Where necessary:

- Ensure notification of customers and recall CRM's/RM's and certificates and documentation within an appropriate period. This also means that any tested samples which can have a significant impact on the operations of customers shall be relayed to the client.
- PT reports already sent to participants are recalled or disregarded.

9. Follow-Up and Closure

• Verify the corrective action eliminated or mitigated the issue
• Retest or recharacterize RM/CRM/sample batch if necessary
• Document closure with review sign-off from operations/lab manager and resume work.
• Update NCW register and archive all records securely
• A planned Internal Audit must include the checking or evaluate the corrective action taken, this it ensures effectiveness of the implementation

ADDED:
• Effectiveness of corrective actions shall be evaluated after a suitable period of implementation to confirm that the non-conformance has not recurred and that process performance has improved.
• Effectiveness shall be verified by review of subsequent test results, repeat stability or homogeneity testing (where applicable), audit findings, or trend analysis of NCW records.
• The outcome of the effectiveness review shall be recorded on the Improvement Action Request form prior to formal closure of the NCW.

10. Timelines

• Major NCs shall be closed within 24 hours from the time of detection.
• Minor NCs shall be closed within 7 days from the time of detection.

ADDED:
• Effectiveness of corrective actions for **Major NCs** shall be reviewed within **1 to 3 months**, depending on the nature and risk of the non-conformance.

Effective date	Rev status	Next review date
02/2026	03	02/2029

• Effectiveness of corrective actions for **Minor NCs** shall be reviewed within **3 to 6 months** after implementation.
• Where justified by the nature of the non-conformance (e.g. long-term stability, trend monitoring), extended review periods may be defined and documented.

11. Record Retention

• NCW records must be stored for minimum of 10 years, linked to batch records.
• Include them in management review for trend analysis.

ADDED:
• Records shall include the nature of the non-conformance, root cause(s), corrective actions taken, planned and actual effectiveness review dates, and results of effectiveness verification.
• Corrective action records shall be available for verification during internal audits and external assessments.

12. Revision

Revision no	Content changed	Reason for the change	Date
02	Update the timelines for the closing of NCs.	NC raised by SANAS	06/02/2026
03	No content changed, just additions to address SANAS findings. (Please see sections 6,8,9,10,11 for the updates- look for "added" subheadings).	To address NC's raised by SANAS.	11/02/2026

Effective date	Rev status	Next review date
02/2026	03	02/2029