NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**	NC Report Number: **AB03**
Area/Scope Assessed: **Management Requirements**	Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

Although a risk assessment to impartiality was done on 19/08/2025,at the time of the assessment it was observed that the risks were not evaluated and graded based on a risk scale where the likelihood and consequences give a scale for each risk

Signed (Representative):

Signed (Team Leader / Inspector): A.Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 4.1.5**

Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

Date:

Signed (Representative):

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS