NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**	NC Report Number: **AB07**
Area/Scope Assessed: **Management Requirements**	Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

At the time of the assessment, it was observed that Document ASFM 7.17 point 7 states that for all NCs, a risk matrix must be used to evaluate and assess the risk of the NC. The NCW register only has major and minor ratings for all NCs.

Signed (Representative):

Signed (Team Leader / Inspector): A. Bougart

MAJOR: [X] MINOR: [ ]	Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 8.7.1 e**	Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

Date:

Signed (Representative):

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS