

	AzureSky		
	AS002TF	Pulveriser Efficiency	
	Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
	Approved by:	**Nonhlanhla Msibi- Operations Manager**	

PULVERIZER EFFICIENCY AND DUST LOSS TESTS

Date: _____ Equipment no: _____

Name of operator/analyst doing the test: _____

Supervisor's signature: _____

Initial sample mass (g): _____
Mass of pulverized sample (g): _____
Mass of oversized material after screening through a 212μm screen (g)_____
Mass of -212μm material after screening through a 212μm screen (g)_____

CALCULATIONS

% Efficiency = $\frac{Mass\ of\ undersize}{Pulverised\ mass} X100$ **= _____ %**

Note: The efficiency should be ≥ 95%.

DUST LOSS

Mass of sample before pulverizing (g): _____
Mass of sample after pulverizing (g)_____

% Dust loss = $\frac{Mass\ of\ sample\ before\ pulverising - mass\ of\ sample\ after\ pulverisng}{Pulverised\ massMass\ of\ sample\ before\ pulverising} X100$ **= _____**

Note: Dust loss ≤ 5%

Document no	Revision Status	Effective Date	Revision date
AS002TF	00	04/2025	04/2028

DUPLICATE RESULTS IF THE INITIAL TEST WENT OUT OF SPECS

DATE: _____ EQUIPMENT NO: _____

NAME OF OPERATOR/ANALYST DOING THE TEST: _____

SUPERVISOR'S SIGNATURE: _____

Initial sample mass (g): _____
Mass of pulverized sample (g): _____
Mass of oversized material after screening through a 212μm screen (g)_____
Mass of -212μm material after screening through a 212μm screen (g)_____

CALCULATIONS

% Efficiency = $\frac{Mass\ of\ undersize}{Pulverised\ mass} X100$ **= _____ %**

Note: The efficiency should be ≥ 95%.

DUST LOSS

Mass of sample before pulverizing (g): _____
Mass of sample after pulverizing (g)_____

% Dust loss = $\frac{Mass\ of\ sample\ before\ pulverising - mass\ of\ sample\ after\ pulverisng}{Pulverised\ massMass\ of\ sample\ before\ pulverising} X100$ **= _____**

Note: Dust loss ≤ 5%

COMMENTS ON ACTION TAKEN:

_____
_____
_____
_____
_____
_____

NB: If the Pulveriser efficiency is still out of spec notify the lab supervmanager.

Document no	Revision Status	Effective Date	Revision date
AS002TF	00	04/2025	04/2028