MICRON LABORATORY SERVICES

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

9 Beatty Avenue, Witbank 1035 Tel: 013 690 1532 Fax: 013 656 1000
Postnet Suite 139 Private Bag X 7260 Witbank 1035
Reg No: 2017/365305/07 Vat nr: 4430102030
www.micronlab.co.za

CERTIFICATE OF CALIBRATION

Certificate no:	AZU F1.2
Date of calibration:	27.02.2026
Date of next calibration:	March 2027
Customer:	AzureSky (Pty) Ltd Unit 3, 17 Markiel Road Wadeville Germiston
Location:	Laboratory
Contact Information	nonhlanhla@azuresky.it.com

Conditions of Issue:

This certificate is issued in accordance with the conditions of accreditation granted by the South African National Accreditation System of South Africa. Subject to the uncertainties specified, this is a correct record of measurements made.

Copyright of this certificate is owned jointly by SANAS and the issuing laboratory. This certificate may not be reproduced other than in full without written approval of SANAS and of the issuing laboratory.

The values in this certificate were correct at the time of calibration. Subsequently the accuracy will depend on such factors as, the care of handling, the use of the instrument and the frequency of use. Re-calibrations should be performed after a period which has been chosen to ensure that the instrument's accuracy remains within the desired limits.

Where statements of conformity are made in the certificate, the following rules are applied:
Within Tolerance: Results ± expanded uncertainty within limits / specification.
Out of Tolerance: Results ± expanded uncertainty exceeds limits / specification.
Indeterminate: Results ± expanded uncertainty overlaps limits / specification.

Measurement results relate only to the item calibrated.

Technical Signatory: J Kruger

Calibrated by: F Rossouw

MICRON LABORATORY SERVICES

t +27 (0)13 690 1532 | f +27 (0)13 656 1000 | e admin@mlabs.co.za | 9 Beatty Avenue, Witbank, 1035
Postnet suite 139, Private bag X7260, Witbank, 1035

sanas Calibration Laboratory
CAL 103-03-00
CAL 103-14-00

CK No: 2017/365305/07

CERTIFICATE OF CALIBRATION

Customer:	AzureSky (Pty) Ltd	**Certificate Number:**	AZU F1.2
Contact person:	nonhlanhla@azuresky.it.com	**Instrument:**	Darvan Brenco Electronic Platform Sc
Location:	Laboratory	**Serial Number:**	120330150
Date of Calibration:	27.02.2026	**Company Number:**	N.N
Date of Next Calibration:	March 2027	**Capacity:** **Kg**	150
Lab Temperature Start of Cal ±2°C:	25.3	**Readability:** **2**	0.01
Lab Temperature End of Cal ±2°C:	25.3	**Condition:**	Good

Calibration procedure: The instrument has been calibrated with standard weights of known value and assumed density of 7100 kg/m³ traceable to the national standard in accordance with our procedure No: P1.2.WI

Traceability of standards Certificate No: SET 5 Calibration expiry date: 30 April 2026

Calibration results: Range calibrated: 100 Results indicated in: Kg

1. **Repeatability test:**
   The standard deviation was found to be 0.01 at a load of 50
2. **Eccentric test:**
   The maximum deviation caused by eccentric loading was 0.00000 at a load of 25
3. **Discrimination (sensitivity) test:**
   A weight of 10 g was added to a load of 50 and caused the indication to change by 0.01 Kg
4. **Accuracy and linearity:**

Before adjustment			After adjustment		
Nominal load	Reading	Error	Nominal load	Reading	Error
0	0.00	0.00	0	0.00	0.00
20	20.00	0.00	20	20.00	0.00
40	40.00	0.00	40	40.00	0.00
60	60.00	0.00	60	60.00	0.00
80	80.00	0.00	80	80.00	0.00
100	100.00	0.00	100	100.00	0.00

Uncertainty: Uncertainty of measurement is: ± 0.03 Kg

5. **Additional calibration points:** Additional calibration points requested

Before adjustment			After adjustment		
Nominal load	Reading	Error	Nominal load	Reading	Error
5	5.00	0.00	5	5.00	0.00
10	10.00	0.00	10	10.00	0.00

Uncertainty: Uncertainty of measurement is: ± 0.03 Kg

The reported uncertainties of measurement are estimated and expressed in accordance with the GUM document and is stated as the calculated standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Remarks: Adjustment method: External calibration
Range calibrated : Partial Calibration: 0-100kg

Authorized by (Technical Signatory) J Kruger

Calibrated by (Metrologist) F Rossouw

End of Certificate