	AzurESky	
AS019.2TF	BALANCE VERIFICATION	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

DAILY BALANCE VERIFICATION FORM

Equipment number:	AS00A31									
Month:	**1.0000 g**	**5.0000 g**	**10.0000g**	**20.0000g**	**50.0000g**	**100.0000 g**	**5.00 kg**	**Done by**	**Supervisor signature**	**Comments**
Date	**± 0.0002g**	**± 0.0003g**	**±0.0005g**	**± 0.0006g**	**±0.0005g**	**± 0.00010g**	**+0.05**			

Document no	Revision Status	Effective Date	Next revision date
AS019.2TF	00	04/2025	04/2028

	AzurESky	
AS019.2TF	**BALANCE VERIFICATION**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

ACTION

- **Tolerance limits exceeded 3 times successively**
- **Action limit exceeded**

Document no	Revision Status	Effective Date	Next revision date
AS019.2TF	00	04/2025	04/2028