# NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**

NC Report Number: **AB09**

Area/Scope Assessed: **Management Requirements**

Assessor / Inspector / Expert: **A. Bougart**

## PART 1: DETAILED OBSERVATION

At the time of the assessment, it was observed that According to ASDP001 point 4.1 states that when a document has been amended a revision table needs to be completed and the content of the change identified, but at the time of the assessment document ASMF 7.17 was amended and no changes documented from version 0 to version 1.

Signed (Representative):

Signed (Team Leader / Inspector): A.Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 8.3.2c**

Date: **08/04/2026**

## PART 2: ROOT CAUSE ANALYSIS

*(Provide an analysis of the extent and cause of the nonconformity)*

## PART 3: CORRECTIVE ACTION TAKEN

*(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)*

Date:

Signed (Representative):

## PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

*(Comment on whether the corrective action is sufficient and appropriate)*

Date:

Signed: Assessor / Inspector/Expert:

## PART 5: FOLLOW-UP AT NEXT ASSESSMENT

*(Comment on the implementation of the corrective actions)*

Date:

Signed: Assessor / Inspector/Expert:


©SANAS