sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**

NC Report Number: **AB11**

Area/Scope Assessed: **Management Requirements**

Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

At the time of the assessment, it was observed that the following agenda points were not discussed in the MRM: outcomes of assurance of the validity of results, monitoring activities and training.

Signed (Representative):

Signed (Team Leader / Inspector): A. Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 8.9.2**

Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

Non-adherence to the MRM procedure, attributed to the lab conducting its first MRM.

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

A memo was issued to all relevant personnel as a reminder of the required agenda points for MRM. A subsequent MRM was conducted ~~were~~ wherein all outstanding agenda points including monitoring activities and training were discussed & documented accordingly.

Date: 22/05/2026

Signed (Representative): Mdeha

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: ABII....

Brainstorming team

Tshepo Maponya

Sifiso Nyambi

Keitumetse Maleka

Problem Statement: During the time of assessment, it was observed that monitoring activities and training were not discussed during the management review meeting despite being listed as required agenda points & the outcomes of assurance of the validity of results, whilst discussed, were not documented under the correct heading, as required by ISO 17025 clause 8.9.2.

Why 1: why were monitoring activities and training not discussed during the MRM despite being listed?

The agenda points for monitoring activities and training were overlooked during the MRM and were therefore not discussed.

Why 2: why were the agenda point overlooked during the meeting?

As this was the laboratory's first MRM, there was insufficient familiarity with all the required agenda points as stipulated by ISO 17025 8.9.2, resulting in monitoring activities and training being ~~not~~ unintentionally omitted from the discussion

Supervisor:


AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: ABII

During the time of assessment, it was observed that monitoring activities & training were not discussed during the management review meeting despite being listed as a required agenda point & the outcome of the assurance of the ~~used~~ validity of results, was discussed under INCORRECT heading



.....

SUPERVISOR

Additional Space for NC's

IRQ no: RB11

It was further noted that the outcomes of assurance of the validity of results were discussed during the first MRM and documented under agenda point 2 - Changes in internal and external issues relevant to PT & RM activities, and were therefore present in the meeting minutes at the time of assessment. Going forward, this item will be discussed & documented under agenda point 5 in accordance with the MR procedure FSMF 8.6.

Supervisor:

5. Inputs

The management review shall include, but is not limited to, the following agenda items:

1. Status of actions from previous management reviews.
2. Changes in internal and external issues relevant to the management system (e.g. revised regulations, SANAS requirements).
3. Results of internal and external audits (including nonconforming work and corrective and preventive actions).
4. Complaints, feedback, and customer satisfaction.
5. Monitoring and measurement of results, including homogeneity and stability studies, as well assuring the validity of result.
6. Nonconforming work, corrective and preventive actions.
7. Assessment of risks and opportunities.
8. Performance of suppliers and subcontractors.
9. Adequacy of resources (staff, training, equipment, facilities).
10. Effectiveness of actions to address risks and opportunities.
11. Opportunities for improvement.

ADDED (to address missing minimum agenda points):

12. Fulfilment of laboratory objectives and targets.
13. Changes in the volume and type of work undertaken by the laboratory.
14. Suitability and effectiveness of policies, procedures, and work instructions.
15. Results of assessments by external bodies (e.g. SANAS, regulators, PT providers, customers).

Effective date	Rev status	Next review date
02/2026	01	02/2029

AzurESky	**Document No.** AS008MF	**Page 1 of 1**
	Effective Date	**April 2026**
	Revision Date	**April 2029**

Title:	Memorandum	
Document by:	Azuresky Pty Ltd	**Revision No**
Compiled by:	Sifiso Nyambi & Tshepo Maponya	**00**
Authorizing Officer: N Msibi	**Signature:**	

Department: Management

Memorandum no: 014

Position: Lab Analyst

Date: 18 May 2026

Title: Management Review Meeting - Required agenda points

Please be advised that in accordance with the MR procedure ASMF 8.6, monitoring activities and training are to be discussed under agenda point or input 9 - Adequacy of resources (staff training, equipment & facilities) during all Management review meetings

AzurESky	ATTENDANCE REGISTER		
	Document no:	Version:	Page:
	ASMF 17043-8.4.1	0	1
	Compiled by:	Sifiso Nyambi - Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	
Date:			
Chair:		Scribe	

Subject:	Management review meeting - required agenda point
Absent:	

Present	Cell Number	Designation	Signature	Email address
NN Mlangeni	0723732558	Lab Analyst		nosipho@Azuresky.it.com
LD Photo	0766923263	Lab Analyst		lilo@azuresky.it.com
K. Maleka	071 724 9031	Lab analyst		kreitumetsc@azuresky.it.com
N. Msibi	0827868149	Ops Manager		nonhlanhla@azuresky.it.com
A. Jama	060 6111 584	Lab analyst		abenathi@azuresky.it.com
S.B. Nyambi	0721741852	Lab Analyst		sifiso@azuresky.it.com

AzurESky	MANAGEMENT REVIEW MEETING		
	Document no: ASMRM 8		
	Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

Date: 21 May 2026 (continued 22,26 May 2026)

Time: 09:54 - 16:03 (11:00 – 16:30)

Location: AzurESky Office

Attendees

- NP Msibi (Operations Manager/Chair)
- SB Nyambi (Lab Manager)
- TMJ Maponya (Lab Supervisor)
- K. Maleka (Lab analyst)

Minutes (Key discussions)

Status of Actions from Previous Management Review

The following action items from the previous meeting were reviewed:

- Item 1
  - Water drainage: Short-term solution of draining into a bucket remains in place. Long-term solution (purchasing appropriate equipment) remains open until Mar 2027
- Item 2
  - Personnel are required to check respirator filters condition and report if not in good condition. Partially addressed and ongoing assessment of personnel.
- Item 3
  - New gloves: Due date 21/05/2026.
- Item 4
  - Internal PPE procurement (main lab): Onboard = 2 PPE; yearly = 1 PPE.
  - Onsite lab: Onboard = 2 PPE; yearly = 1 PPE

- Build PPE stock inventory – for onsite lab.
- PPE define – wait for mine (Vryheid) requirement.
- Decision: Wait until mine starts up, then buy PPE for both main and onsite lab. 2 standard PPE & 1 Freezer Jacket.
- Also obtain safety pack from mine.
- NB: Due date: pending until conclusion of contract with Vryheid mine.

- Item 5
  - Stability monitoring reporting frequency: Bimonthly for CV and Vol and monthly for IM.

- Item 6
  - Crucibles: Issue resolved. Spares were brought.
  - Request for crucible tray: two months from today.
  - Spares requirements Ongoing - To be monitored.
  -

- Item 7
  - Lab supervisor suggested adding onto submission form that late submission (later than 25th of every month) will result in the forfeiting of money paid and no participation for that month. Due date: 30 June 2026

- Item 8
  - Closed: Effective.

- Item 9
  - Closed: Effective.

- Item 10
  - Discussed that disciplinary procedure is to be drafted by next week Friday (29 May 2026).
  - After internal review, an external review is required.

Effective date	Rev status	Next review date
05/2026	00	05/2029

- Item 11
  - Addressed and effective.
- Item 12
  - Closed and addressed.
- Item 13
  - Currently 30 open NCs
  - Update NC register by the 29 May 2026.
- Item 14
  - To be discussed under Point 13 on new agenda.
- Item 15
  - Addressed.
- Item 16
  - Risk register review open until 29 May 2029

New Agenda discussion

Changes in Internal and External issues relevant to the management system.

- Several changes were documented resulting from the recent SANAS document review and site visit audit such as the Internal procedure and Non-Conforming Works
- Extended scope of work with PSD & TM. TM was included in the accreditation scope.
- Potential extension of work - on-site lab (Vryheid mine).
- Review periodically - Regulation of tax: EMP501 due 31 May 2026; letter of good standing and BBBEE certificate were updated and in order. These should be updated May of every year.

- Safety: Expiration of medicals for N Msibi, S Nyambi and K Maleka, these are to be renewed immediately when executing the safely file for Vryheid mine. Due date July 2026

Results of Internal and External Audits

- External audit performed by SANAS in April 2026: The existing system was deemed adequate; however, its implementation was found to be lacking, which resulted in COF status.
- Six NCs still pending corrective actions and they should be sent to SANAS by 22 May – 26th May 2026.
- Recent internal audit (May): No formal NCs were recorded.

Complaints, Feedback and Customer Satisfaction.

NB: Personnel feedback to be addressed as well

- Noko Analytical services complaint: An NC (NC No. 022/2026) was opened, investigated and addressed.
- Six feedback forms were brought back. (It was noted that the feedback forms returned by the six laboratories were sourced from the main laboratories only and therefore represent not only the main laboratories but also their respective onsite laboratories.)
- Overall evaluation of feedback: 4.8/5 (under quantitative analysis – positive reviews only). As per the criteria, this round has been deemed a successful evaluation.
- Internal personnel feedback: A grievance process needs to be set up to enable personnel to lodge grievances and the overall issues handled during the year can be discussed at the next MRM.

Monitoring and measurement of results (Performance of parameters), including homogeneity and stability studies

- Bias was detected on Vol, IM and CV. An investigation was done on all parameters and new control plotted after discovery.
- Quick-Ash: Potential negative bias from PT participation (trending low).

- Environmental conditions: it's winter and temperatures have dropped, early in the morning A/C settings can be set to warmer settings. On-going monitoring is conducted and recorded on monitoring sheets.
- Stability monitoring: Update required 26 May 2026.
- Check this month's PT to ensure it follows stability verification.
- 0.5g mass piece purchased.
- Pulveriser, crushers, homogeneity drum, Furnaces, Sieves verification are all in good condition and working order.

Non-conforming work, corrective and preventive actions

- Currently 30 open NCs (external).
- All internal NCs closed: follow-up has been done/addressed/verified.

Assessment of risks and opportunities

After reviewing the risk assessment to impartiality. The following risk were marked as to monitor:

- ***Risk***: Clients don't stick to timeline of submission of results. ***Control***: Sending reminders to clients by the 20th every month.
- ***Financial risk:*** Financial dependence from a single client. ***Control***: MD to set up a fund to combat such a risk from emerging. Ongoing monitoring.
- ***Dual roles risk:*** same person is doing the test and approving the results. ***Control***: results are checked by TS who did not compile the report. Mitigated.
- ***Dual roles risk:*** Lack of personnel. ***Control:*** No existing control however daily until mitigation measure is available.
- ***Methodological Risk:*** Poor communication and oversight. ***Control***: any changes to methods/procedures to be relayed via memo.
- ***Remuneration:*** Dissatisfaction with salary package. ***Control:*** MRM will include reviewing of salaries. Open communication lines with Managing Director is implemented to enable a discussion of the issue when they arise.
- ***Internal Audit:*** Self-audit + technical conflict. ***Control:*** Cross-departmental auditing and external peer review. Resource constraints – OP to audit system.

Effective date	Rev status	Next review date
05/2026	00	05/2029

Performance of suppliers and subcontractors

- Current suppliers are in good standing in accordance with doc 6.6.3 criteria and they met internal checklist.
- Linberts will be used for calibration methods since they met internal criteria. Currently using both Linberts and Micron.

Adequacy of resources and training

- Personnel (management) raised several grievances:
  - Workload: Too much for the management team.
  - Lack of resources: No personnel for sample prep, causing management to fill in.
  - Complaint about additional work added to scope (business proposal & ISO 9001).
  - OM made decision based on personal reasons on appointing students.
  - OM said the decision is final and will not change regarding appointed of students.
- Training plan for: safety rep, fire marshals, first aiders for on-site/onboard lab. These are to be implemented when safety file requirements are have been received from Vryheid mine.
- Training matrix review discussed and no changes were made on the existing plan.
- Equipment: Currently adequate; however there is a need to add a Sulphur equipment which shall be planned and executed before Mar 2027.
- Facilities: Overall fine; however, PPE inventory room needs to be added in the future (November 2026).

Assessment/Effectiveness of actions to address risks and opportunities

Risk ID	Identified Risk	Mitigation / Control Actions Taken	Status
1	Uncalibrated equipment	Established and implemented calibration schedules. Validated service provider performance. Calibrated Vernier Caliper and verified against in-house criteria. Procured 0.5g mass piece to cover the full working range.	Mitigated
2	Unlabelled/Improperly received samples	Implemented laboratory intake procedures. Limited sampling trips to the mine have ensured control; all completed trips have maintained sample integrity.	Partially Controlled
3	Use of unverified equipment	Verified equipment for -10mm and -3mm/4.75mm coal fractions. Action: Procurement of raw coal >40mm scheduled for completion by end of June 2026.	Ongoing

	MANAGEMENT REVIEW MEETING	
AzurESky	Document no: ASMRM 8	
Compiled by:	Sifiso Nyambi - Laboratory Manager	
Approved by:	Nonhlanhla Msibi - Operations Manager	

Risk ID	Identified Risk	Mitigation / Control Actions Taken	Status
4	Health risks (lack of ventilation/extraction)	Risk assessed. Mitigation: Provided appropriate respirators as interim control while extraction fan procurement is pending.	Ongoing
5	Back pain due to heavy lifting	Conducted risk assessment (Risk Rating: D3). Implemented engineering controls: Procured steel trolley. Administrative controls: Staff training scheduled for Q4 2026.	Ongoing
6	Insufficient equipment availability	Procurement plan initiated; remaining balance of required equipment to be acquired by year-end 2026.	Ongoing
7	Late deliveries of PT (Proficiency Testing) items	Systemic issues resolved; risk effectively managed and removed from active registry.	Closed
8	Ineffective communication	Standardized communication via WhatsApp group. Analysts provided with clear instructions and deadlines; no deviations	Mitigated

Risk ID	Identified Risk	Mitigation / Control Actions Taken	Status
		reported in analysis quality.	
9	Late deliveries of equipment	Operations normalized; lab now maintains fully effective capability for PT preparation. Risk closed.	Closed

Removed risks: Late deliveries of PT samples. This risk has been removed because the lab's operational strategy to send PTs ahead of schedule has been impeccable for the past 6-8 months. Thus, risk has been fully removed.

Likelihood x Consequence = Risk Score

Fulfilment of Laboratory objectives and targets

- 2026 Objectives have been set using ASMF 1703-8.9.213, they will be rated after six months (bi-annually). Next rating: October 2026.

Changes in the volume and type of Work undertaken by the laboratory

- Testing work business proposal drafted and submitted for Vryheid mine onsite lab management. Commissioning is expected to happen July 2026.
- ISO 9001system documentation required by Malefo mining services. Awaiting for deposit then work will start.

Suitability and Effectiveness of policies, procedures and work instructions

Effective date	Rev status	Next review date
05/2026	00	05/2029

- Under impartiality: Quality policy was amended to include internal audit. To be checked after monitoring period.
- Changes made in prep are effective since they passed audit.
- Working instructions: Changes still effective. A lot of procedures were changed after SANAS – to be checked after COF.

Results of assessment by external bodies

- SANAS found the laboratory is not conforming to its own management system.
- Technical aspect of the CMB was found to be sufficiently implemented, given the number of NCs.
- PT provider: Lab performs well – SANAS was impressed.

	MANAGEMENT REVIEW MEETING	
AzurESky	Document no: ASMRM 8	
Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

Management Review Action Register

All improvement actions identified during this meeting are captured below. Progress shall be reviewed at the next Management Review Meeting.

No.	Action / Improvement Required	Source	Responsible Person	Target Date
1	Air-con: Drain water into bucket & replace with appropriate equipment (long-term: buy appropriate equipment)	Previous MRM action item	NP Msibi	Open until (new air-con) equipment purchased
2	New gloves	Previous MRM action item	NP Msibi	21/05/2026
3	Build PPE stock inventory	Previous MRM / PPE procurement	NP Msibi	Wait for mine startup July 2026
5	Request for crucible tray (2 months from today)	Previous MRM action item	NP Msibi	July 2026
6	Draft disciplinary procedure; internal review then external review	Previous MRM action item	NP Msibi	29/05/2026 (draft); June 2026 (close)

Effective date	Rev status	Next review date
05/2026	00	05/2029

	MANAGEMENT REVIEW MEETING	
AzurESky	Document no: ASMRM 8	
Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

No.	Action / Improvement Required	Source	Responsible Person	Target Date
7	Update NC register – currently 30 open NCs (external);	Internal/External audit results	Keitumetse	29/05/2026
8	Send outstanding 6 NCs to SANAS	External audit	NP Msibi	22 May - 26 May
9	Set up internal personnel feedback/grievance system for next MRM	Complaints & feedback review	NP Msibi	July 2026
10	Establish PPE inventory room	Adequacy of resources	NP Msibi	November 2026
11	Document an Ethics Policy	Risk assessment – management influence	NP Msibi	29/05/2026

Minutes/Record authorised
NP Msibi
Date:27.05.2026

Effective date	Rev status	Next review date
05/2026	00	05/2029

AzurESky	ATTENDANCE REGISTER		
	Document no:	Version:	Page:
	ASMF 17043-8.4.1	0	1
	Compiled by:	Sifiso Nyambi - Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	
Date:	21, 22 & 26 May 2026		
Chair:	N. Msibi	Scribe	K. Maleka
Subject:	MRM		
Absent:			

Present	Cell Number	Designation	Signature	Email address
N. Msibi	082 786 8149	Ops Manager		nonhlanhla@azuresky.it.com
K Maleka	071 724 9031	Lab analyst	K. Maleka	keitumetsem@azuresky.it.com
TMJ Mapanya	076 455 0672	Lab Supervisor		tshepo@azuresky.it.com
SB Nyambi	0721741852	Lab Manager		SIFISO@azuresky.it.com