sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**

NC Report Number: **AB17**

Area/Scope Assessed: **Management Requirements**

Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

At the time of the assessment, it was observed that Audit records used for the internal audit dated 29/01-20/02/2026 were not signed by the auditor i.e SANAS Form F44. Furthermore an uncontrolled form named vertical assessment was used by the auditor and other scrap sheets of paper attached.

Signed (Representative):

Signed (Team Leader / Inspector): A. Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 8.8.2e**

Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

- Administrative oversight.
- Inadequate document hygiene.
- Lack of form standardisation.

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

- Sign SANAS F44 form for audit conducted b/w 29/01-20/02.
- Document clean-up.
- Standardisation of vertical assessments.
- Audit team briefing

Date: 22/04/26

Signed (Representative):

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: AB17

Brainstorming team

Sifiso Nyambi
Stuppo Maponya

Problem Statement It was observed that audit records used for the internal audit dated 29/01 - 20/02 2026 were not signed by the auditor (i.e. SANAS Form F44-11). Furthermore an uncontrolled form named vertical assessment was used by the auditor & other scraps of papers attached.

Why 1? Why were the offical audit records (F44) left unsigned? The audit team concluded the technical assessment but overlooked the final administrative requirment to sign the form.

Why 2? Why was the administrative step overlooked? There was no step for verifying the completeness of records (signed & authorised) before filing the audit pack.

Why 3? Why uncontrolled forms and scrap papers included in official audit records? These were used as temporary working aids during the audit, and there was no protocol to ensure their removal after the data was transferred.

Why 4? Why was there a lack of protocol for document disposal? The distinction between informal working notes and quality controlled records had not been clearly defined within the audit team.

Why 5? Why the administrative lapses occur? Inadequate training on document hygiene and lack of standardised verification process for the finalisation of audit.


Supervisor:

AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: AB17.



.....

SUPERVISOR

AS009MF 04/26

Additional Space for NC's

IRQ no: AB17

Extent of NC: Observation was specific to the audit records generated during the internal audit done b/w 29/01-20/02 2026.

Root cause:

① While the audit activities were performed, a final verification step was missed, resulting in the SANAS F44 records remaining unsigned after audit was concluded.

② Personnel used uncontrolled vertical assessment forms and informal scrap paper as Temporary working aids to navigate the system during the audit process. These were intended as personal guidance notes but were mistakenly filed with the Official QMS records instead of being discarded once data has transferred to formal docs.

③ There was a lack of clarity regarding which specific form should be used for vertical audits, leading to the use of external unapproved templates

Corrective action

① The SANAS F44 form has been signed.

② Remove uncontrolled documents and informal scraps of paper, these have been removed from the file. Audit pack now contains only controlled, signed QMS records.

③ Lab has officially adopted the F44 form as one of the official forms used for internal audits. This ensures uniform approach & prevents the use of uncontrolled external templates

④ A meeting was held with audit team reinforcing "document hygiene" protocols. Team was instructed that informal notes are acceptable to gather data, but only formalised controlled forms like F44 constitute valid audit evidence and are permitted in the final audit file.

Supervisor:

sanas South African National Accreditation System	**VERTICAL ASSESSMENT** **ISO/IEC 17025:2017 LABORATORIES**	Organisation's SANAS No/s.

Date/s of Assessment	29/01 – 20/02 2026
Assessor/s & Observers	Nkiso Nyambi
Laboratory	Mzuretsky Pty Ltd.
Area / field of operation	Testing
Laboratory Representative	Keitumetse Maleka, Lulo Rabe, Nosiplo Hlongwane of Akwatek Jama

Certificate or Report *(Select one or more final Report /Certificate(s). Record at least the number, date and the accredited parameters, as on the Scope of Accreditation, measured. Include a short description of calibrations/tests performed,*

AZ PT 001

Name of analyst / metrologist *and* technical signatory

Nkiso Nyambi

Clause	**REQUIREMENTS AND COMMENTS** *Compliance = C, Non-compliance = NC, Not applicable = NA* ***NB1: Indicate WHAT has been checked and HOW requirements have been implemented. The order of assessment need not follow the order of the checklist. Assessors are expected to know & have the standard, this checklist is designed as guidance to prompt detailed recording of the process.*** ***NB2*: Where a clause is marked as NA, reason must be provided as to why it's not applicable*** ***REFER TO ISO/IEC 17025:2017 FOR DETAIL AND FOR CLARIFICATION NOTES.***	**C NC NA***
7.5 / 7.11	**Technical Records** *(state which data and calculations were checked)*	
7.5.1	Records include the date and the identity of personnel responsible	C
7.5.1	Records contain sufficient information	C
7.5.2	Amendments to technical records traceable to previous versions or to original observations	C
7.5.2	Original, amended data and files are kept, including the date of alteration and person responsible	C
7.11.1	Access to the data and information needed to perform laboratory activities	C
7.11.2	LIMS validated for functionality before introduction	C
7.11.6	Calculations and data transfers checked in an appropriate and systematic manner	C
6	**Personnel - Training** *Check training records of identified operator(s) for accredited work from selected Results*	
6.2	Operator/s identified as competent for the work and is proof of competence available	C
6.2.2	Personnel competence documented for each activity	C
6.2.3	Personnel competent for the work and able to evaluate the significance of deviations	C
6.2.4	Duties, responsibility and authority communited to personnel	C
6.2.5	Records retained for: training; supervision; authorisation; monitoring of competence	C
6.2.6	Personnel authorized to perform specific function	C

Additional Assessor Notes *(This may be used for rough notes as well)*

©SANAS

7.2.1-3/ 7.6	**Performance capability of selected methods**	
7.2.1.1 - 7	Proof of confirmation of proper operation of -standard methods, -laboratory developed methods -non-standard methods	C C C
7.2.2.1 - 4	Methods validated and availability of performance capability Modifications to the validation development plan approved and authorized. Validation records retained (procedure; requirements; performance characteristics; results & fitness for purpose statement)	C C
7.6.1	Contributions to measurement uncertainty identified	C
7.6.2	Calibration laboratories (including in-house calibration) - evaluate the measurement uncertainty for all calibrations	C
7.6.3 ***Notes Std***	Testing laboratories - evaluate measurement uncertainty.	C

7.7	**Ensuring the validity of results**	
7.7.1 (a – k)	Indicate how the laboratory monitors results Monitoring data suitably recorded (e.g. control charts), evaluated reviewed.	C
7.7.2	Monitor its performance by comparison with results of other laboratories	C
7.7.3	Monitoring data analysed, used to control and/or improve the laboratory's activities Action taken for incorrect results or results found outside pre-defined criteria	C

6.4 / 6.5	**Calibration of equipment and/or standards used** *(Indicate which equipment/standards)*	
6.4.6	Measuring equipment are calibrated	C
6.4.7	Appropriateness of calibration and verification programmes, cover operating ranges	C
6.4.8	Calibrated equipment are correctly labelled, coded or identified	C
6.4.11	Calibration and reference material data are updated and implemented	C
6.4.13	Records of calibration and verification complete, tolerances appropriate	C
6.5.1	Establish and maintain metrological traceability of its measurement results	C
6.5.2	Traceability to national standards	C
6.5.3	Metrological traceability to an appropriate reference e.g. CRMs, Consensus standards	C
6.4.10	Intermediated checks - reference, primary, transfer and working standards	C

Additional Assessor Notes *(This may be used for rough notes as well)*

©SANAS

6.4	**Equipment maintenance and operation**	
6.4.3 6.4.4 6.4.9 6.4.12	Handling/transport/storage/use to prevent contamination/unintended adjustment/deterioration of - equipment and - std/reference materials	C
6.4.13 a-h	Records for equipment which can influence laboratory activities retained	C

7.8	**Test Report / Certificate**	
7.8.1.1 / R03	Results reviewed and authorized prior to release	C
7.8.1.2	Simplified report issued with the agreement of the customer	C
7.8.2.1	Report/certificate include at least following information, unless the laboratory has valid reasons:	
a)	a title (e.g. "Test Report" or "Calibration Certificate").	C
b)	name and address of the laboratory	C
c)	the location of performance of the laboratory activity.	C
d)	unique identification	C
e)	the name and contact information of the customer	C
f)	identification of the method used	C
g)	a description, unambiguous identification, and, when necessary, the condition of the item	C
h)	the date of receipt of the test or calibration item(s), and the date of sampling, where this is critical to the validity and application of the results;	C
i)	the date(s) of performance of the laboratory activity;	C
j)	the date of issue of the report;	C
k)	reference to the sampling plan and sampling method used by the laboratory or other bodies where these are relevant to the validity or application of the results;	
l)	statement to the effect that the results relate only to the items tested, calibrated or sampled;	C
m)	the results with, where appropriate, the units of measurement;	C
n)	additions to, deviations, or exclusions from the method;	
o)	identification of the person(s) authorizing the report	C
p)	clear identification when results are from external providers	C
7.8.2.2	Data provided by a customer clearly identified	C
	A disclaimer is put on the report when the information is supplied by the customer and can affect the validity of results	C
	If lab is not responsible for the sampling stage, it is stated in the report that the results apply to the sample as received	
R04	Correct use of the SANAS Accreditation Symbol (SANAS R04)	C

The lab is currently not accredited.

7.8.3	**Specific requirements for test reports**	
7.8.3.1	In addition to 7.8.2, do test reports, where necessary for Interpretation of test results, include:	
a)	information on specific test conditions, such as environmental conditions	
b)	where relevant, a statement of conformity with requirements or specifications	
c)	where applicable, the measurement uncertainty presented in the same unit as that of the measurand or in a term relative to the measurand	
d)	where appropriate, opinions and interpretations (see 7.8.7);	
e)	additional information which may be required by specific methods, authorities, customers or groups of customers	
7.8.4	**Specific requirements for calibration certificates**	
7.8.4.1	In addition to the requirements listed in 7.8.2, calibration certificates shall include the following	
a)	the measurement uncertainty of the measurement result presented in the same unit as that of the measurand or in a term relative to the measurand	C
b)	the conditions (e.g. environmental) under which the calibrations were made that have an influence on the measurement results	C
c)	a statement identifying how the measurements are metrologically traceable	C
d)	the results before and after any adjustment or repair, if available	C
e)	where relevant, a statement of conformity with requirements or specifications	
f)	where appropriate, opinions and interpretations	
7.8.4.2	Is the laboratory responsible for sampling? NO.	
7.8.4.3	Calibration certificate or calibration label shall not contain any recommendation on the calibration interval except where this has been agreed with the customer.	
	7.8.4.1 - Applies to certificates given by the calibrating lab.	
7.8.5	**Reporting sampling – specific requirements**	
Where the laboratory is responsible for the sampling activity, in addition to the requirements listed in 7.8.2, reports shall include the following, where necessary for the interpretation of results:		
a)	the date of sampling	C
b)	unique identification of the item or material sampled (including the name of the manufacturer, the model or type of designation and serial numbers as appropriate)	C
c)	the location of sampling, including any diagrams, sketches or photographs	C
d)	a reference to the sampling plan and sampling method	C
e)	details of any environmental conditions during sampling that affect the interpretation of the test results	C
f)	information required to evaluate measurement uncertainty for subsequent testing or calibration	N/A

The lab hasn't sampled any of its samples. However, the above details have been included in the report structure for future endeavours.

Additional Assessor Notes *(This may be used for rough notes as well)*

7.8.6	**Reporting statements of conformity**	
7.8.6.1	Decision rule documented when statement of conformity to a specification or standard is provided	
7.8.6.2	Statement of conformity clearly identifies:	
a)	to which results the statement of conformity applies	
b)	which specifications, standards or parts thereof are met or not met	
c)	the decision rule applied (unless it is inherent in the requested specification or standard).	
7.8.7	**Reporting opinions and interpretations**	
7.8.7.1 Notes Std	Only authorised personnel express opinions and interpretations	
7.8.7.2	Opinions and interpretations expressed are based on the results obtained from the tested or calibrated item and clearly identified	
7.8.7.3	Record of the dialogue with the customer retained	
7.8.8	**Amendments to reports.**	
	Any change of information clearly identified and, where appropriate, the reason for for the change included in the report	C
	Amendments to a report after issue only made in the form of a further document, or data transfer which includes the statement "Amendment to Report, serial number or an equivalent form of wording	C
	A complete new report is uniquely identified and contain a reference to the original that it replaces	C

for 7.8.8. see report AZ PT 006 .

6.3	**Facilities and environmental conditions**	
6.3.1 ***Notes Std***	Suitability of facilities and environmental conditions for the laboratory activities	C
6.3.2	Requirements for facilities and environmental conditions documented	C
6.3.3	Environmental conditions monitored, controlled and recorded	C
6.3.4	Measures for access control, cross-contamination prevention and effective separation implemented; monitored and reviewed periodically	C
6.3.5	Suitability of facilities or sites outside laboratory's permanent control	C

Additional Assessor Notes (This may be used for rough notes as well)

Laboratory representatives signatures:

Keitumetse Maleka
Abonathi Jama

Signed: Technical Assessor			
Signed : Team Leader	Apanya	Date	22/04/2026

Before

Conduct laboratory audit

The audit team shall assess effectiveness of the Quality Management System and compliance to the revelent codes of practice, Laboratory developed methods and Legal Agreements. The audit team shall ensure the audit scope covers technical and/or management system audits.

The Lead Auditor shall on the day of the audit perform the following:

- Conduct the opening meeting
- Introduce the audit team
- Conduct the audit as planned
- Attain all signed off non-conformances (Hardcopy and/or Emailed copy).
- Conduct the closing meeting.
- Issue an audit report to all parties concerned.

Auditors shall on the day of the audit perform the following:

- Conduct the audit as planned
- Record all non-conformances on ASMF 8.7.3 improvement action request
- Classify the non-conformances either as major, minor or observation
- Discuss the outcome of the audit with the Auditee.

Compile report

The report for laboratory audit shall be compiled as follows:

- The Lead Auditor shall compile a draft report.
- The draft report shall be reviewed by any other management team member for correctness and truthfulness of the record.
- The report shall be authorized by any other senior management team member and distribution to all personnel within 12 working days.

The audit report shall provide an accurate and clear record of the audit process and shall include the following:

CONTROLLED DOC

- The Lead Auditor shall generate the scope of the audit and any requirements to the Auditee at least 24 hours prior to the day of the audit or unless otherwise there is urgent need to conduct the audit with immediate effect.

Conduct laboratory audit

The Lead Auditor shall on the day of the audit perform the following:

- Conduct the opening meeting
- Introduce the audit team
- Conduct the audit as planned
- Issue non-conformances when they arise. (Hardcopy and/or Emailed copy).
- Conduct the closing meeting and share the findings.

Auditors shall on the day of the audit perform the following:

- Conduct the audit as planned
- Record all non-conformances on the SANAS F03-09 document.
- Classify the non-conformances either as major, minor or observation
- Discuss the outcome of the audit with the Auditee.
- Some of the documents that can be used during the audit include but not limited to SANAS F44, F48. F49, F90, F91, F66, F176

- The results of the internal audit shall be documented in a comprehensive Internal Audit Results Pack. This pack serves as the official record of the audit and final report to management. The Results Pack consists of:
  - **SANAS Assessment Forms**: Relevant SANAS technical checklist (Including but not limited to F44, F48, F49, F91, F66, and F176) as selected by the auditor based on the scope of the audit, Or any document the auditor deems fit for the scope they are auditing.
  - **Non-Conformances records (NC):** Any non-conformances raised to be documented on the SANAS F03-09

In accordance with ISO 17025 Clause 8.8.2 c, the reporting of results to relevant management is evidenced by the mandatory signatures of the laboratory representative/management on both the F

CONTROLLED DOC

AzurESky	**Document No.** AS008MF	**Page 1 of 1**
	Effective Date	**April 2026**
	Revision Date	**April 2029**

Title:	Memorandum	
Document by:	Azuresky Pty Ltd	**Revision No**
Compiled by:	Sifiso Nyambi & Tshepo Maponya	**00**
Authorizing Officer: N Msibi	**Signature:**	

Department: Management

Memorandum no: 07

Position: Laboratory Manager

Date: 22/04/2026

Title: The use of authorised, controlled & signed documents

Kindly note that the use of personal notes and other scraps of paper is allowed while navigating the audit process. However, only authorised/controlled document are to be used for the audit pack at the end of each assessment. As a rule of thumb, only file controlled documents as proof for all Technical records.

AzurESky	ATTENDANCE REGISTER		
	Document no:	Version:	Page:
	ASMF 17043-8.4.1	0	1
	Compiled by:	Sifiso Nyambi - Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	
Date:	22/04/2026		
Chair:	Sifiso M.	Scribe	
Subject:	Use of authorised, controlled signed docs.		
Absent:			

Present	Cell Number	Designation	Signature	Email address
Sifiso Nyambi	0721741852	Lab Manager		Sifiso@azuresky.it.com
TMJ Mapanya	076 4550692	Lab Supervisor		tshepo@azuresky.it.com
K. Maleka	071 724 9031	Lab analysts		Keitumetse@azuresky.it.com
A. Jama	060 6111 584	Lab analyst		Abdinasir@Azuresky.it.com
N. Msibi	0827681149	Ops. Manager		nonhlanhla@azuresky.it.com