sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: **T1195**

Organisation Name: **Azuresky Laboratory**

NC Report Number: **AB22**

Area/Scope Assessed: **Management Requirements**

Assessor / Inspector / Expert: **A. Bougart**

PART 1: DETAILED OBSERVATION

At the time of the assessment, it was observed that although the lab uses one supplieras a service provider for procurement of instruments, calibrations and servicing of analytical equipment, there is no evidence of any communication regarding the products and services to be provided, the acceptance criteria, competence required or activities that the lab intends to perform at the customer's premises.

Signed (Representative):

Signed (Team Leader / Inspector): A. Bougart

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: **ISO/IEC 17025 requirement 6.6.3**

Date: **08/04/2026**

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

Oversight given the intent of applying the standard subclause correctly by designating a service provider that is accredited for the services the lab requires.

Incomplete system documentation since the process of selection of service provider has not been documented, it was only verbally agreed to by management & the service provider.

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supportig evidence of implementation)

>A checklist & approval of suppliers has been documented and sent to the service provider as agreement between the service provider & the lab that they will be the primary services provider for all listed equipment. ~~signatures~~ TMS, competence requirements house been met by the service provider as stipulated,

>Equipment procedure 6.4 has been ammended to include the fulfillment of the subclause in question.

Date: 08/05/2026

Signed (Representative): Abel Apanja

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: AB22

Brainstorming team

Tshepo Maponya
Sizwe Nyambi
Keitumetse Maleka

What went wrong? It was observed that although the lab uses one suppliers a services provider for procurement of instruments, calibrations and servicing of analytical equipment, there is no evidence of any communication regarding the products and services to be provided, the acceptance criteria, required competence or activities that the lab intends to perform at the customers' premises.

Why was there no communication regarding the products & services to be provided, the acceptance criteria, required competence among the customers & service provider.

> Through previous established engagement, the lab appointed micron automatically as its service provider and failed to follow the right channels as standardized

Why did the lab not follow the right channels in appointing micron?

> The lab automized the appointment of Micron based on previous experience, overlooking the need to follow standard practiced

What caused this deviation?

> Oversight, given the lab had intent of applying the standard correctly by going for a service provider that is accredited for the services the lab requires. Moreover, previous experience shows that indeed this service provider is competent in ~~proving~~ TMS providing the needed services.

Incomplete System documentation.


Supervisor: Maponya.

AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: AB22



.....

SUPERVISOR

BEFORE

	Equipment Management	
AzurESky "ensure quality of test"	Document no: ASMF 6.4 Proc	
Compiled by:	Tshepo Maponya - Laboratory Supervisor	
Approved by:	Nonhlanhla Msibi - Operations Manager	

1. Purpose & Scope

This procedure outlines the selection, identification, calibration, and maintenance of all laboratory equipment. This applies to all "critical equipment"—any hardware or software that can significantly influence the accuracy of laboratory results.

2. Selection and Verification

Before any equipment is placed into service (whether it is brand new or returning from repair), it must be verified.

- **Action:** The lab Manager/Supervisor compares the equipment's performance against the manufacturer's specifications or the laboratory's method requirements.
- **Record:** Identification & Verification results are noted in the **Equipment Register (ASMF 6.3 Equipment Register).**

3. Identification and Labelling

All equipment listed in the register shall be uniquely identified with a dedicated equipment number.

- **Calibration Labels:** Equipment must be labelled with its calibration status, indicating the **date of last calibration** and the **due date for the next calibration,** where necessary.

4. Handling, Transport, and Storage

To prevent contamination, damage, or deterioration:

- Equipment must be used only by authorized personnel.
- Sensitive instruments (e.g., balances) must be protected from vibrations and extreme temperature shifts.
- If equipment is transported outside the lab, it must be re-verified upon return before being used for testing.

AFTER

AzurESky "ensure quality of test"	Equipment Management		
	Document no: ASMF 6.4 Proc		
	Compiled by:	Tshepo Maponya - Laboratory Supervisor	
	Approved by:	Nonhlanhla Msibi - Operations Manager	

1. Purpose & Scope

This procedure outlines the selection, identification, calibration, and maintenance of all laboratory equipment. This applies to all "critical equipment"—any hardware or software that can significantly influence the accuracy of laboratory results.

2. Selection and Verification

Before any equipment is placed into service (whether it is brand new or returning from repair), it must be verified.

- **Action:** The lab Manager/Supervisor compares the equipment's performance against the manufacturer's specifications or the laboratory's method requirements.
- **Record:** Identification & Verification results are noted in the **Equipment Register (ASMF 6.3 Equipment Register).**

3. Mandated Service Provider

To maintain technical consistency and metrological traceability, the Laboratory has appointed Micron Equip as the mandated provider for all equipment procurement, servicing, and calibration as detailed in the External Provider Checklist (Section 6.6.3).

3.1. Communication of Requirements

In accordance with ISO/IEC 17025 Clause 6.6.3, all technical and quality requirements are communicated to the provider via the signed checklist and supplementary purchase orders. This includes:

- **Technical Specifications:** Defined in the equipment table, including specific calibration setpoints and tolerances.
- **Acceptance Criteria:** Verification against the technical requirements of this procedure before equipment is commissioned.
- **Competence:** Mandatory requirement for **SANAS accreditation**, which in turn validates personnel competence as assigned technical signatories to their scope of activities listed in the accreditation certificate.

- **Access Rights:** The Laboratory reserves the right to witness activities at the provider's premises to verify compliance.

3.2. Ongoing Monitoring

The evaluator assesses the service provider annually; if no issues are found, the agreement with the service provider automatically rolls over to the next year. If issues arise during the review, a new service provider shall be evaluated the same way and approved for the services the lab requires when the criteria and competence requirements are met.

4. Identification and Labelling

All equipment listed in the register shall be uniquely identified with a dedicated equipment number.

- **Calibration Labels:** Equipment must be labelled with its calibration status, indicating the **date of last calibration** and the **due date for the next calibration,** where necessary.

4. Handling, Transport, and Storage

To prevent contamination, damage, or deterioration:

- Equipment must be used only by authorized personnel.
- Sensitive instruments (e.g., balances) must be protected from vibrations and extreme temperature shifts.
- If equipment is transported outside the lab, it must be re-verified upon return before being used for testing.

AFTER

9. Revision

Revision no	Content changed	Reason for revision
01	• Added section 8, which controls the use of external equipment. • Added a section that covers the mandating of the equipment service provider (section 3)	Ensure compliance to clause 6.4.2 of ISO 17025 Ensures compliance to subclause 6.6.3 of ISO 17025

AzurESky	**PROVISION OF EQUIPMENT, SERVICES, AND SUPPLIES CHECKLIST**		
	Document no: ASMF6.6.3		
	Compiled by:	**Tshepo Maponya - Laboratory Supervisor**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

Supplier name: Micron Equip

Supplier contact: andile@micronequip.com

Product/service supplied:

The laboratory has designated Micron Equip as the exclusive vendor for the procurement, servicing, and calibration of the equipment outlined in the Table below:

Equipment	Description	Calibration Required?
25mm Crusher	25mm 3-phase jaw crusher, capable of crushing ±50mm coal down to size -25mm. The crusher must have adjustable jaws, have a guard mesh housing all mechanical components, and a ±100mm receiving mouth.	No, efficiency to be conducted by the lab
10mm Crusher	10mm 3-phase jaw crusher, capable of crushing ±25mm coal down to size ±25mm. The crusher must have adjustable jaws, have a guard mesh housing all mechanical components, and a ±50mm receiving mouth	No, efficiency to be conducted by the lab
3mm Crusher	3mm 3-phase jaw crusher, capable of crushing ±10mm coal down to size ±3mm. The crusher must have adjustable jaws, have a guard mesh housing all mechanical components, and a ±20mm receiving mouth.	No, efficiency to be conducted by the lab
Pulverisers	Pulverisers capable of milling coal from -3mm to -212µm. They must either use a cross-beat or hammers, possess a 212µm sieve with a receiver. Both 3-phase & 2-phase are acceptable.	No, efficiency to be conducted by the lab.
Homogenizer	A 3-phase homogenizer capable of processing a 210 L drum of pulverised coal.	No, efficiency done by the lab.
Oven	A 700L-1400L oven capable of reaching temperatures of 110-120 °C with ±2 heating fans.	Yes, at setpoints 40°C & 107°C

Effective date	Rev status	Next review date
05/2026	00	05/2029

AzurESky	**PROVISION OF EQUIPMENT, SERVICES, AND SUPPLIES CHECKLIST**		
	Document no: ASMF6.6.3		
	Compiled by:	**Tshepo Maponya - Laboratory Supervisor**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

Oven	A 50L oven capable of reaching temperatures of ±110°C.	Yes, at setpoint 107°C
Furnace	A muffle furnace capable of reaching temperatures of up to ±815°C	Yes, at setpoints 500°C & 815°C, ramping profile.
Furnace	A muffle furnace capable of reaching temperatures of up to ±900°C.	Yes, at setpoint 900°C
Sieve Screens	- Sieve screens with a 450mm circumference and sizes in mm (50,40,25,22.4,10,6,4.75,3,1,0.5). - Sieve screen with ±100mm circumference and size 212µm. - 450mm circumference receiver. - ±100mm circumference receiver.	No, verification done by the lab.
Drycal Calorimeter System	- Drycal Single Vessel Calorimeter System - Single Vessel Drycal Calorimeter Complete with: 1 x Computer complete 1 x Calorimeter 1x Controller box 2x Bombs 1x Filling station 1x Cooler/ washer 1x Circulation Tank 1x UPS 1x Startup seal kit 1x Operation manual - Water Chiller System Complete - Drycal Oxygen Regulator Complete Fittings & Piping - Drycal Crucibles - Benzoic Acid Tablets, mass at ±0.5g, CV at 26.423 MJ/KG	No, calibration to be done by the lab using Benzoic acid tablets.

Effective date	Rev status	Next review date
05/2026	00	05/2029

CONTROLLED DOC

AzurESky	**PROVISION OF EQUIPMENT, SERVICES, AND SUPPLIES CHECKLIST**	
	Document no: ASMF6.6.3	
	Compiled by:	**Tshepo Maponya - Laboratory Supervisor**
	Approved by:	**Nonhlanhla Msibi - Operations Manager**

Analytical Balance	Capacity: 120g Readability: 0.1mg Pan Size: 80mm x 80mm	Yes
Top Pan Scale	Capacity: 8000g Readability: 0.1g Pan Size: 300mm x 210mm	Yes
Platform Scale	Capacity: 150kg Readability: 10g Pan Size: 400mm x 500mm	Yes
Mass Pieces	Class F1 Weights(100g,50g,20g,10g,5g,1g,0.5g) Mass Piece Cast Iron (20kg, 10kg, 5kg)	Yes
Cone & Quarter Cross	A cone & quarter steel cross at ±7m for both the legs	No
Desiccator	±20L desiccator, with moisture-absorbing silica gel desiccants.	No

Note: Consumables for any equipment purchased at Micron shall also be purchased at Micron Equip.

To be filled by the service provider

I, Andile Simelane on behalf of **Micron Equip** hereby mandate that **Micron Equip** will be able to undertake the Supply, Maintenance & Calibration of the aforementioned equipment.

Signature: _____

Date: 14.05.2025

To Be Filled By Azuresky Pty Ltd

Competence & Acceptance Criteria

Effective date	Rev status	Next review date
05/2026	00	05/2029

CONTROLLED DOC

AzurESky	**PROVISION OF EQUIPMENT, SERVICES, AND SUPPLIES CHECKLIST**		
	Document no: ASMF6.6.3		
	Compiled by:	**Tshepo Maponya - Laboratory Supervisor**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

The overall competence & acceptance criteria required for the purchase & calibration of equipment is a SANAS accreditation (mass & temperature metrology) & ± 5 years of supply, servicing & calibration of equipment.

Remarks: This appointment is based on Micron Equips' demonstrated technical competence and a proven track record of reliability established through previous institutional engagements. Consequently, Micron Equip is recognised as a SANAS-accredited, pre-approved, trusted supplier for all laboratory technical requirements.

Supplier approval

☒ **Approved**
☐ **Conditionally Approved (e.g., for low-risk items or pending documents)**
☐ **Not Approved**

Evaluator Name: Tshepo Maponya

Signature: 
Date: 15/05/2026

CONTROLLED DOC

Effective date	Rev status	Next review date
05/2026	00	05/2029

AzurESky	**Document No.** AS008MF	**Page 1 of 1**
	Effective Date	**April 2026**
	Revision Date	**April 2029**
Title:	Memorandum	
Document by:	Azuresky Pty Ltd	**Revision No**
Compiled by:	Sifiso Nyambi & Tshepo Maponya	**00**
Authorizing Officer: N Msibi	**Signature:**	

Department: Management

Memorandum no: 013

Position: Lab Supervisor

Date: 08/05/2026

Title: Notice of amandment to equipment procedure & introdeceion of Checklist 6.6.3.

Kindly note that procedure 6.4 has been amended to address clause 6.6.3 of ISO 17025 which the lab has mandated Micron Equip for procurement, servicing & calibration of all technical equipment, with the use of checklist 6.6.3.

AzurESky	ATTENDANCE REGISTER		
	Document no:	Version:	Page:
	ASMF 17043-8.4.1	0	1
	Compiled by:	Sifiso Nyambi - Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	
Date:	08/05/2026		
Chair:	TMJ Maponya	Scribe	
Subject:	Introduction of checklist ~~6.6.5~~ 6.6.3 & commencement to procedure 6.4 Equipment TMJ		
Absent:			

Present	Cell Number	Designation	Signature	Email address
N. Msibi	0829868149	Ops Manager		nonhlanhla@azuresky.it.com
TMJ Maponya	0764550692	Lab Supervisor	TMJ Maponya	tshepo@azuresky.it.com
SB Nyambi	0721741852	Lab Manager		sifiso@azuresky.it.com