sanas South African National Accreditation System

NON-CONFORMANCE, CORRECTIVE ACTION AND CLEARANCE REPORT

SANAS No.: T1159

Organisation Name: AZURSKY

NC Report Number: NC07

Area/Scope Assessed: Coal Testing

Assessor / Inspector / Expert: N Chilisa

PART 1: DETAILED OBSERVATION

Although the CAB is participating in the PT for sample preparation and total moisture this is not included in the Lab PT-Plan QM 5-AS001 PT

Signed (Representative): Appanga

Signed (Team Leader / Inspector): AR

MAJOR: [X] MINOR: [ ]

Reference No. of relevant Guide / Standard / Requirement: ISO/IEC 17025 7.7 SANAS R80

Date: 08/04/2026

PART 2: ROOT CAUSE ANALYSIS

(Provide an analysis of the extent and cause of the nonconformity)

Oversight given that the lab was participating in the required PT scheme for the methods, however the CAB forgot to update the PT plan to include methods added after the drafting of the system, resulting in the system not completely document.

PART 3: CORRECTIVE ACTION TAKEN

(Clearly indicate what corrective action was taken, and attach supporting evidence of implementation)

The PT plan was revised to include sample preparation & total moisture, with their frequency of participation included in the plan. Communication of the changes in the plan also attached.

Date: 13/04/2026

Signed (Representative): Appanga

PART 4: CORRECTIVE ACTION VERIFIED AND CLEARED

(Comment on whether the corrective action is sufficient and appropriate)

Date:

Signed: Assessor / Inspector/Expert:

PART 5: FOLLOW-UP AT NEXT ASSESSMENT

(Comment on the implementation of the corrective actions)

Date:

Signed: Assessor / Inspector/Expert:

©SANAS


AS001MF

07/25

ROOT CAUSE ANALYSIS FORM

IRQ NO: NC-07

Although the CAB is participating in PT for sample preparation and total moisture this is not included in the lab PT-Plan QMS-AS001 PT



Mponya

SUPERVISOR

AS004MF 07/25

BRAINSTORMING FORM

IRQ no: NC-07

Brainstorming team

Tshepo Maponya
Sizwiso Nyambi
Nosipho Hlongwane
Mbenabhi Jama

NC / Investigation question: Although the CAB is participating in the PT for sample preparation and total moisture, this is not included in the lab PT Plan QMS-AS001 PT

Why 1: Why was Sample preparation & total moisture not included in the PT plan QMS AS001 PT?

> Oversight given that the lab was participating in the required PT scheme for the method, however the CAB forgot to update the PT plan to included methods added after the drafting of the system, resulting in the system not completely documented.

TMJ

Supervisor: Maponya

AzurESky

BEFORE

Proficiency Testing (PT) Participation Plan (2025 – 2029)

Laboratory Name: Azuresky Pty Ltd

Document No.: QMS-AS001 PT

Revision: 00

Effective Date: 10/10/2025

Next Review: 10/10/2026

Prepared by: Laboratory Manager Signature: _____

Approved by: Operations Manager Signature: _____

1. Objective

To define the laboratory's 5-year strategy for participation in internal and external proficiency testing schemes. The plan ensures continuous verification of analytical competence for coal-related parameters under controlled, low-throughput conditions where PT complements other quality-assurance (QA) measures.

2. Scope

Parameter	Method/Standard
Quick Ash	In-house method (based on ISO 1171)
ISO Ash	ISO 1171
ISO Volatile Matter	ISO 562
ISO Moisture (Inherent Moisture)	ISO 589 / SANS 5925
Calorific Value	ISO 1928

3. Laboratory Context

Operational focus: Quality assurance and validation of homogenised reference coal materials. Testing frequency is occasional, primarily for CRM stability checks and verification of homogenised lots. Duplicate analyses are performed for each batch, accompanied by CRMs to monitor drifts.

Internal QC practices include weekly verification of crushers, furnaces, and ovens (every Monday). Furnace and oven calibration is externally done by Micron, while the calorimeter is calibrated in-house. External technicians service crushers and pulverisers.

AzurESky

4. Quality-Assurance Measures Supporting PT

QA / IQC Measure	Frequency	Purpose
CRM analysis with each duplicate batch	Each testing event	Detect and quantify drift or bias
Duplicate sample analysis	Each testing event	Check repeatability
Crusher verification	Weekly (Monday)	Validate sample-preparation uniformity
Furnace & oven verification	Weekly (Monday)	Confirm temperature accuracy
CRM stability check	Monthly	Detect degradation of stored reference materials
External calibration (Micron)	Annual	Ensure metrological traceability
External servicing of mechanical units	As scheduled	Maintain performance stability

5. Risk-Based Determination of PT Frequency

Risk Factor	Risk Level	Mitigation / QA Action
Limited sample throughput	Low	Monthly CRM verification plus quarterly PT participation
Staff turnover	Low	Cross-training and annual competency review
Instrument drift	Moderate	Duplicate,CRM monitoring, weekly verifications
Method complexity	Moderate	Periodic bias trending and external PT
External PT availability	Readily available	Enrolled quarterly

AzurESky

6. Areas of Technical Competence

Technical Area	Measurement Process	Parameters Included	Key Equipment
Proximate Analysis	Gravimetric/ Thermogravimetric	Quick Ash, ISO Ash, ISO Volatile Matter, ISO Moisture	Muffle furnace, drying oven
Calorific-Value Determination	Combustion Calorimetry	Calorific Value	Bomb calorimeter

7. PT Participation Strategy and Frequency

Parameter	External PT (CoalConcept)	Internal PT (CRM Checks)	Total Frequency	Rationale
Quick Ash	Quarterly	Monthly CRM check + duplicates	4 external + 12 internal	Moderate risk; temperature-dependent
ISO Ash	Quarterly	Monthly CRM check	4 external + 12 internal	Linked to calorific correlation
ISO Volatile Matter	Quarterly	Monthly CRM check	4 external + 12 internal	Sensitive to furnace behaviour
ISO Moisture	Quarterly	Monthly CRM check	4 external + 12 internal	Influenced by humidity
Calorific Value	Quarterly	Monthly CRM check	4 external + 12 internal	Key coal quality indicator

8. Graphical Summary – PT Cycle

The diagram below illustrates the PT cycle and review process:

Weekly Equipment Verification → Monthly CRM/Duplicate Check → Quarterly External PT (CoalConcept) → Bias & Trend Evaluation → Corrective Actions if Needed → Annual Management Review → PT Plan Update

9. Conclusion

This PT Plan demonstrates a sound, risk-based approach suited to a non-commercial coal reference laboratory with limited sample throughput. Through structured internal CRM verification, weekly equipment checks, and quarterly external PT participation, the laboratory ensures ongoing validation of analytical competence and readiness for SANAS accreditation.

AzurESky

After

Proficiency Testing (PT) Participation Plan (2025 – 2029)

Laboratory Name: Azuresky Pty Ltd

Document No: QMS-AS001 PT

Revision: 01

Effective Date: 13/04/2026

Next Review: 04/2030

Prepared by: Laboratory Manager Signature: _____

Approved by: Operations Manager Signature: _____

1. Objective

To define the laboratory's 5-year strategy for participation in internal and external proficiency testing schemes. The plan ensures continuous verification of analytical competence for coal-related parameters under controlled, low-throughput conditions where PT complements other quality-assurance (QA) measures.

2. Scope

Parameter	Method/Standard
Quick Ash	In-house method (based on ISO 1171)
ISO Ash	ISO 1171
ISO Volatile Matter	ISO 562
ISO Moisture (Inherent Moisture)	SANS 5925
Calorific Value	ISO 1928
Total Moisture	ISO 589
Preparation	ISO 13909-4

3. Laboratory Context

Operational focus: Quality assurance and validation of homogenised reference coal materials. Testing frequency is occasional, primarily for CRM stability checks and verification of homogenised lots. Duplicate analyses are performed for each batch, accompanied by CRMs to monitor drifts.

Internal QC practices include weekly verification of crushers, furnaces, and ovens (every Monday). Furnace and oven calibration is externally done by Micron, while the calorimeter is calibrated in-house. External technicians service crushers and pulverisers.

4. Quality-Assurance Measures Supporting PT

QA / IQC Measure	Frequency	Purpose
CRM analysis with each duplicate batch	Each testing event	Detect and quantify drift or bias
Duplicate sample analysis	Each testing event	Check repeatability
Crusher verification	Weekly (Monday)	Validate sample-preparation uniformity
Furnace & oven verification	Weekly (Monday)	Confirm temperature accuracy
CRM stability check	Monthly	Detect degradation of stored reference materials
External calibration (Micron)	Annual & every two years for mass pieces	Ensure metrological traceability
External servicing of mechanical units	As scheduled or when required	Maintain performance stability

5. Risk-Based Determination of PT Frequency

Risk Factor	Risk Level	Mitigation / QA Action
Limited sample throughput	Low	Monthly CRM verification plus quarterly PT participation
Staff turnover	Low	Cross-training and annual competency review
Instrument drift	Moderate	Duplicate,CRM monitoring, weekly verifications
Method complexity	Moderate	Periodic bias trending and external PT
External PT availability	Readily available	Enrolled quarterly

AzurESky

6. Areas of Technical Competence

Technical Area	Measurement Process	Parameters Included	Key Equipment
Proximate Analysis	Gravimetric/ Thermogravimetric	Quick Ash, ISO Ash, ISO Volatile Matter, ISO Moisture	Muffle furnace, drying oven
Calorific-Value Determination	Combustion Calorimetry	Calorific Value	Bomb calorimeter
Total Moisture	Thermogravimetric	Total Moisture	Drying oven, Top Pan Balance
Sample Preparation	Crushing, Splitting & Milling	Preparation of Coal	Crushers, Splitters, Pulverizers

7. PT Participation Strategy and Frequency

Parameter	External PT (Coal Concept)	Internal PT (CRM Checks)	Total Frequency	Rationale
Quick Ash	Quarterly	Monthly CRM check + duplicates	4 external + 12 internal	Moderate risk; temperature-dependent
ISO Ash	Quarterly	Monthly CRM check	4 external + 12 internal	Linked to calorific correlation
ISO Volatile Matter	Quarterly	Monthly CRM check	4 external + 12 internal	Sensitive to furnace behaviour
ISO Moisture	Quarterly	Monthly CRM check	4 external + 12 internal	Influenced by humidity
Calorific Value	Quarterly	Monthly CRM check	4 external + 12 internal	Key coal quality indicator
Total Moisture	Bi-monthly	-	6 per year	Moderate risk, supplier-provided frequency.

Sample Preparation	Bi-monthly	-	6 per year	Moderate risk, supplier-provided frequency.

8. Graphical Summary – PT Cycle

The diagram below illustrates the PT cycle and review process:

Weekly Equipment Verification → Monthly CRM/Duplicate Check → Quarterly External PT (Coal Concept) → Bias & Trend Evaluation → Corrective Actions if Needed → Annual Management Review → PT Plan Update

9. Conclusion

This PT Plan demonstrates a sound, risk-based approach suited to a testing, proficiency testing provider & certified reference material producing laboratory with limited sample throughput. Through structured internal CRM verification, weekly equipment checks, quarterly & bi-monthly external PT participation, the laboratory ensures ongoing validation of analytical competence and readiness for SANAS accreditation compliance.

Date	Revision no	Content Changed
13/04/2026	01	No changes, only the following additions made: • PT plan scope increased to include PT participation for Total moisture & sample preparation • Areas of technical competence & PT participation section updated to include total moisture and sample preparation. • Conclusion updated to include bi-monthly pt participation of sample preparation & total moisture

AzurESky	**Document No.** AS008MF	**Page 1 of 1**
	Effective Date	**April 2026**
	Revision Date	**April 2029**

Title:	Memorandum	
Document by:	Azuresky Pty Ltd	**Revision No**
Compiled by:	Sifiso Nyambi & Tshepo Maponya	**00**
Authorizing Officer: N Msibi	**Signature:**	

Department: Management

Memorandum no: 002-2026

Position: Lab Supervisor (Maponya TMJ)

Date: 13/04/2026

Title: Additions to the laboratory PT-Plan

Kindly take note of the additions made in the PT Plan QMS-AS001 PT which now includes preparation and Total moisture methods.

Moreover, personell are to adhere to the frequency of the participation and make certain that samples are ordered on time, processed and reported on time with the relevant service provider.

AzurESky	ATTENDANCE REGISTER		
	Document no:	Version:	Page:
	ASMF 17043-8.4.1	0	1
	Compiled by:	Sifiso Nyambi - Laboratory Manager	
	Approved by:	Nonhlanhla Msibi - Operations Manager	
Date:	13/04/2026		
Chair:	Tshepo Mapanya	Scribe	
Subject:	Additions to the lab PT Plan QMS-AS001 PT		
Absent:			

Present	Cell Number	Designation	Signature	Email address
TMJ Mapanya	076 455 0692	Lab Supervisor	Mapanya	tshepo@azuresky.it.com
A Jama	060 611 5594	Lab analyst		Abenethi@Azuresky.it.com
NN Hlongwane	0723732558	Lab Analyst		nosipho@Azuresky.it.com
SB Nyambi	072 174 1852	Lab Manager		Sifiso@azuresky.it.com
N. Msibi	0827655149	Lab Operations M		nonhlanhla@azuresky.it.com