	AzureSky		
	AS005IW ROOM TEMPERATURE CONTROL AND HUMIDITY		
	Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
	Approved by:	**Nonhlanhla Msibi- Operations Manager**	

1. *OBJECTIVE*

To ensure that the room temperature and humidity fall within the acceptable range, to improve the accuracy and validity of results.

2. *RESPONSIBILITY*

- Lab Analyst
- Laboratory Supervisor

3. *PRINCIPLE*

3.1. Use calibrated thermometers, hygrometers, and data loggers to ensure that samples are kept in a controlled environment (one that aligns with the specified methods).

3.2. Variable humidity greatly affects the inherent moisture content in coal, thus should be carefully monitored.

3.3. All samples should be conditioned in a controlled environment.

4. *PROCEDURE*

4.1. Check if the hygrometer and/or thermometer are still working at the beginning of the shift. Also make a note as to when the calibration of the equipment is due.

4.2. Look at the readings (humidity and temperature) and record them on the AS018TF FORM at the beginning of every shift.

4.3. If either the temperature or humidity falls outside the acceptable range, adjust the air conditioner accordingly (i.e. if the temperature is too low, slightly increase the temperature of the air conditioner and vice versa).

4.4. If the changes in temperature are too drastic, changes should be implemented immediately.

4.5. Check the readings after an hour. Should the changes still not be effective, inform the supervisor immediately. Record what actions you took under the comments section.

4.6. All the forms must be checked and signed by the lab supervisor daily.

4.7. The laboratory temperature should lie between 19°C and 26°C and humidity 40-60 %.

Effective date	Revision status	Next Review date
04/2025	00	04/2028

	AzureSky		
	AS005IW ROOM TEMPERATURE CONTROL AND HUMIDITY		
	Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
	Approved by:	**Nonhlanhla Msibi- Operations Manager**	

5. TECHNICAL CHANGES

Revision number	Date	Reason for revision

Effective date	Revision status	Next Review date
04/2025	00	04/2028