	AzurESky	
AS023TF	**STORAGE ROOM TEMPERATURE**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

LABORATORY RELATIVE HUMIDITY AND TEMPERATURE RECORDING

Date: _____ Location: Storage Room

				TIMES READINGS TAKEN	READINGS TAKEN BY
Temp. (^{0}C)	SPECIFICATIONS: 16-20	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25-60	ACTUAL READING			
Temp. (^{0}C)	SPECIFICATIONS: 16-20	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25-60	ACTUAL READING			
Temp. (^{0}C)	SPECIFICATIONS: 16-20	ACTUAL READING			
RH (%)	SPECIFICATIONS: 25-60	ACTUAL READING			

Document no	Revision Status	Effective Date	Revision date
AS023TF	01	04/2025	04/2028

AzurESky		
AS023TF	**STORAGE ROOM TEMPERATURE**	
Compiled by:	**Sifiso Nyambi- Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi- Operations Manager**	

*N/B: READINGS TO BE **RECORDED ON MONDAY AND FRIDAY EVERY WEEK**, HOWEVER, PERSONNEL ARE TO CHECK VISUALLY THE ENVIRONMENTAL CONDITIONS FOR COMPLIANCE AT LEAST THREE TIMES PER SHIFT AT INTERVALS. IF OUT OF SPEC REPORT TO SUPERVISOR.*

Shift Supervisor sign**:** _____

Revision table

Content changed	Reason	Date
Humidity changed from 25-45% to 25-60%	Initial humidity values were provisional, practical work revealed that the initial conditions could not be fulfilled.	12/2025

Document no	Revision Status	Effective Date	Revision date
AS023TF	01	04/2025	04/2028