AzurESky "ensure quality of test"	**7.8. REPORTING OF RESULTS**		
	Document no: ASMF 7.8		
	Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

LABORATORY POLICY MEMORANDUM

TITLE: Reporting, Review, and Authorization of Test Results

REFERENCE: ISO/IEC 17025:2017 Clause 7.8 and SANAS R03/4

DOCUMENT NUMBER: 7.8.

1. PURPOSE

To ensure that all test results, certificates, and reports issued by the laboratory are accurate, traceable, authorized, and compliant with the requirements of ISO/IEC 17025:2017 Clause 7.8 and SANAS R03/4.

2. SCOPE

This procedure applies to all laboratory activities related to the preparation, review, approval, issue, amendment, and re-issue of test reports, and sampling reports.

3. RESPONSIBILITIES

Role	Responsibilities
Analyst / Technician	Prepare raw data and draft results in accordance with approved methods.
Technical Signatory / Authorised Person	Review, verify, and authorise all reports prior to release.
Operations Manager	Ensure compliance with ISO 17025 and SANAS R03 requirements; maintain document control.
Laboratory Manager	Approve any amendments, oversee issuance, and ensure only authorised signatories release reports.

AzurESky "ensure quality of test"	**7.8. REPORTING OF RESULTS**		
	Document no: ASMF 7.8		
	Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
	Approved by:	**Nonhlanhla Msibi - Operations Manager**	

4. PROCEDURE

4.1 Review and Authorization of Results

- All test results shall be reviewed for accuracy, completeness, and traceability to the sample and method used.
- Review includes verification of calculations, data transcription, QC results, and correct documentation.
- Reports shall only be released after authorization by an approved Technical Signatory.
- Electronic systems must have secure authentication and audit trails.

4.2 Common Requirements for Test Reports

- Title, laboratory details, report ID, page numbers, client name, sample identification, test dates, methods, results, authorized signatures, and SANAS statement must be included.

4.3 Specific Requirements for Test Reports

- Include deviations, influencing factors, method reference, measurement uncertainty, and sampling details if applicable.
- If any information taken from the client can affect the validity of results, a disclaimer must accompany the report outlining the said information.
- Clearly state within the report that the laboratory is responsible for all the information provided in the report.
- When the lab has not participated in the sampling process (e.g. sample brought by the client) , the report must clearly state that the results only apply to the sample as received.

4.4 Reporting Sampling Requirements

- Include reference to sampling plan, date/location, sampler identification, and relevant conditions.

4.5 Statement of Conformity(Refer to AS010IW)

- Clearly state the decision rule applied and ensure it considers measurement uncertainty where applicable.
- Indicate if results meet or fail specified limits and document the rationale.

AzurESky "ensure quality of test"	**7.8. REPORTING OF RESULTS**	
	Document no: ASMF 7.8	
Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

4.6 Reporting Opinions and Interpretations

- Opinions or interpretations must be clearly identified, supported by documented rationale, and issued only by authorized personnel.

4.7 Amendments to Reports

- Revised reports are issued titled Amended Test Report. The unique report number remains the same in order to make the link to the original test report. All changes are stated clearly on the "NOTE" section with reasons for the change. Any technical signatory can amend a test report.

4.8 Requirements as per SANAS R04

- The SANAS mark must only appear on accredited activities and comply with R04 conditions.
- Non-accredited results must be clearly identified and accompanied by the standard disclaimer.

4.9 Confidentiality and Record Control

- All reports and data shall be securely stored, access restricted, and retention aligned with policy and client agreements.

5. REFERENCES

• ISO/IEC 17025:2017, Clause 7.8

• SANAS R04: Conditions for Use of Accreditation Symbol, Reference to Accreditation and Combined Marks

- SANAS R03: Nominated Representative, Technical Signatories, qualifications, and responsibilities.

• Laboratory QMS Manual

6. REVISION HISTORY

Revision	Date	Description of Change	Approved By

	7.8. REPORTING OF RESULTS	
	Document no: **ASMF 7.8**	
Compiled by:	**Sifiso Nyambi - Laboratory Manager**	
Approved by:	**Nonhlanhla Msibi - Operations Manager**	

00	17/10/2025	Initial issue	Nonhlanhla Msibi