

Kaunas University of Technology
Faculty of Electrical and Electronics Engineering

Functional Near-Infrared Spectroscopy and Machine Learning For Neurocognitive and Mental-Health Screening

Literature Review

Gerson David Pinto Chica
Project Author

Dr. Vidas Raudonis
Supervisor

Dr. Valdas Grigaliūnas
Advisor

Kaunas, 2026

Contents

List of Figures **3**

1 Introduction **4**
1.1 Research Object and Planned Research Workflow 4

2 Literature Review **6**
2.1 Scope of This Literature Analysis 6
2.2 fNIRS Signal Composition and Haemodynamic Estimates 6
2.3 Methodological Evidence From fNIRS Studies 7
2.3.1 Physiological Confounds and Short-Channel Regression 7
2.3.2 Signal Structure and Representation for Machine Learning 8
2.3.3 GASF-Based Task-State Classification 9
2.3.4 Classical Machine Learning and Deep Learning Models 10
2.3.5 CNN-LSTM Neural Network Model for fNIRS Workload Classification . . . 11
2.3.6 Task-State, Workload, and Walking-Task Classification 12
2.3.7 Cognitive-Effort Modelling Framework 12
2.3.8 Leakage-Free Evaluation of Models 13
2.3.9 Reproducibility Guidelines, Standardisation, and Wearable fNIRS 14
2.3.10 Connectivity-Aware and Graph-Based Extensions 15
2.3.11 Cross-Study Interpretation 15
2.3.12 Methodological Implications 16

3 Methodology **17**
3.1 Methodological Risks 17
3.2 Thesis Positioning 17

4 Conclusions **18**

References **19**

List of Figures

Fig. 1. Planned research workflow for the fNIRS acquisition, preprocessing, and subject-wise machine-learning validation pipeline. 4

Fig. 2. Conceptual short- and long-separation fNIRS channels. Short channels estimate superficial physiology; long channels contain both superficial and cortical components. 8

Fig. 3. Sequence-preserving fNIRS representation used for model input. 9

Fig. 4. Example convolutional neural network used for GASF-based fNIRS task classification [1]. 10

Fig. 5. Example CNN-LSTM architecture for fNIRS workload classification [2]. 11

Fig. 6. Example fNIRS sensor placement used in walking-task classification [3]. 12

Fig. 7. Leakage-safe evaluation pipeline for fNIRS machine learning. 14

1. Introduction

1.1. Research Object and Planned Research Workflow

The research object in this work is an fNIRS-based neurocognitive state measurement and analysis system, workflow from signal acquisition to haemodynamic state classification using machine learning algorithms [4, 5, 2]. The research target is not only the construction of a measuring device and a classifier. Rather, it is the creation of the complete engineering and data processing chain including signal acquisition, preprocessing, haemodynamic response estimation, and machine-learning based classification of neurocognitive states [4, 5, 2]. Hence, the system should not be viewed as a clinical diagnostic method nor a haemodynamic biomarker test [6, 7, 8, 9].

The data to be analysed would consist of multi-wavelength intensity signals recorded from a frontal or prefrontal fNIRS montage, together with synchronised task-event markers. Optical intensity signals would be processed to produce optical density, followed by calculation of haemoglobin concentration changes using the modified Beer-Lambert law [6, 8, 9]. The primary processed variables would be HbO and HbR; HbDiff and HbT could serve as auxiliary haemodynamic variables.

The planned workflow is shown in Fig. 1. At first, the experimental protocol will be designed, ethical approval will be received, and volunteers will be recruited for the experiment. Informed consent will be obtained before placing the fNIRS cap or headset over the selected montage area of the brain. Then, signal acquisition will begin after calibration and quality check of the signal.




Fig. 1. Planned research workflow for the fNIRS acquisition, preprocessing, and subject-wise machine-learning validation pipeline.

Evaluation of machine learning classifiers would be carried out using subject-wise validation [5, 10]. The data from the same subject would never be included into both train and validation set. This is required since otherwise, the classifier might actually learn personal characteristics of each subject, which would make the classification meaningless [5, 10]. Logistic regression, support vector machines and random forests can be used as baseline models [5, 4]. Other types of more complex machine learning models could be considered only when the collected data amount is sufficient [5, 2, 11].

The planned equipment is the Cortivision Spectrum C23 fNIRS system. The device would be useful in this research since it allows stationary and mobile fNIRS recordings with user-configurable op-

tode placement. Furthermore, the device would allow task event synchronisation using hardware or software markers, as well as short-channel measurements using configured emitters [12].Acquisition would be done using the acquisition computer and Cortivision software, whereas preprocessing and machine learning tasks would require the use of MATLAB or Python languages.It is planned to organise the data set in accordance with the Brain Imaging Data Structure (BIDS) format or the similar one [13, 14].

The planned study will be performed as a pilot study with healthy volunteers. The planned target of recruitment is about 25 participants. However, in order to achieve sufficient statistical power, the minimum number of valid recordings should be no less than 20. Volunteers would be recruited via university announcement, posters, invitations in classes, and direct offers for those who expressed their interest in participating.

2. Literature Review

In recent years, functional near-infrared spectroscopy (fNIRS) has become an increasingly valuable neuroimaging technique for the investigation of brain functionality and the dynamics of neurocognitive processes [15, 16]. One of the advantages of fNIRS is its portability and suitability for studies in realistic environments that cannot accommodate bulky scanner setups. In neurocognitive and mental-health screening contexts, however, fNIRS should be considered cautiously. Continuous-wave fNIRS does not estimate brain activity directly and does not provide a diagnostic biomarker by itself. Instead, it provides an indirect estimate of relative haemodynamic responses based on changes in oxyhaemoglobin and deoxyhaemoglobin levels [6, 7, 8].

Therefore, the scientifically defensible question is not whether fNIRS and machine-learning algorithms can directly diagnose mental-health conditions or neurodevelopmental disorders. A more accurate question is whether relative changes in frontal haemodynamic patterns can be preprocessed, represented, and classified under valid methodological conditions. This view aligns with current best practices and recent reproducibility guidance, which show that fNIRS outcomes are strongly influenced by signal quality, short-channel correction, data structuring, terminology, and evaluation strategy [6, 17, 9, 5, 10].

2.1. Scope of This Literature Analysis

The key challenge is not only obtaining high accuracy with a classifier. The more important question is whether the signal entered into the classifier is physiologically plausible. Current fNIRS methodology stresses that systemic physiology, superficial blood flow, motion, optode contact, preprocessing choices, and participant-wise evaluation must be controlled before a classification result can be interpreted [6, 7, 18, 17].

Several papers cited in this review contain examples of fNIRS classification and deep learning [1, 2, 4, 3, 19, 11]. Their relevance is conditional on the limitations of fNIRS data. The thesis therefore relies most strongly on methodological evidence concerning short-channel regression, physiological confounds, benchmarking, reproducible data organization, and data leakage [20, 21, 18, 22, 5, 10, 13].

2.2. fNIRS Signal Composition and Haemodynamic Estimates

An fNIRS signal is obtained by detecting near-infrared light intensity after scattering and absorption in biological tissue. Optical density (OD) can be calculated by subtracting the logarithm of the ratio between recorded and reference intensities:

$$\mathrm{OD}(\lambda,t) = -\log\left(\frac{I(\lambda,t)}{I_0(\lambda)}\right). \tag{1}$$

The corresponding change in optical density is defined as:

$$\Delta \mathrm{OD}(\lambda, t) = \mathrm{OD}(\lambda, t) - \mathrm{OD}(\lambda, t_0). \tag{2}$$

For continuous-wave fNIRS, optical-density signals are transformed into relative haemodynamic signal changes by applying the modified Beer-Lambert law. For a given wavelength λ, this relation can be written as:

$$\Delta A(\lambda) = \mathrm{DPF} \cdot r \cdot \left[\varepsilon_{\mathrm{HbO}}(\lambda)\Delta\mathrm{HbO} + \varepsilon_{\mathrm{HbR}}(\lambda)\Delta\mathrm{HbR}\right], \tag{3}$$

where DPF is the differential pathlength factor, r is the source-detector distance, $\varepsilon_{\mathrm{HbO}}$ and $\varepsilon_{\mathrm{HbR}}$ are the extinction coefficients of oxyhaemoglobin and deoxyhaemoglobin, and ΔHbO and ΔHbR are relative concentration changes [6, 8, 9]. For two wavelengths, the same relation can be represented as:

$$\begin{bmatrix} \Delta A(\lambda_1) \\ \Delta A(\lambda_2) \end{bmatrix} = \mathrm{DPF} \cdot r \begin{bmatrix} \varepsilon_{\mathrm{HbO}}(\lambda_1) & \varepsilon_{\mathrm{HbR}}(\lambda_1) \\ \varepsilon_{\mathrm{HbO}}(\lambda_2) & \varepsilon_{\mathrm{HbR}}(\lambda_2) \end{bmatrix} \begin{bmatrix} \Delta\mathrm{HbO} \\ \Delta\mathrm{HbR} \end{bmatrix}. \tag{4}$$

The fNIRS features used in this thesis are based on the following relative haemodynamic estimates:

$$\mathrm{HbDiff} = \Delta\mathrm{HbO} - \Delta\mathrm{HbR}, \tag{5}$$

$$\mathrm{HbT} = \Delta\mathrm{HbO} + \Delta\mathrm{HbR}. \tag{6}$$

These quantities should not be confused with oxygen saturation, because saturation depends on an estimation algorithm and hardware-specific assumptions [8, 9]. For example, a saturation quantity could be calculated as:

$$\mathrm{rSO_2 \text{ or } TSI} = \frac{\mathrm{HbO_2}}{\mathrm{HbO_2} + \mathrm{Hb}} \times 100, \tag{7}$$

but this term should be used only when it is actually calculated by the device or analysis software. Thus, the present research uses ΔHbO, ΔHbR, HbDiff, and HbT as relative haemodynamic features.

2.3. Methodological Evidence From fNIRS Studies

2.3.1. Physiological Confounds and Short-Channel Regression

One disadvantage of fNIRS is that long-distance channels receive mixed signals from cortical haemodynamics and superficial tissues, such as the scalp and skin. Superficial physiology includes systemic activity, blood-pressure changes, breathing, skin and scalp blood flow, motion artefacts, Mayer waves, and optode-contact changes [7, 18]. If these variables are not controlled, they can become the main drivers of machine-learning model output.

One way to address this limitation is to use short-distance channels, because short channels are more sensitive to superficial physiological noise than long channels [20, 21, 22]. The following equation illustrates how a short-channel signal can serve as a nuisance variable for correcting a long-channel signal:

$$y_{\text{clean}} = y_{\text{long}} - \alpha y_{\text{short}}, \qquad \alpha = \frac{\langle y_{\text{short}}, y_{\text{long}} \rangle}{\langle y_{\text{short}}, y_{\text{short}} \rangle}. \tag{8}$$

Here, y_{long} is a measured long-distance signal, y_{short} is a short-channel signal, and y_{clean} is the corrected long-distance signal. The coefficient α defines the proportion of superficial haemodynamic activity shared between short and long channels.

More generally, this correction can be expressed using a general linear model:

$$Y = X_{\text{task}}\beta_{\text{task}} + X_{\text{SDC}}\beta_{\text{SDC}} + X_{\text{phys}}\beta_{\text{phys}} + \varepsilon, \tag{9}$$

where X_{task} includes task-related regressors, X_{SDC} includes short-distance-channel regressors, X_{phys} includes additional physiological nuisance variables where available, and ε includes the residual signal. This decomposition is especially important in mental-health applications, where intergroup task-related differences can be subtle and easily confounded by systemic physiology [7, 17, 18].

Recent reproducibility studies confirm the need to treat noise and preprocessing as part of scientific validity rather than as secondary details. The FRESH initiative showed that fNIRS findings are sensitive to signal quality, preprocessing steps, and researcher experience [17]. For this thesis, short-channel regression, artefact handling, and explicit reporting of preprocessing are therefore key steps before machine-learning techniques are applied.




Fig. 2. Conceptual short- and long-separation fNIRS channels. Short channels estimate superficial physiology; long channels contain both superficial and cortical components.

2.3.2. Signal Structure and Representation for Machine Learning

One of the most important steps in machine learning with fNIRS data is signal representation. Unlike images, raw fNIRS data are not naturally represented in two-dimensional spatial coordinates. They are multichannel time series. For an analysis window, the input can be defined as:

$$X \in \mathrm{R}^{C \times T \times K}, \tag{10}$$

where C is the number of fNIRS channels, T is the number of time samples per channel, and K is the number of haemodynamic signals, such as ΔHbO, ΔHbR, HbDiff, and HbT. For N analysis windows, the full dataset can be written as:

$$\mathscr{X} \in \mathrm{R}^{N\times C\times T\times K}. \tag{11}$$

This representation is preferable to aggressive reshaping into a generic image format. Modern fNIRS deep-learning models show that CNN-LSTM, 1D convolutional, attentional, and temporal convolutional models can process temporal and spatial aspects of fNIRS signals without discarding their channel-by-time structure [2, 11, 4]. Classical machine-learning models can also use extracted features from these windows, provided that feature extraction and model evaluation are leakage-safe [5].




Fig. 3. Sequence-preserving fNIRS representation used for model input.

2.3.3. GASF-Based Task-State Classification

Some studies use image-based transformations of fNIRS data, such as Gramian Angular Summation Field (GASF). Wickramaratne and Mahmud used fNIRS signals transformed with GASF for three-class task classification, including mental arithmetic, motor imagery, and idle state [1]. This confirms that fNIRS data can contain task-related haemodynamic activity, although it also raises questions about whether GASF preserves physiologically meaningful structure.

The GASF representation is created by first normalizing the signal:

$$\tilde{x}_i = \frac{x_i - \min(X)}{\max(X) - \min(X)}. \tag{12}$$

Each sample is then converted into a polar-angle representation:

$$\varphi_i = \arccos(\tilde{x}_i), \tag{13}$$

and the matrix is constructed as:

$$\mathrm{GASF}_{ij} = \cos(\varphi_i + \varphi_j), \tag{14}$$

$$\mathrm{GASF}_{ij} = \tilde{x}_i\tilde{x}_j - \sqrt{1-\tilde{x}_i^2}\sqrt{1-\tilde{x}_j^2}. \tag{15}$$

Although GASF may be useful for designing image-based deep-learning models, this technique can

distort the original physiological structure of fNIRS data. In this thesis, GASF is therefore treated as one possible representation from the literature rather than as the preferred approach. To retain physiological meaning, the window length, sampling frequency, channel distribution, and haemodynamic signal structure should remain explicit [2, 11, 5].




Fig. 4. Example convolutional neural network used for GASF-based fNIRS task classification [1].

2.3.4. Classical Machine Learning and Deep Learning Models

Classical machine-learning techniques remain important because they provide interpretable and reproducible baselines for deep-learning models. Common classical methods in fNIRS studies include linear discriminant analysis, logistic regression, support vector machines with linear or kernel methods, decision trees, random forests, and k-nearest neighbours [5, 4]. These models are usually applied to manually extracted features such as mean, slope, peak value, time-to-peak, variance, area under the curve, or spectral features. A feature extraction operation can be written as:

$$z_i = g(X_i), \tag{16}$$

where X_i is a haemodynamic window and z_i is a feature vector. The classifier then maps the feature vector into class probabilities:

$$\hat{y}_i = f_\theta(z_i). \tag{17}$$

Deep-learning models allow more flexible feature extraction without manual feature design. However, this increased flexibility also increases the risk of overfitting and leakage. Two common fNIRS deep-learning architectures use convolutional layers and recurrent layers, such as LSTM layers. CNN layers extract spatial or local temporal structures, whereas LSTM layers model temporal dependencies. One-dimensional convolution over time can be written as:

$$Y[i] = f\left(\sum_{j=1}^{k} X[i+j-1] \cdot W[j] + b\right), \tag{18}$$

where k is the convolution-kernel size, W is the kernel, b is the bias, and $f(\cdot)$ is a nonlinear activation function. A commonly used activation function is the rectified linear unit:

$$\mathrm{ReLU}(x) = \max(0, x). \tag{19}$$

Pooling can be applied to reduce the temporal dimension:

$$Y[i] = \max(X[i \cdot p : (i+1) \cdot p]). \tag{20}$$

CNN-LSTM models are relevant because fNIRS responses evolve over time. Recent research also explores spatiotemporal convolution, temporal convolutional networks, and attention mechanisms [4, 11]. In general, the final classifier of such a model can be expressed as:

$$\hat{y} = f_\theta(X), \quad X \in \mathrm{R}^{C \times T \times K}. \tag{21}$$

The loss function commonly used for classification problems is cross-entropy:

$$\mathscr{L} = -\frac{1}{N} \sum_{i=1}^{N} \sum_{m=1}^{M} y_{im} \log(\hat{p}_{im}), \tag{22}$$

where M is the number of classes, y_{im} is the class indicator, and $\hat{p}_{im}$ is the predicted probability for class m.

2.3.5. CNN-LSTM Neural Network Model for fNIRS Workload Classification

Convolutional and LSTM layers are potentially useful for building an architecture capable of processing two forms of variation in fNIRS data. CNN layers learn local patterns, whereas LSTM layers model dependencies between samples [2]. Nevertheless, high reported accuracy in workload classification cannot imply diagnostic ability, because workload classes are experimentally defined and may reflect many non-clinical factors.




Fig. 5. Example CNN-LSTM architecture for fNIRS workload classification [2].

2.3.6. Task-State, Workload, and Walking-Task Classification

Modern literature supports fNIRS machine learning for task-state, workload, and walking-task classification more strongly than for standalone diagnosis. Khan et al. show the benefits of deep learning for cognitive workload modelling [2]. This is relevant because workload is related to mental effort. Nevertheless, workload classification cannot be equated with mental-health screening: a classification task is defined experimentally, whereas a diagnosis is defined by clinical criteria.

Ma et al. developed a deep-learning model to distinguish walking tasks in older adults [3]. This work demonstrates the potential of fNIRS and deep learning for classification in motor-cognitive paradigms. At the same time, motion artefacts become especially important during walking, which increases the need for careful preprocessing and leakage-free evaluation [3, 7, 10].




Fig. 6. Example fNIRS sensor placement used in walking-task classification [3].

2.3.7. Cognitive-Effort Modelling Framework

A cognitive-effort modelling framework is relevant because it allows the analysis to consider individual effort in addition to task performance [19]. Cognitive effort must be defined independently from performance. Standardization can be written as:

$$z = \frac{x - \mu}{\sigma}. \tag{23}$$

For performance and cognitive-load variables, this becomes:

$$P_z(s,i) = \frac{P(s,i) - \mu(P_i)}{\sigma(P_i)}, \tag{24}$$

$$M_z(s,i) = \frac{M(s,i) - \mu(M_i)}{\sigma(M_i)}. \tag{25}$$

Relative neural efficiency (RNE) and relative neural involvement (RNI) can then be written as:

$$\mathrm{RNE} = \frac{P_z - M_z}{\sqrt{2}}, \tag{26}$$

$$\mathrm{RNI} = \frac{P_z + M_z}{\sqrt{2}}. \tag{27}$$

RNE and RNI may help distinguish participants in terms of cognitive effort, but they cannot serve as reliable indicators of clinical disorders. They are therefore useful only as possible neurocognitive task-classification variables.

2.3.8. Leakage-Free Evaluation of Models

Several machine-learning issues create major challenges for fNIRS research, especially leakage and overfitting. When time windows from the same participant are divided between training, validation, and test sets, a classifier may learn participant-specific features, superficial physiology, motion patterns, or optode-placement differences instead of task-related haemodynamic responses [5, 10].

A leakage-free procedure can be described by splitting subjects first:

$$S = S_{\text{train}} \cup S_{\text{val}} \cup S_{\text{test}}, \tag{28}$$

where:

$$S_{\text{train}} \cap S_{\text{val}} = \emptyset, \quad S_{\text{train}} \cap S_{\text{test}} = \emptyset, \quad S_{\text{val}} \cap S_{\text{test}} = \emptyset. \tag{29}$$

If the time windows from participant i are denoted as $w_{i,t}$, then:

$$i \in S_k \Rightarrow w_{i,t} \in D_k \quad \forall t. \tag{30}$$

This means that windows from a participant can belong to the training set, validation set, or test set, but cannot be divided across those sets. Splitting should occur before scaling, feature selection, dimensionality reduction, hyperparameter tuning, or model selection [5].

BenchNIRS provides a useful background for leakage-safe fNIRS machine learning because it eval-

uates several models using public fNIRS Hub datasets and nested cross-validation [5]. It also shows that performance on unseen data is lower than many optimistic reports suggest. Nested cross-validation can be expressed as:

$$\theta^* = \arg\min_{\theta} \frac{1}{K} \sum_{k=1}^{K} \mathscr{L}_{\text{val}}^{(k)}(\theta), \tag{31}$$

where θ^* denotes hyperparameters selected in the inner loop. The selected model is then tested on the outer set of participants:

$$\text{Score}_{\text{test}} = \text{Metric}(f_{\theta^*}(X_{\text{test}}), y_{\text{test}}). \tag{32}$$

If classes are unbalanced, balanced accuracy is more informative than ordinary accuracy:

$$\text{Balanced Accuracy} = \frac{1}{M} \sum_{m=1}^{M} \frac{\text{TP}_m}{\text{TP}_m + \text{FN}_m}, \tag{33}$$

where M is the number of classes, TP_m is the number of true positives for class m, and FN_m is the number of false negatives. F1-score, sensitivity, specificity, and confusion matrices should also be reported.




Fig. 7. Leakage-safe evaluation pipeline for fNIRS machine learning.

2.3.9. Reproducibility Guidelines, Standardisation, and Wearable fNIRS

The diversity of fNIRS preprocessing and modelling approaches makes reproducibility a central concern. FRESH shows that findings can vary with signal quality, analysis pipelines, and researcher expertise [17]. Consequently, machine-learning models should not be compared unless preprocessing and data quality are reported clearly.

NIRS-BIDS is important for this thesis because it extends the Brain Imaging Data Structure to near-infrared spectroscopy [13]. It supports standardisation of subject folders, tasks, events, channels, optodes, coordinate systems, and SNIRF-compatible data. The broader BIDS framework also provides a model for reproducible neuroimaging data organization [14]. Organizing fNIRS data according to a BIDS-compatible or BIDS-inspired scheme is therefore an important future step for the thesis.

Wearable fNIRS is another reason why this research direction may become more relevant in the future. Dense-sampling wearable platforms may support repeated measurements in natural environments and precision functional neuroimaging [23]. However, current evidence remains preliminary. The safest conclusion is that wearable fNIRS may enable longitudinal studies and individual functional mapping, but more diverse clinical cohorts are required before claims about disorder screening can be made.

2.3.10. Connectivity-Aware and Graph-Based Extensions

Most fNIRS classification studies treat channels as independent variables. Recent work has started to model fNIRS data as networks, where nodes represent channels or cortical areas and edges represent relationships between those areas [24]. This may be relevant because brain functions are not isolated to single cortical locations.

A graph convolutional layer can be represented as:

$$H^{(l+1)} = \sigma\left(\tilde{D}^{-\frac{1}{2}}\tilde{A}\tilde{D}^{-\frac{1}{2}}H^{(l)}W^{(l)}\right), \tag{34}$$

where $\tilde{A}$ is an adjacency matrix with self-edges, $\tilde{D}$ is a degree matrix, $H^{(l)}$ is the feature matrix at layer l, $W^{(l)}$ is a trainable weight matrix, and σ is a nonlinear activation function. Tang et al. predicted cortical-thalamic functional connectivity from fNIRS-derived functional data using graph convolutional networks [24]. This approach does not remove the depth limitation of fNIRS, but it shows that graph-based methods can model connectivity-aware information. In this thesis, graph-based approaches should therefore be considered future additions rather than the immediate methodological step.

2.3.11. Cross-Study Interpretation

The literature provides a cautious but sound basis for the thesis position. fNIRS combined with machine-learning techniques can be used to classify task patterns, cognitive workload, and cognitive effort in walking tasks [1, 2, 3, 19, 11, 5]. At the same time, fNIRS and machine learning are not sufficient for conclusive diagnosis of mental disorders or other brain conditions. This caution is supported by broader neuroimaging evidence showing that clinical classification can remain weak even in large multi-site datasets and that human neurobiology does not map cleanly onto simple diagnostic labels [25, 26].

Therefore, the validity of fNIRS machine learning should be tested by evaluating the entire pipeline:

$$\begin{aligned}&\text{Raw intensity} \rightarrow \text{Optical density} \rightarrow \Delta\text{HbO}, \Delta\text{HbR} \rightarrow \text{Short-channel correction}\\ &\rightarrow \text{Windowed representation} \rightarrow \text{Leakage-safe model training} \rightarrow \text{Unseen-participant evaluation.}\end{aligned} \tag{35}$$

If one step in this chain is problematic, classification may produce apparently accurate results without physiological meaning. For this reason, the thesis should evaluate classifier performance by generalisation to unseen subjects, stability across folds, comparison with baseline methods, and interpretability of channel and haemodynamic contributions.

2.3.12. Methodological Implications

Traditional machine-learning models are important baselines. At minimum, logistic regression or linear discriminant analysis, linear support vector machine, RBF-kernel support vector machine, and random forest should be tested. CNN, LSTM, CNN-LSTM, and temporal-convolutional models become useful only if they generalize better than these baselines [5, 2, 11].

The reviewed literature supports further development of this research only under careful methodological conditions. The thesis should avoid claiming that fNIRS diagnoses mental-health or neurodevelopmental conditions. A stronger and more defensible claim is that task-evoked fNIRS signals, analysed with short-channel correction and participant-wise validation, can reveal relative haemodynamic patterns associated with selected neurocognitive task conditions.

3. Methodology

The thesis topic sits in an attractive but risky research space. fNIRS offers a portable and non-invasive way to observe relative haemodynamic activity during cognitive, motor, educational, or social tasks. However, mental-health and neurodevelopmental conditions rarely map to one clean physiological marker. A model may find statistical structure in the signal without learning a clinically meaningful pattern.

For that reason, the defensible framing is not that fNIRS diagnoses mental disease. The stronger position is that task-evoked fNIRS, combined with leakage-safe machine learning, may contribute to objective assessment by modelling relative neurocognitive response patterns.

3.1. Methodological Risks

The current methodology must control four risks. First, extracerebral systemic physiology can contaminate long-separation fNIRS channels, so short-distance channels should be used as nuisance regressors where available [7, 20, 21, 22]. Second, fNIRS should be represented as channel-by-time haemodynamic data rather than arbitrarily resized into a generic image grid [2, 11]. Third, HbDiff and HbT are relative haemodynamic quantities and should not be described as tissue oxygen saturation [6, 8]. Fourth, model evaluation must avoid leakage from participant identity, preprocessing, hyperparameter tuning, and overlapping windows [5, 10].

3.2. Thesis Positioning

The reviewed literature supports task-state, workload, walking-task, and cognitive-effort modelling more strongly than stand-alone clinical diagnosis [1, 4, 3, 19]. A strong and defensible aim is to investigate whether task-evoked fNIRS signals, analysed with short-channel correction and subject-wise validation, can reveal objective relative haemodynamic patterns associated with selected neurocognitive indicators.

4. Conclusions

This literature analysis shows that the thesis methodology can be supported by a modest, but solid claim: the combination of fNIRS and machine learning is currently better supported for classifying relative haemodynamic task patterns than for clinical diagnosis.

To that end, the updated methodology considers the following factors: short-distance channel regression for eliminating extracerebral effects, preservation of channel-by-time representation, distinction of relative haemodynamic patterns, and reporting of oxygenation only when explicitly estimated by the instrument and algorithm.

The future research should address leakage problems, employ nested hyperparameter optimization, perform preprocessing in a leakage-safe manner, use robust baselines, and implement model interpretability techniques.

References

[1] S. D. Wickramaratne and M. S. Mahmud. A deep learning based ternary task classification system using gramian angular summation field in fnirs neuroimaging data, 2021. URL https://arxiv.org/abs/2101.05891.

[2] Mehshan Ahmed Khan, Houshyar Asadi, Mohammad Reza Chalak Qazani, Adetokunbo Arogbonlo, Siamak Pedrammehr, Adnan Anwar, Asim Bhatti, Saeid Nahavandi, and Chee Peng Lim. Enhancing cognitive workload classification using integrated lstm layers and cnns for fnirs data analysis. *Computers*, 14(2), 2025. doi: 10.3390/computers14020073. URL https://www.mdpi.com/2073-431X/14/2/73.

[3] Dongning Ma, Meltem Izzetoglu, Roee Holtzer, and Xun Jiao. Deep learning based walking tasks classification in older adults using fnirs. *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 31:3437–3447, 2023. doi: 10.1109/TNSRE.2023.3306365. URL https://arxiv.org/abs/2102.03987.

[4] Condell Eastmond, Aseem Subedi, Suvranu De, and Xavier Intes. Deep learning in fnirs: A review. *Neurophotonics*, 9(4), 2022. doi: 10.1117/1.NPh.9.4.041411. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC9301871/.

[5] Johann Benerradi, Jeremie Clos, Aleksandra Landowska, Michel F. Valstar, and Max L. Wilson. Benchmarking framework for machine learning classification from fnirs data. *Frontiers in Neuroergonomics*, 4, 2023. doi: 10.3389/fnrgo.2023.994969. URL https://www.frontiersin.org/articles/10.3389/fnrgo.2023.994969/full.

[6] Meryem A. Yucel et al. Best practices for fnirs publications. *Neurophotonics*, 8(1), 2021. doi: 10.1117/1.NPh.8.1.012101. URL https://discovery.ucl.ac.uk/id/eprint/10119544/.

[7] Felix Scholkmann, Ilias Tachtsidis, Martin Wolf, and Ursula Wolf. Systemic physiology augmented functional near-infrared spectroscopy: A powerful approach to study the embodied human brain. *Neurophotonics*, 9(3), 2022. doi: 10.1117/1.NPh.9.3.030801. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC9272976/.

[8] Caterina Amendola, Davide Contini, Rebecca Re, Lorenzo Spinelli, Lorenzo Frabasile, Pietro Levoni, and Alessandro Torricelli. Robustness of tissue oxygenation estimates by continuous wave space-resolved near infrared spectroscopy. *Journal of Biomedical Optics*, 28(7), 2023. doi: 10.1117/1.JBO.28.7.075002. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC10351877/.

[9] Katharina Stute, Louisa K. Gosse, Samuel Montero-Hernandez, Guy A. Perkins, and Meryem A. Yucel. The fnirs glossary project: A consensus-based resource for functional near-infrared spectroscopy terminology. *Neurophotonics*, 12(2), 2025. doi: 10.1117/1.NPh.12.2.027801. URL https://doi.org/10.1117/1.NPh.12.2.027801.

[10] Geoffrey Brookshire, Jake Kasper, Nicholas M. Blauch, Yunan Charles Wu, Ryan Glatt, David A. Merrill, Spencer Gerrol, Keith J. Yoder, Colin Quirk, and Che Lucero. Data leakage in deep learning studies of translational eeg. *Frontiers in Neuroscience*, 18, 2024. doi: 10.3389/fnins.2024.1373515. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC11099244/.

[11] Yuxin Qin, Baojiang Li, Wenlong Wang, Xingbin Shi, Cheng Peng, and Yifan Lu. Classification algorithm for fnirs-based brain signals using convolutional neural network with spatiotemporal feature extraction mechanism. *Neuroscience*, 542:59–68, 2024. doi: 10.1016/j.neuroscience.2024.02.011. URL https://www.sciencedirect.com/science/article/pii/S0306452224000617.

[12] Cortivision. Spectrum fnirs system (c23): Product specifications and capabilities, 2026. URL https://cortivision.com/. Accessed: 2026-06-08.

[13] Robert Luke, Robert Oostenveld, Helena Cockx, Guiomar Niso, Maureen J. Shader, Felipe Orihuela-Espina, Hamish Innes-Brown, Stephen Tucker, David Boas, Meryem A. Yucel, et al. Nirs-bids: Brain imaging data structure extended to near-infrared spectroscopy. *Scientific Data*, 12, 2025. doi: 10.1038/s41597-024-04136-9. URL https://www.nature.com/articles/s41597-024-04136-9.

[14] Russell A. Poldrack, Christopher J. Markiewicz, Stefan Appelhoff, Yoni K. Ashar, Tibor Auer, Sylvain Baillet, and Krzysztof J. Gorgolewski. The past, present, and future of the brain imaging data structure. *Imaging Neuroscience*, 2:1–19, 2024. doi: 10.1162/imag_a_00103. URL https://direct.mit.edu/imag/article/doi/10.1162/imag_a_00103/119672/The-past-present-and-future-of-the-Brain-Imaging.

[15] Lirui Yang and Zehua Wang. Applications and advances of combined fmri-fnirs techniques in brain functional research. *Frontiers in Neurology*, 16, 2025. doi: 10.3389/fneur.2025.1542075. URL https://www.frontiersin.org/articles/10.3389/fneur.2025.1542075.

[16] Zehui Zhan, Qinchen Yang, Lixia Luo, and Xia Zhang. Applying functional near-infrared spectroscopy (fnirs) in educational research: A systematic review. *Current Psychology*, 43(11):9676–9691, 2024. doi: 10.1007/s12144-023-05094-y. URL https://doi.org/10.1007/s12144-023-05094-y.

[17] Meryem A. Yucel, Robert Luke, Rickson C. Mesquita, Alexander von Luhmann, et al. fnirs reproducibility varies with data quality, analysis pipelines, and researcher experience. *Communications Biology*, 8, 2025. doi: 10.1038/s42003-025-08412-1. URL https://www.nature.com/articles/s42003-025-08412-1.

[18] Franziska Klein. Optimizing spatial specificity and signal quality in fnirs: An overview of potential challenges and possible options for improving the reliability of real-time applications. *Frontiers in Neuroergonomics*, 5, 2024. doi: 10.3389/fnrgo.2024.1286586. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC11188482/.

[19] Shayla Sharmin, Mohammad Fahim Abrar, Gael Lucero-Palacios, Aditya Raikwar, and Roghayeh Leila Barmaki. Beyond cognitive load: Ai-based estimation of cognitive effort using brain signals during digital tasks. *Delaware Journal of Public Health*, 12(1):32–43,

2026. doi: 10.32481/djph.2026.03.08. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC13082466/.

[20] Dominik G. Wyser, Christoph M. Kanzler, Lena Salzmann, Olivier Lambercy, Martin Wolf, Felix Scholkmann, and Roger Gassert. Characterizing reproducibility of cerebral hemodynamic responses when applying short-channel regression in functional near-infrared spectroscopy. *Neurophotonics*, 9(1), 2022. doi: 10.1117/1.NPh.9.1.015004. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC8901194/.

[21] Franziska Klein, Michael Luhrs, Amaia Benitez-Andonegui, Philipp Roehn, and Cornelia Kranczioch. Performance comparison of systemic activity correction in functional near-infrared spectroscopy for methods with and without short distance channels. *Neurophotonics*, 10(1), 2023. doi: 10.1117/1.NPh.10.1.013503. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC9555616/.

[22] Jiahui An, Pulkit Goyal, Andreas R. Luft, and Josef G. Schonhammer. Functional near-infrared spectroscopy short-channel regression improves cortical activation estimates of working memory load. *Neurophotonics*, 12(3), 2025. doi: 10.1117/1.NPh.12.3.035009. URL https://pmc.ncbi.nlm.nih.gov/articles/PMC12368614/.

[23] Ali Rahimpour Jounghani, Anupam Kumar, Laura Moreno Carbonell, Ester Patrize Lopez Aguilar, Tulla Bee Picardi, Seth Crawford, Audrey K. Bowden, and S. M. Hadi Hosseini. Wearable fnirs platform for dense sampling and precision functional neuroimaging. *npj Digital Medicine*, 8, 2025. doi: 10.1038/s41746-025-01690-3. URL https://www.nature.com/articles/s41746-025-01690-3.

[24] Lingkai Tang, Lilian M. N. Kebaya, Homa Vahidi, Paige Meyerink, Sandrine de Ribaupierre, Soume Bhattacharya, Keith St. Lawrence, and Emma G. Duerden. Predicting cortical-thalamic functional connectivity using functional near-infrared spectroscopy and graph convolutional networks. *Scientific Reports*, 14, 2024. doi: 10.1038/s41598-024-79390-3. URL https://www.nature.com/articles/s41598-024-79390-3.

[25] Vladimir Belov, Tracy Erwin-Grabner, Moji Aghajani, Andre Aleman, Alyssa R. Amod, et al. Multi-site benchmark classification of major depressive disorder using machine learning on cortical and subcortical measures. *Scientific Reports*, 14, 2024. doi: 10.1038/s41598-023-47934-8. URL https://www.nature.com/articles/s41598-023-47934-8.

[26] Emily Beam, Christopher Potts, Russell A. Poldrack, and Amit Etkin. A data-driven framework for mapping domains of human neurobiology. *Nature Neuroscience*, 24:1733–1744, 2021. doi: 10.1038/s41593-021-00948-9. URL https://www.nature.com/articles/s41593-021-00948-9.