Dynamic Systems Inc.

www.Gleeble.com
Info@Gleeble.com

Gleeble System Service Report

Company Name:	University of Witwatersrand		
Gleeble Model Number:	3500	Gleeble Serial Number:	504
Service Engineer:	Garrick Palermo	Date(s) on Site:	2/28 - 3/2

Reason for visit: Annual Service and Calibration

Final Software Specifications			
QNX:	6.5.0	Quiksim:	2.4.5918
GTC:		Origin:	
Simcon:	2.3.20	CCT:	Desktop
SimconPh:	2.0.25	Gleeble Menu:	
MDSP:			

Services Performed			
☐	New Machine Installation	☐	Software Update
☐	System Upgrade	☐	Hardware Repair
☐	Training	✔	Calibration
☐	New Device Installation	✔	Maintenance
☐		☐	

Please select all that apply

Measurement Devices Calibrated			
General Purpose MCU			
✔ Stroke	✔ Force	✔ Thermocouples	✔ C-Gauge
☐ L-Gauge	✔ Dilatometer	☐ Micrometer	☐ 39070
✔ 39071	☐ 39072	☐ 39060	☐ 5K Load Cell
☐	☐	☐	☐
Hydrawedge MCU			
✔ Wedge	✔ Jaw	✔ Force	☐
Torsion MCU			
☐ Thrust	☐ Torque	☐	☐
MAXStrain MCU			

Dynamic Systems Inc.
323 NY 355 Poestenkill, NY 12140
Phone: (518) 283-5350 Fax: (518) 283-3160

GLEEBLE

	Wedge		Force				
Notes:							

Maintenance Performed			
✔	Changed Vacuum Pump Oil	✔	Change Diffusion Pump Oil
✔	Checked/Filled Accumulators (Including all MCU's)	✔	Replaced Hydraulic Pump Filter
✔	Replaced Load Unit Hydraulic Filter	✔	Replaced all MCU Hydraulic Filters
	Replaced Hydraulic Fluid		Replaced MOOG Valve Filter
	Replaced Chiller Fluid	✔	Replaced Water Filter
	Load Unit Servo Tuning	✔	MCU Servo Tuning
Notes:			

Computer Settings			
Desktop IP Address:		Does this pc have internet access?	
Embedded IP Address:		Is it accessible for remote work?	
Any Special Network or User Login Information:			

Summary of the Work Performed and System Recommendations

Please include all necessary photos of items to be replaced!

2/28 -

-Calibrated Hydrawedge: Wedge, Jaw and Force.
-Changed Load Unit and MCU oil filters. Changed Load Unit water filter. Checked and recharged pilot and main accumulator in Load Unit and MCU. Changed Diffusion and Rough pump oil. Hydraulic Pump oil filter
-Installed Pocket Jaw: Stroke, Force, TC, C-Gauge, Dilatometer, HZT071

3/1 -

- Installed Hydrawedge
- HW Servo Tuning
- Autoloader troubleshooting
- Additional training

Summary and Recommendations Continued…

3/2 -

- Continued training
- Autoloader Troubleshooting

michael Bodunrin _____ For University of Witwatersrand

Garrick Palermo _____ For DSI

Note: Per customer request, LForce5K and HZT060 were not calibrated.

Pocket Jaw Flow Stress Anvil Kit: 9555-100
Super Bolt Installation Tool: G389X
TC Ceramic Insulator Sleeve (2 hole): TC026
R-Type TC Wire Insulation Sleeve:
-Black: TU011
-White: TU010
Anvils:
-Cap: 9219-062 (pack of 5)
-Base: 9555-500 (pack of 2)
-Inner Sleeve: 9555-501
-Outer Sleeve: 9555-502
-Spring Clip: 9555-506
-Complete: 9555-102
Imperial Allen Key Set: TOOL023
+J Type TC Wire: TC030
-J Type TC Wire: TC031
Jaw Transducer (Complete): TRN006
Jaw Transducer Extension Shaft: 33646
Black Part Only: TRN006B

Gleeble Support: Support.Gleeble.com
Gleeble Catalog: Gleeble.com/catalog

- 3/2/2022