Dynamic Systems Inc.

HYDRAWEDGE MCU CERTIFICATE OF CALIBRATION

Customer University of Witwatersrand Serial No. 504

HYDRAWEDGE MCU SYSTEM CALIBRATION

CHECKLIST AND SUPPORT DOCUMENTATION

Company Name University of Witwatersrand

Gleeble Model Number 3500 Gleeble Serial Number 504

Room Temperature 22C

Calibration services were provided on the above listed Hydrawedge in accordance with standard factory specified procedures. Calibrations performed have been recorded in this document.

This is to certify that the measurement system of the Hydrawedge listed has been calibrated using transfer standards traceable to the National Institute of Standards and Technology. The accuracy of the thermocouple wire is not included in this calibration process.

Accepted:

michael Bodunrin
(Customer)

Garrick Palermo
(Dynamic Systems Inc.)

3/2/2022
(Date)

Unless otherwise noted, calibration expires one year from above date.

Dynamic Systems Inc.

HYDRAWEDGE MCU CERTIFICATE OF CALIBRATION

Customer University of Witwatersrand Serial No. 504

HYDRAWEDGE MCU STROKE CALIBRATION

REFERENCE INSTRUMENT	MODEL NUMBER	SERIAL NUMBER	CALIBRATION DUE DATE
Mitutoyo	0-2”	11121501	11/12/2022

DEVICE TYPE BEING CALIBRATED SERIAL NO. 121121B71817						
TYPE		RANGE			MOUNTING	
X	Stroke	X	+/-18.00	mm	X	Installed
	Diametrial Extensometer			in.		Free Standing
	Lengthwise Extensometer			kN		In Calibration Fixture
	Load Cell			N		
				kg		
				lbs.		

CALIBRATION IN INCHES 0.1 ____ OR MILLIMETERS X ____

REFERENCE DISTANCE	AS FOUND	OOT	AFTER CALIBRATION
-18.00	-17.89		-17.89
-15.00	-15.00		-15.00
-12.00	-11.99		-11.99
-9.00	-9.00		-9.00
-6.00	-6.00		-6.00
-3.00	-3.00		-3.00
0.00	0.00		0.00
3.00	3.00		3.00
6.00	6.00		6.00
9.00	9.00		9.00
12.00	11.99		11.99
15.00	14.98		14.98
18.00	17.99		17.99

Date of Calibration: 2/28/2022

Calibrated by: G. Palermo

Dynamic Systems Inc.

HYDRAWEDGE MCU CERTIFICATE OF CALIBRATION

Customer University of Witwatersrand Serial No. 504

HYDRAWEDGE 10 TON MCU FORCE SYSTEM CALIBRATION

REFERENCE INSTRUMENT	MODEL NUMBER	SERIAL NUMBER	CALIBRATION DUE DATE
Transducer Techiques	LBC-50K	333592	12/16/2022

DEVICE TYPE BEING CALIBRATED			SERIAL NO. 452371			
TYPE		RANGE			MOUNTING	
	Stroke			mm	X	Installed
	Diametrial Extensometer			in.		Free Standing
	Lengthwise Extensometer	X	-111 kN	kN		In Calibration Fixture
X	Load Cell			N		
				kg		
				lbs.		

CALIBRATION IN KILONEWTONS X NEWTONS ____ KILOGRAMS ____ POUNDS ____

	COMPRESSION		
REFERENCE LOAD	AS FOUND	OOT	AFTER CALIBRATION
0.00	0.00		0.00
-10.00	-9.8		-10.14
-20.00	-19.25		-19.87
-30.00	-28.59	x	-29.6
-40.00	-38.19	x	-39.73
-50.00	-47.79	x	-49.75
-60.00	-57.65	x	-59.60
-70.00	-67.49	x	-69.45
-80.00	-77.32	x	-79.27
-90	88.25	x	-89.12

Date of Calibration: 2/28/2022

Calibrated by: G. Palermo

Dynamic Systems Inc.

HYDRAWEDGE MCU CERTIFICATE OF CALIBRATION

Customer University of Witwatersrand Serial No. 504

HYDRAWEDGE L-STRAIN CALIBRATION

REFERENCE INSTRUMENT	MODEL NUMBER	SERIAL NUMBER	CALIBRATION DUE DATE
Mitutoyo	0-2”	11121501	11/12/2022

DEVICE TYPE BEING CALIBRATED		SERIAL NO. 121121B72318				
TYPE		RANGE			MOUNTING	
	Stroke	X	+/-18.00	mm	X	Installed
	Diametrial Extensometer			in.		Free Standing
X	Lengthwise Extensometer			kN		In Calibration Fixture
	Load Cell			N		
				kg		
				lbs.		

CALIBRATION IN INCHES 0.1 ____ OR MILLIMETERS X ____

REFERENCE DISTANCE	AS FOUND	OOT	AFTER CALIBRATION
-18.00	-17.77	x	-18.06
-15.00	-14.81	x	-15.01
-12.00	-11.94		-11.94
-9.00	-8.89		-8.89
-6.00	-5.92		-5.92
-3.00	-2.97		-2.97
0.00	0.00		0.00
3.00	2.98		2.98
6.00	5.98		5.98
9.00	8.92		8.92
12.00	11.88		11.88
15.00	14.84		14.84
18.00	17.80	x	18.06

Date of Calibration: 2/28/2022

Calibrated by: G. Palermo