

Standard Operating Procedure (SOP_003)

Title: Equipment Calibration and Maintenance

Version: 1.0

Effective Date: 2025-01-01

Purpose:

To ensure laboratory equipment operates within specified tolerances through regular calibration and preventive maintenance.

Scope:

Applies to all analytical and measuring equipment in the laboratory.

Procedure:

1. Maintain a calibration log for each instrument with calibration date and technician initials.
2. Perform balance calibration weekly using certified weights.
3. Pipettes must be checked monthly and serviced every 6 months.
4. Document maintenance activities including part replacements or repairs.
5. Equipment found out of tolerance must be tagged "Do Not Use" and reported.

Notes:

Regular calibration ensures data accuracy and compliance with quality standards.