

[COMPANY LOGO]

Residual CHO DNA Sample Analysis Report

Assay Run Information

Field	Value
Assay Name	—
SA Run Number	—
Assay Date	—
Request Number	—
Test Method	—
Notebook Number	—
Laboratory	—
Laboratory Address	—

System Suitability

System Suitability: FAIL

Item	R ²	R ² Pass	Efficiency %	Efficiency Pass
Curve Fit	1.00	Pass	100.89	Pass

Item	Ct %CV	Ct %CV Pass	Back-Calc %Bias	Back-Calc Pass
STD 0.003	0.47	Pass	5.06	Pass
STD 0.03	0.11	Pass	-4.04	Pass
STD 0.3	0.29	Pass	-1.75	Pass
STD 3	0.34	Pass	-2.85	Pass
STD 30	0.19	Pass	-1.25	Pass
STD 300	0.46	Pass	3.93	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
ERC - 10 (ERC)	1.12	Pass	68.58	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
ERC - 30 (ERC)	1.91	Pass	66.91	Pass
ERC - 60 (ERC)	3.06	Pass	70.84	Pass
HPC - 190 (HPC)	10.26	Pass	112.88	Pass
LPC - 5 (LPC)	0.70	Pass	109.67	Pass
MPC - 10 (MPC)	5.17	Pass	109.87	Pass

Item	Criteria	Value	Status
NEC - FB (NEC)	Undetermined, or Ct ≥ LOQ Ct	0/3	Fail
NEC - PBS (NEC)	Undetermined, or Ct ≥ LOQ Ct	3/3	Pass
NTC (NTC)	Undetermined, or Ct ≥ LOQ Ct	3/3	Pass

Sample Suitability

Sample	Dilution Factor	Quantity Mean	STD Range Suitability (0 .0028–288.20 29)	Quantity %CV	Replicates Used	Quantity %CV Suitability	Linearity %Bias	Linearity Suitability	Total DNA (ng/mL)	Protein Con centration (mg/mL)	DNA per Protein (ng/mg)	Suitability
S1 D1	1.0	0.0004	Fail	17.24	3	Pass	—	N/A	0.0004	1.825	0.0002	Fail
S1 D2	10.0	0.0	Fail	21.52	3	Pass	—	N/A	0.0004	1.825	0.0002	Fail
S1 D3	100.0	—	N/A	—	1	N/A	—	N/A	0.0007	1.825	0.0004	Fail
S1 D4	1000.0	—	N/A	—	0	N/A	—	N/A	0.0	1.825	0.0	Fail

Final Sample Results

Sample #	Sample ID	Sample Name	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Dilutions Averaged	≤ 15 ng/mg Status
S1			0.0004*	1.825	0.0002*	S1 D1, S1 D2, S1 D3, S1 D4 (n=4)	Below

Values marked with an asterisk (*) were calculated from a sample concentration (Quantity) below the assay's limit of quantification (LOQ = 0.0028 pg/uL), or from replicates with no detectable signal. Treat these results as low confidence.

Signatures

Submitter	Reviewer
Name: —	Name: —
Signature: _____	Signature: _____
Date: _____	Date: _____