

[ COMPANY LOGO ]

## Residual CHO DNA Sample Analysis Report

### Assay Run Information

| Field              | Value           |
|--------------------|-----------------|
| Assay Name         | —               |
| SA Run Number      | sample_data_001 |
| Assay Date         | —               |
| Request Number     | —               |
| Test Method        | —               |
| Notebook Number    | —               |
| Laboratory         | —               |
| Laboratory Address | —               |

### System Suitability

**System Suitability: PASS**

| Item      | R <sup>2</sup> | R <sup>2</sup> Pass | Efficiency % | Efficiency Pass |
|-----------|----------------|---------------------|--------------|-----------------|
| Curve Fit | 1.00           | Pass                | 100.94       | Pass            |

| Item          | Ct %CV | Ct %CV Pass | Back-Calc %Bias | Back-Calc Pass |
|---------------|--------|-------------|-----------------|----------------|
| STD 6 - 0.003 | 0.31   | Pass        | 8.39            | Pass           |
| STD 5 - 0.03  | 0.15   | Pass        | -3.16           | Pass           |
| STD 4 - 0.3   | 0.12   | Pass        | -2.73           | Pass           |
| STD 3 - 3     | 0.25   | Pass        | -11.16          | Pass           |
| STD 2 - 30    | 0.28   | Pass        | -3.13           | Pass           |
| STD 1 - 300   | 0.32   | Pass        | 9.89            | Pass           |

| Item           | Quantity %CV | CV Pass | % Recovery | Recovery Pass |
|----------------|--------------|---------|------------|---------------|
| ERC - 30 (ERC) | 1.50         | Pass    | 79.69      | Pass          |

| Item            | Quantity %CV | CV Pass | % Recovery | Recovery Pass |
|-----------------|--------------|---------|------------|---------------|
| HPC - 190 (HPC) | 5.82         | Pass    | 99.75      | Pass          |
| LPC - 5 (LPC)   | 3.91         | Pass    | 89.61      | Pass          |
| MPC - 10 (MPC)  | 2.28         | Pass    | 93.53      | Pass          |

| Item            | Criteria                          | Value | Status |
|-----------------|-----------------------------------|-------|--------|
| NEC - PBS (NEC) | Undetermined, or Ct $\geq$ LOQ Ct | 3/3   | Pass   |
| NTC (NTC)       | Undetermined, or Ct $\geq$ LOQ Ct | 3/3   | Pass   |

### Sample Suitability

| Sample # | Sample Dilution | Dilution Factor | Quantity %CV | Quantity %CV Suitability | Linearity %Bias | Linearity Suitability | Total DNA (ng/mL) | Protein Concentration (mg/mL) | DNA per Protein (ng/mg) | Suitability |
|----------|-----------------|-----------------|--------------|--------------------------|-----------------|-----------------------|-------------------|-------------------------------|-------------------------|-------------|
| S1       | S1 D1           | 1000.00         | 3.87         | Pass                     | -9.13           | Pass                  | 23049.0494        | 0.9100                        | 25328.6257              | Pass        |
| S1       | S1 D2           | 10000.00        | 3.59         | Pass                     | -1.43           | Pass                  | 25002.2260        | 0.9100                        | 27474.9737              | Pass        |
| S1       | S1 D3           | 20000.00        | 1.69         | Pass                     | 0.00            | Pass                  | 25363.9181        | 0.9100                        | 27872.4375              | Pass        |
| S2       | S2 D1           | 1000.00         | 4.76         | Pass                     | -3.94           | Pass                  | 18574.1577        | 1.2400                        | 14979.1594              | Pass        |
| S2       | S2 D2           | 10000.00        | 3.95         | Pass                     | 0.00            | Pass                  | 19336.7314        | 1.2400                        | 15594.1383              | Pass        |
| S2       | S2 D3           | 20000.00        | 4.24         | Pass                     | -2.58           | Pass                  | 18836.9866        | 1.2400                        | 15191.1182              | Pass        |
| S3       | S3 D1           | 1000.00         | 2.41         | Pass                     | 4.92            | Pass                  | 29976.5790        | 1.4300                        | 20962.6427              | Pass        |
| S3       | S3 D2           | 10000.00        | 1.52         | Pass                     | 0.00            | Pass                  | 28570.2785        | 1.4300                        | 19979.2157              | Pass        |
| S3       | S3 D3           | 20000.00        | 2.78         | Pass                     | -7.75           | Pass                  | 26354.6975        | 1.4300                        | 18429.8584              | Pass        |
| S4       | S4 D1           | 1000.00         | 6.10         | Pass                     | 0.00            | Pass                  | 2.7433            | 0.7000                        | 3.9190                  | Pass        |
| S4       | S4 D2           | 10000.00        | 15.04        | Pass                     | -24.75          | Fail                  | 2.0642            | 0.7000                        | 2.9489                  | Fail        |
| S4       | S4 D3           | 20000.00        | 33.61        | Fail                     | 9.68            | Pass                  | 3.0088            | 0.7000                        | 4.2982                  | Fail        |
| S5       | S5 D1           | 1000.00         | 3.39         | Pass                     | 2.80            | Pass                  | 17102.3960        | 0.8200                        | 20856.5805              | Pass        |
| S5       | S5 D2           | 10000.00        | 0.99         | Pass                     | 0.00            | Pass                  | 16636.0621        | 0.8200                        | 20287.8806              | Pass        |
| S5       | S5 D3           | 20000.00        | 2.62         | Pass                     | -11.00          | Pass                  | 14805.4977        | 0.8200                        | 18055.4850              | Pass        |
| S6       | S6 D1           | 1000.00         | 5.15         | Pass                     | 12.30           | Pass                  | 40960.0932        | 1.2200                        | 33573.8469              | Pass        |
| S6       | S6 D2           | 10000.00        | 3.25         | Pass                     | 0.00            | Pass                  | 36473.8814        | 1.2200                        | 29896.6241              | Pass        |
| S6       | S6 D3           | 20000.00        | 3.45         | Pass                     | -1.01           | Pass                  | 36107.2381        | 1.2200                        | 29596.0968              | Pass        |

### Final Sample Results

| Sample # | Sample ID | Sample Name | Total DNA (ng/mL) | Protein Concentration (mg/mL) | DNA per Protein (ng/mg) | Dilutions Averaged        | $\leq 15$ ng/mg Status |
|----------|-----------|-------------|-------------------|-------------------------------|-------------------------|---------------------------|------------------------|
| S1       |           |             | 24471.7312        | 0.9100                        | 26892.0123              | S1 D1, S1 D2, S1 D3 (n=3) | Above                  |

| Sample # | Sample ID | Sample Name | Total DNA (ng/mL) | Protein Concentration (mg/mL) | DNA per Protein (ng/mg) | Dilutions Averaged        | ≤ 15 ng/mg Status |
|----------|-----------|-------------|-------------------|-------------------------------|-------------------------|---------------------------|-------------------|
| S2       |           |             | 18915.9586        | 1.2400                        | 15254.8053              | S2 D1, S2 D2, S2 D3 (n=3) | Above             |
| S3       |           |             | 28300.5183        | 1.4300                        | 19790.5723              | S3 D1, S3 D2, S3 D3 (n=3) | Above             |
| S4       |           |             | 2.7433            | 0.7000                        | 3.9190                  | S4 D1 (n=1)               | Below             |
| S5       |           |             | 16181.3186        | 0.8200                        | 19733.3154              | S5 D1, S5 D2, S5 D3 (n=3) | Above             |
| S6       |           |             | 37847.0709        | 1.2200                        | 31022.1893              | S6 D1, S6 D2, S6 D3 (n=3) | Above             |

Signatures

| Submitter        |  | Reviewer         |  |
|------------------|--|------------------|--|
| Name: —          |  | Name: —          |  |
| Signature: _____ |  | Signature: _____ |  |
| Date: _____      |  | Date: _____      |  |