

[COMPANY LOGO]

Residual CHO DNA Sample Analysis Report

Assay Run Information

Field	Value
Assay Name	—
SA Run Number	sample_data_002
Assay Date	—
Request Number	—
Test Method	—
Notebook Number	—
Laboratory	—
Laboratory Address	—

System Suitability

System Suitability: PASS

Item	R ²	R ² Pass	Efficiency %	Efficiency Pass
Curve Fit	1.00	Pass	101.87	Pass

Item	Ct %CV	Ct %CV Pass	Back-Calc %Bias	Back-Calc Pass
STD 6 - 0.003	0.24	Pass	2.74	Pass
STD 5 - 0.03	0.29	Pass	0.90	Pass
STD 4 - 0.3	0.22	Pass	-2.18	Pass
STD 3 - 3	0.10	Pass	-4.76	Pass
STD 2 - 30	0.26	Pass	-2.62	Pass
STD 1 - 300	0.38	Pass	5.28	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
ERC - 30 (ERC)	2.91	Pass	79.52	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
HPC - 190 (HPC)	0.77	Pass	98.95	Pass
LPC - 5 (LPC)	2.44	Pass	92.07	Pass
MPC - 10 (MPC)	3.95	Pass	95.37	Pass

Item	Criteria	Value	Status
NEC - PBS (NEC)	Undetermined, or Ct ≥ LOQ Ct	3/3	Pass
NTC (NTC)	Undetermined, or Ct ≥ LOQ Ct	3/3	Pass

Sample Suitability

Sample #	Sample Dilution	Dilution Factor	Quantity %CV	Quantity %CV Suitability	Linearity %Bias	Linearity Suitability	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Suitability
S1	S1 D1	1.00	2.26	Pass	0.00	Pass	8.5834	8.1240	1.0565	Pass
S1	S1 D2	100.00	4.41	Pass	4.40	Pass	8.9614	8.1240	1.1031	Pass
S1	S1 D3	200.00	7.19	Pass	2.02	Pass	8.7566	8.1240	1.0779	Pass
S2	S2 D1	1.00	4.94	Pass	0.00	Pass	0.0024	13.4400	0.0002	Pass
S2	S2 D2	100.00	32.12	Fail	33.33	Fail	0.0032	13.4400	0.0002	Fail
S2	S2 D3	200.00	47.71	Fail	16.67	Pass	0.0028	13.4400	0.0002	Fail
S3	S3 D1	1.00	4.86	Pass	0.00	Pass	0.0102	14.0000	0.0007	Pass
S3	S3 D2	100.00	21.64	Pass	-24.51	Fail	0.0077	14.0000	0.0005	Fail
S3	S3 D3	200.00	28.74	Fail	-1.96	Pass	0.0100	14.0000	0.0007	Fail
S4	S4 D1	1.00	1.07	Pass	0.00	Pass	6.4572	5.4040	1.1949	Pass
S4	S4 D2	100.00	3.64	Pass	4.03	Pass	6.7172	5.4040	1.2430	Pass
S4	S4 D3	200.00	2.76	Pass	-3.27	Pass	6.2462	5.4040	1.1559	Pass
S5	S5 D1	1.00	2.65	Pass	0.00	Pass	0.0122	11.8500	0.0010	Pass
S5	S5 D2	100.00	7.63	Pass	-31.97	Fail	0.0083	11.8500	0.0007	Fail
S5	S5 D3	200.00	27.74	Fail	5.74	Pass	0.0129	11.8500	0.0011	Fail
S6	S6 D1	1.00	5.67	Pass	0.00	Pass	0.0023	14.0000	0.0002	Pass
S6	S6 D2	100.00	32.38	Fail	17.39	Pass	0.0027	14.0000	0.0002	Fail
S6	S6 D3	200.00	—	N/A	—	N/A	0.0019	14.0000	0.0001	N/A

Final Sample Results

Sample #	Sample ID	Sample Name	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Dilutions Averaged	≤ 15 ng/mg Status
S1			8.7671	8.1240	1.0792	S1 D1, S1 D2, S1 D3 (n=3)	Below

Sample #	Sample ID	Sample Name	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Dilutions Averaged	≤ 15 ng/mg Status
S2			0.0024	13.4400	0.0002	S2 D1 (n=1)	Below
S3			0.0102	14.0000	0.0007	S3 D1 (n=1)	Below
S4			6.4735	5.4040	1.1979	S4 D1, S4 D2, S4 D3 (n=3)	Below
S5			0.0122	11.8500	0.0010	S5 D1 (n=1)	Below
S6			0.0023	14.0000	0.0002	S6 D1 (n=1)	Below

Signatures

Submitter		Reviewer	
Name: —		Name: —	
Signature: _____		Signature: _____	
Date: _____		Date: _____	