

[COMPANY LOGO]

Residual CHO DNA Sample Analysis Report

Assay Run Information

Field	Value
Assay Name	â€”
SA Run Number	â€”
Assay Date	â€”
Request Number	â€”
Test Method	â€”
Notebook Number	â€”
Laboratory	â€”
Laboratory Address	â€”

System Suitability

System Suitability: FAIL

Item	RÂ²	RÂ² Pass	Efficiency %	Efficiency Pass
Curve Fit	1.00	Pass	101.87	Pass

Item	Ct %CV	Ct %CV Pass	Back-Calc %Bias	Back-Calc Pass
STD 6 - 0.003	0.24	Pass	2.74	Pass
STD 5 - 0.03	0.29	Pass	0.90	Pass
STD 4 - 0.3	0.22	Pass	-2.18	Pass
STD 3 - 3	0.10	Pass	-4.76	Pass
STD 2 - 30	0.26	Pass	-2.62	Pass
STD 1 - 300	0.38	Pass	5.28	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
ERC - 30 (ERC)	2.91	Pass	79.52	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
HPC - 190 (HPC)	0.77	Pass	98.95	Pass
LPC - 5 (LPC)	2.44	Pass	92.07	Pass
MPC - 10 (MPC)	3.95	Pass	95.37	Pass

Item	Criteria	Mean Ct	Status
NEC - PBS (NEC)	Undetermined, or Ct ≥ 29.8884	29.7784	Fail
NTC (NTC)	Undetermined, or Ct ≥ 29.8884	37.6273	Pass

Sample Suitability

Sample	Dilution Factor	Quantity Mean	STD Range Suitability (0.0029~284.1697)	Quantity %CV	Replicates Used	Quantity %CV Suitability	Linearity %Bias	Linearity Suitability	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Suitability
S1 D1	1.0	8.5834	Pass	2.26	3	Pass	0.00	Pass	8.5834	8.124	1.0565	Pass
S1 D2	100.0	0.0896	Pass	4.41	3	Pass	4.40	Pass	8.9614	8.124	1.1031	Pass
S1 D3	200.0	0.0438	Pass	7.19	3	Pass	2.02	Pass	8.7566	8.124	1.0779	Pass
S2 D1	1.0	0.0024	Fail	4.94	3	Pass	—	N/A	0.0024	13.44	0.0002	Fail
S2 D2	100.0	0.0	Fail	32.12	3	Fail	—	N/A	0.0032	13.44	0.0002	Fail
S2 D3	200.0	0.0	Fail	47.71	3	Fail	—	N/A	0.0028	13.44	0.0002	Fail
S3 D1	1.0	0.0102	Pass	4.86	3	Pass	0.00	Pass	0.0102	14.0	0.0007	Pass
S3 D2	100.0	0.0001	Fail	21.64	3	Pass	-24.51	Fail	0.0077	14.0	0.0005	Fail
S3 D3	200.0	0.0001	Fail	0.0	2	Pass	96.08	Fail	0.02	14.0	0.0014	Fail
S4 D1	1.0	6.4572	Pass	1.07	3	Pass	0.00	Pass	6.4572	5.404	1.1949	Pass
S4 D2	100.0	0.0672	Pass	3.64	3	Pass	4.03	Pass	6.7172	5.404	1.243	Pass
S4 D3	200.0	0.0312	Pass	2.76	3	Pass	-3.27	Pass	6.2462	5.404	1.1559	Pass
S5 D1	1.0	0.0122	Pass	2.65	3	Pass	0.00	Pass	0.0122	11.85	0.001	Pass
S5 D2	100.0	0.0001	Fail	7.63	3	Pass	-31.97	Fail	0.0083	11.85	0.0007	Fail
S5 D3	200.0	0.0001	Fail	0.0	2	Pass	63.93	Fail	0.02	11.85	0.0017	Fail
S6 D1	1.0	0.0023	Fail	5.67	3	Pass	—	N/A	0.0023	14.0	0.0002	Fail
S6 D2	100.0	0.0	Fail	32.38	3	Fail	—	N/A	0.0027	14.0	0.0002	Fail
S6 D3	200.0	—	N/A	—	2	N/A	—	N/A	0.0019	14.0	0.0001	Fail

Final Sample Results

Sample #	Sample ID	Sample Name	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Dilutions Averaged	≤ 15 ng/mg Status	Next Step
S1			8.7671	8.124	1.0792	S1 D1, S1 D2, S1 D3 (n=3)	Below	Reportable
S2			0.0028 - LOQ	13.44	0.0002 - LOQ	S2 D1, S2 D2, S2 D3 (n=3)	Below	LOQ - Spike Test
S3			0.0102	14.0	0.0007	S3 D1 (n=1)	Below	Reportable
S4			6.4735	5.404	1.1979	S4 D1, S4 D2, S4 D3 (n=3)	Below	Reportable
S5			0.0122	11.85	0.0010	S5 D1 (n=1)	Below	Reportable
S6			0.0023 - LOQ	14.0	0.0002 - LOQ	S6 D1, S6 D2, S6 D3 (n=3)	Below	LOQ - Spike Test

Values marked "- LOQ" were calculated from a sample concentration (Quantity) below the assay's limit of quantification (LOQ = 0.0029 pg/uL), or from replicates with no detectable signal. Treat these results as low confidence.

Signatures

Submitter		Reviewer	
Name: "		Name: "	
Signature: _____		Signature: _____	
Date: _____		Date: _____	