

[COMPANY LOGO]

Residual CHO DNA Sample Analysis Report

Assay Run Information

Field	Value
Assay Name	â€”
SA Run Number	â€”
Assay Date	â€”
Request Number	â€”
Test Method	â€”
Notebook Number	â€”
Laboratory	â€”
Laboratory Address	â€”

System Suitability

System Suitability: PASS

Item	RÂ²	RÂ² Pass	Efficiency %	Efficiency Pass
Curve Fit	1.00	Pass	100.67	Pass

Item	Ct %CV	Ct %CV Pass	Back-Calc %Bias	Back-Calc Pass
STD 6 - 0.003	0.36	Pass	2.96	Pass
STD 5 - 0.03	0.09	Pass	-0.31	Pass
STD 4 - 0.3	0.19	Pass	-0.94	Pass
STD 3 - 3	0.19	Pass	-4.42	Pass
STD 2 - 30	0.20	Pass	-3.48	Pass
STD 1 - 300	0.23	Pass	5.56	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
ERC - 30 (ERC)	1.59	Pass	96.80	Pass

Item	Quantity %CV	CV Pass	% Recovery	Recovery Pass
HPC - 190 (HPC)	5.05	Pass	108.45	Pass
LPC - 5 (LPC)	2.27	Pass	98.77	Pass
MPC - 10 (MPC)	2.54	Pass	101.87	Pass

Item	Criteria	Mean Ct	Status
NEC - PBS (NEC)	Undetermined, or Ct ≥ 30.3026	33.9551	Pass
NTC (NTC)	Undetermined, or Ct ≥ 30.3026	39.8641	Pass

Sample Suitability

Sample	Dilution Factor	Quantity Mean	STD Range Suitability (0.0029~283.3238)	Quantity %CV	Replicates Used	Quantity %CV Suitability	Linearity %Bias	Linearity Suitability	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Suitability
S1 D1	1.0	0.0014	Fail	6.91	3	Pass	—	N/A	0.0014	11.8	0.0001	Fail
S1 D2	10.0	0.0001	Fail	0.0	2	Pass	—	N/A	0.001	11.8	0.0001	Fail
S2 D1	1.0	0.0075	Pass	4.72	3	Pass	0.00	Pass	0.0075	9.59	0.0008	Pass
S2 D2	10.0	0.0007	Fail	8.76	3	Pass	-2.67	Pass	0.0073	9.59	0.0008	Fail
S3 D1	1.0	0.0026	Fail	4.7	3	Pass	—	N/A	0.0026	10.13	0.0003	Fail
S3 D2	10.0	0.0003	Fail	6.75	3	Pass	—	N/A	0.0028	10.13	0.0003	Fail
S4 D1	1.0	0.0031	Pass	1.16	3	Pass	0.00	Pass	0.0031	4.487	0.0007	Pass
S4 D2	10.0	0.0003	Fail	5.17	3	Pass	0.00	Pass	0.0031	4.487	0.0007	Fail
S5 D1	1.0	0.0072	Pass	4.35	3	Pass	0.00	Pass	0.0072	9.02	0.0008	Pass
S5 D2	10.0	0.0007	Fail	10.51	3	Pass	-1.39	Pass	0.0071	9.02	0.0008	Fail
S6 D1	1.0	0.0337	Pass	1.91	3	Pass	0.00	Pass	0.0337	10.13	0.0033	Pass
S6 D2	10.0	0.0037	Pass	3.97	3	Pass	10.09	Pass	0.0371	10.13	0.0037	Pass
S7 D1	1.0	0.0063	Pass	3.42	3	Pass	0.00	Pass	0.0063	11.327	0.0006	Pass
S7 D2	10.0	0.0005	Fail	10.47	3	Pass	-15.87	Pass	0.0053	11.327	0.0005	Fail
S8 D1	1000.0	48.0312	Pass	6.7	3	Pass	23.93	Fail	48031.2258	1.63	29467.0097	Fail
S8 D2	10000.0	4.4637	Pass	5.84	3	Pass	15.17	Pass	44636.9394	1.63	27384.6254	Pass
S8 D3	20000.0	1.9379	Pass	2.82	3	Pass	0.00	Pass	38757.2479	1.63	23777.4527	Pass

Final Sample Results

Sample #	Sample ID	Sample Name	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Dilutions Averaged	≤ 15 ng/mg Status	Next Step
S1			0.0012 - LOQ	11.8	0.0001 - LOQ	S1 D1, S1 D2 (n=2)	Below	LOQ - Spike Test

Sample #	Sample ID	Sample Name	Total DNA (ng/mL)	Protein Concentration (mg/mL)	DNA per Protein (ng/mg)	Dilutions Averaged	≤ 15 ng/mg Status	Next Step
S2			0.0075	9.59	0.0008	S2 D1 (n=1)	Below	Reportable
S3			0.0027 - LOQ	10.13	0.0003 - LOQ	S3 D1, S3 D2 (n=2)	Below	LOQ - Spike Test
S4			0.0031	4.487	0.0007	S4 D1 (n=1)	Below	Reportable
S5			0.0072	9.02	0.0008	S5 D1 (n=1)	Below	Reportable
S6			0.0354	10.13	0.0035	S6 D1, S6 D2 (n=2)	Below	Reportable
S7			0.0063	11.327	0.0006	S7 D1 (n=1)	Below	Reportable
S8			41697.0937	1.63	25581.0390	S8 D2, S8 D3 (n=2)	Above	Reportable

Values marked "- LOQ" were calculated from a sample concentration (Quantity) below the assay's limit of quantification (LOQ = 0.0029 pg/uL), or from replicates with no detectable signal. Treat these results as low confidence.

Signatures

Submitter	Reviewer
Name:	Name:
Signature: _____	Signature: _____
Date: _____	Date: _____