

LaunchFSW

Execution time
11/20/2023 3:04:23 PM
Operating system
Windows 10 64bit
Run configuration name
TestRun

Computer name
IRAYSERVER
Screen dimension
1600x900
Username
I-Ray

Language
en-US
Duration
1.3m



■ 2 x Failed
■ 1 x Blocked

Total warnings
0

Total errors
2

Table of Contents

■ SmartFolder.....	2
■ SmartFolder.....	2
■ TestCase.....	2
■ KillOpenApplication.....	2
■ LaunchFSW.....	2
■ TestCase.....	3
■ Simulate.....	3
■ TestCase.....	3



SmartFolder

Rows: 1

1.3m



SmartFolder

Data Row: 1

1.3m



First Name: Test

Last Name: Patient

Device Name: LT977-DW

TestCase



TestCase

1.2m



KillOpenApplication

38ms



Time	Level	Category	Message
------	-------	----------	---------



LaunchFSW

1.2m



Time	Level	Category	Message
------	-------	----------	---------

00:00.120 Info Data Current variable values:

\$First_Name = 'Test', \$Last_Name = 'Patient'

00:00.160 Info Application Run application 'C:\Program Files (x86)\ReSound\SmartFit\SmartFitSA.exe' in normal mode.

00:00.589 Info Delay Waiting for 12s.

00:12.624 Info Mouse Mouse Left Click item 'Directionality.StandAloneAutomationIdsAdd2Action' at Center.


01:12.756 Info User Item 'StandAloneAutomationIdsAdd2Action' was found using Robust path.


01:12.906 Error Module Failed to find item 'LaunchFSWRepository.Directionality.StandAloneAutomationIdsAdd2Action'.





No element found for path '/form[@name='ReSound Smart Fit 1.17']/button[@automationid='StandAloneAutomationIds.Add2Action']' at step 'button[@automationid='StandAloneAutomationIds.Add2Action']' after 352 attempts within 1m.

TestCase


 **TestCase**

7.62s 


 **Simulate**

7.61s 

Time	Level	Category	Message
01:12.982	Info	Data	Current variable values: \$Device_Name = 'LT977-DW'
01:13.003	Info	Mouse	Mouse Left Click item 'SmartLauncher.PARTEditor' at Center.
01:20.097	Error	Module	The test run has been aborted by the user.



TestCase

 **TestCase**

0ms 