

Electrophoresis (Information Catalogue)

PURPOSE

CBQAP aims to reduce the high variability observed in routine diagnostic testing. The Electrophoresis module, part of the CBQAP proficiency testing, will help reduce the variability observed and ensure that the results produced showcase accurate description of the participant testing.

DISCLAIMER

CBQAP shall strictly maintain participant confidentiality. Any participant detail shall not be revealed to other participants. The data gathered from participants will only be used for analysis and research purposes.

Participants must test the sample(s) as per their routine diagnostic testing. The tests should be performed on the same method and analyser as mentioned in the CBQAP enrolment. The protocol for the same is provided in this document. Participants are advised not to share their webpage credentials, reports and other sensitive information.

SAMPLE CONTENT

The contents provided in this module comprise two amber-colored glass vials containing lyophilised serum.

The samples have been screened and found to be negative for tests such as Hepatitis B, Hepatitis C, and HIV. The bottles are labelled with the respective program numbers for identification. For the Electrophoresis program, the bottle will be labelled as “**Electrophoresis Module**” and will comprise of program code starting with “**EP**”. This module is separated into **QL** and **QN**, which are packaged as separate sample bottles (further information in the protocol).

STORAGE

Samples are to be stored at a temperature of 2-8°C. The samples can withstand a temperature of -20°C, and can remain stable when kept at 2-8°C till the given expiration date. Once opened and reconstituted the samples remains stable in liquid form for 5 days at 2-8°C, provided repeated freeze thawing is refrained from.

Please ensure the sample is received in a good condition. If any breakage in sample or tampering in consignment packaging is observed, kindly report it to CBQAP.

Electrophoresis (Information Catalogue)

PROTOCOL

In the Electrophoresis Program, two samples will be provided: **QL and QN**. Both samples cover different set of tests to be performed. Use the **QL** and **QN** samples as follows:

QL: SPE and Immunotyping

QN: Immunoglobulin Quantification, Free Light Chain Quantification, Total Protein Quantification

Proceed with proficiency testing with these samples accordingly, with the tests enrolled in CBQAP.

Diluent To Be Used: Normal Saline

Protocol

1. The sample provided by CBQAP should be treated as a patient sample and all protocols for testing the sample(s) should be carried out according to routine diagnostic procedures.
2. Take the vial labelled “**Electrophoresis Module**”. Open the metallic seal and then carefully open the rubber cap.
3. Reconstitute with exactly **1 mL** of diluent and keep for 30 minutes at room temperature.
4. Test for the markers enrolled in the CBQAP program in the analysers and method registered.
5. After testing, discard the sample and the vials as per the routine diagnostic procedures.

A total of four samples to be assayed over a twelve-month period, with one sample tested per quarter. Each sample is labelled accordingly, by which results should be submitted to CBQAP.

The tests and parameters to be performed (based on the enrolment) is provided with the consignment.

Electrophoresis (Information Catalogue)

LIMITATIONS

1. This product should not be used past expiration date.
2. If there is evidence of microbial contamination or excessive turbidity in the product, discard the vial and report the same.
3. This product is not intended for use as a standard.
4. Serum components may interfere with assay signal.

RESULT REPORTING & DATA SUBMISSION REQUIREMENTS

1. All test results shall be submitted through the CBQAP online portal within the specified timeline. The portal will be made available shortly.
2. Results that are not reported will be excluded from statistical analysis and considered unacceptable.
3. Samples yielding results beyond the analytical measurement range of the method shall be managed in accordance with established laboratory procedures for handling out-of-range values.
4. Only results submitted on or before the assigned due date shall be included in the evaluation.
5. The last date of submitting the results is mentioned in the enrolment information card, dispatched along with the sample. Late submissions may be considered; however, they shall be flagged in the revised or updated analytical reports.
6. Results submitted after the closure of the evaluation cycle shall not be accepted under any circumstances.
7. Measurement uncertainty and coverage factor will be available in the sample reports, denoting any deviation from homogeneity, stability and repeatability of the sample for a given parameter.

PRECAUTIONS



Level 2 Biohazard. Treat as potentially infectious.

Any unexpected results, abnormal sample characteristics, or suspected reagent issues should be promptly reported and investigated before proceeding further.

Electrophoresis (Information Catalogue)

Contact Details

Address: Department of Biochemistry, Tata Memorial Hospital, Jerbai Wadia Lane, Parel East, Mumbai, Maharashtra 400012.

Phone No: 24177000 | Extension: 4339

Email: ncgcbqap@gmail.com | Eqas-biochem.tmh@tmc.gov.in