Validated IC method

Nitrate in water

Reference SD: 0.10 mg/L

Treated as known here

→

Today's repeatability check

One homogeneous sample



Six independent results

Estimate today's spread

→

Analytical question

Has precision worsened?

Or could larger spread occur by chance?

Use a sampling reference

A column problem is suspected, not established by the test.

We test repeatability, not whether the nitrate mean is correct.