SystemTP-603 SystemTP

Portview

SystemTP

Report no. : SystemTP-603

Rev. : 1.6

By	Title	Name	Sign
Prepared	Manager	S. R. Lim	
	Staff	J. B. Kim	
Reviewed	General Manager	S. I. Choi	
Approved	CTO (Director)	K. Y. Ro	

GENORAY Co., Ltd.

Revision History

Rev#	Description	Date	Remark
1.5	Controlled system-level procedure baseline maintained with the Portview verification plan set	2024.01.29	
1.6	Related verification-document revisions updated for the current Portview baseline	2025.07.24	

Table of Contents

1. Purpose 4
2. References 4
3. Execution Basis 4
4. System-Level Coverage Model 4
    4.1 Scope Boundaries 5
    4.2 What Stays At System Level 5
    4.3 What Does Not Belong Here 5
5. System Procedure Set 5
    5.1 Approved System Procedures 5
    5.2 Requirement Mapping 6
    5.3 Procedure Intent 7
6. Procedure Summary 7
7. Planned Evidence Output 8
8. Traceability Use 8

Document Overview

This document defines the system-level verification procedures for Portview.

1. Purpose

This document defines the system-level verification procedures used to demonstrate that the integrated Portview product baseline satisfies the applicable system-level expectations.

2. References

This system procedure is governed primarily by:

- `SV-603-04` Software Verification Plan
- `RS-603` RS for Portview
- `TM-603` Traceability Matrix

3. Execution Basis

Field	Value
Product under test	`Portview`
Software baseline	`Portview 2.2.5.16`
Test platform	`Genoray testing platform`
Configuration	`Default`
Verification level	`System`

4. System-Level Coverage Model

The system-level procedure baseline covers verification of the integrated Portview product configuration rather than isolated software functions.

4.1 Scope Boundaries

Scope Area	Coverage Intent
Product-baseline verification	Confirm the integrated Portview baseline behaves as intended on the controlled platform
System-level functional support	Confirm the supporting system-level functions are properly available
End-to-end behavior	Confirm the complete user-facing workflow remains acceptable at integrated system level

4.2 What Stays At System Level

System-level verification should stay focused on behaviors that need the released baseline, configured environment, and full user-facing workflow.

- startup, launch, and baseline availability on the designated workstation configuration
- end-to-end patient and image workflow across multiple integrated functions
- integrated viewer behavior under representative clinical workflow sequences
- output and communication behavior in the complete product configuration
- user-visible recovery, logging, and localized interaction at product level

4.3 What Does Not Belong Here

The following should remain in unit or integration procedure sets and should not be repeated here as isolated low-level checks.

- single-tool annotation behavior already covered by `TP-603`
- isolated import or export edge cases already covered by `TP-603`
- unit-level module behavior already covered by `STP-603`
- design-allocation confirmation that belongs to `SDS-603` or `TM-603`

5. System Procedure Set

The authored system procedure set should use a small number of release-oriented procedure items instead of reproducing every lower-level check.

5.1 Approved System Procedures

System Procedure ID	Title	System-Level Intent
`SYSP-001`	Product baseline startup and availability	Confirm the released Portview baseline launches and presents the expected operational context on the controlled platform
`SYSP-002`	End-to-end patient and acquisition workflow	Confirm patient selection, source selection, acquisition, and image presentation operate as one product workflow
`SYSP-003`	End-to-end viewer and diagnostic workflow	Confirm mounted-view, compare-view, tooth-map, navigation, and representative annotation workflow remain coherent at product level
`SYSP-004`	End-to-end output and communication workflow	Confirm print, export, media, and DICOM paths operate correctly in the full integrated baseline
`SYSP-005`	Device integrity and recovery workflow	Confirm device status, sensor connectivity, USB integrity, network dependence, and recovery guidance are acceptable at product level
`SYSP-006`	Logging and localization workflow	Confirm activity logging and multilingual GUI behavior remain acceptable in the released product workflow
`SYSP-007`	Installation and authorized-version acceptance	Confirm the released product baseline is installed on an acceptable workstation configuration and unsupported or unauthorized installation states are detected
`SYSP-008`	Operational audit and log-review workflow	Confirm representative product operations produce audit-relevant logs that support troubleshooting and review

System Procedure ID	Title	System-Level Intent
`SYSP-009`	Data persistence and reopen workflow	Confirm patient and image data remain attributable, recoverable, and reusable after save, close, and reopen workflow transitions

5.2 Requirement Mapping

System Procedure ID	Primary RS Coverage
`SYSP-001`	`RS-001`, `RS-002`
`SYSP-002`	`RS-003` to `RS-009`, `RS-024`
`SYSP-003`	`RS-010` to `RS-014`
`SYSP-004`	`RS-015` to `RS-018`
`SYSP-005`	`RS-019`, `RS-020`, `RS-022`, `RS-023`
`SYSP-006`	`RS-021`, `RS-025`
`SYSP-007`	`RS-001`, `RS-019`, `RS-020`, `RS-023`
`SYSP-008`	`RS-021`, `RS-023`
`SYSP-009`	`RS-003`, `RS-010`, `RS-016`, `RS-024`

5.3 Procedure Intent

System Procedure ID	Setup	Core Procedure	Acceptance Criteria
`SYSP-001`	Released Portview baseline on controlled workstation configuration	Launch product, confirm baseline availability, review startup and operational context	Product launches correctly and the expected operational context is available without blocking issue

System Procedure ID	Setup	Core Procedure	Acceptance Criteria
`SYSP-002`	Controlled patient dataset, configured device source, released baseline	Select patient, acquire image from representative source, and confirm image presentation in the integrated workflow	End-to-end acquisition workflow completes and the resulting image is associated with the correct patient context
`SYSP-003`	Released baseline with representative patient image set	Perform mounted-view, compare-view, tooth-map, and representative annotation workflow	Viewer workflow remains coherent and representative diagnostic actions are usable at product level
`SYSP-004`	Released baseline with output destinations configured	Execute representative print, export, media, and DICOM-send actions	Output and communication paths complete successfully or fail with controlled user-visible handling
`SYSP-005`	Released baseline with representative device and connection state transitions	Review connectivity, integrity, and recovery behavior under representative operational changes	Device-related issues are detected, communicated, and recoverable without uncontrolled behavior
`SYSP-006`	Released baseline with enabled logging and language resources	Execute representative workflow while reviewing logging and localized user interaction	Product logs the expected workflow activity and localized GUI behavior remains coherent
`SYSP-007`	Released installation package and supported workstation configuration	Install or verify the released baseline, confirm displayed version, and review unsupported or unauthorized installation conditions	The released version is identifiable, acceptable workstation conditions are met, and unsupported or unauthorized installation states are detected with controlled handling

System Procedure ID	Setup	Core Procedure	Acceptance Criteria
`SYSP-008`	Released baseline with representative workflow activity	Execute representative clinical workflow, then review the produced operational logs for completeness and troubleshooting value	Audit-relevant workflow actions are reflected in the log set with sufficient detail to support review and problem analysis
`SYSP-009`	Released baseline with representative patient and image dataset	Create or open patient data, acquire or load images, save results, close the workflow, and reopen the data set	Patient and image data remain associated, recoverable, and reusable after the full save-close-reopen sequence

6. Procedure Summary

The system-level verification procedure set should:

- execute the controlled integrated Portview configuration
- confirm the complete product baseline on the designated test platform
- verify representative end-to-end workflows rather than isolated low-level checks
- confirm installation, auditability, and persistence expectations that are only meaningful at released product level
- record pass, fail, or justified outcomes for the integrated system-level behavior



7. Planned Evidence Output

Procedure	Expected Output
`SYSP-001` to `SYSP-009`	`SystemTR-603` system-level result record
`SystemTP-603`	Controlled system procedure baseline governing `SYSP-001` to `SYSP-009`

8. Traceability Use

This procedure document is intended to support:

- system-level planning references in `SV-603-04` Software Verification Plan
- system-level execution summaries in `SV-603-05` Software Verification Report
- system-level traceability rows in `TM-603` Traceability Matrix