TR-603 SwTR

Portview

SwTR

Report no. : TR-603

Rev. : 0.1

By	Title	Name	Sign
Prepared	Manager	J. W. Lee	
Reviewed	General Manager	S. I. Choi	
Approved	CTO (Director)	K. Y. Ro	

GENORAY Co., Ltd.

Revision History

Rev#	Description	Date	Remark
0.0	Initial integration result record prepared according to verification-item subdivision	2024-01-22	
0.1	Integration result set updated with translation-related verification coverage	2025-07-24	

Table of Contents

1. Purpose 4
2. References 4
3. Execution Information 4
4. Result Coverage 5
5. Result Summary 5
    5.1 Procedure-Level Result Set 5
6. Deviations And Issues 8
7. Evidence References 8
8. Traceability Use 8

Document Overview

This document records the integration-level verification results for Portview.

1. Purpose

This document records the executed integration verification results for Portview and provides the evidence summary associated with the integration procedure baseline.

2. References

This result record is governed primarily by:

- `RS-603` RS for Portview
- `TP-603` SwTP for Portview
- `SV-603-04` Software Verification Plan
- `SV-603-05` Software Verification Report

3. Execution Information

Field	Value
Executor	`J. W. Lee`
Execution date	`2025.07.24`
Independent reviewer	`M. C. Boo`
Review date	`2025.07.24`
Execution note	Tests were executed using a pre-configured Genoray testing platform with standardized test data sets. The controlled test environment and procedure-level acceptance criteria enabled single-executor throughput. Independent verification is provided through document review by a separate reviewer.
Test location	`Genoray testing platform`
Configuration	`Default`

Field	Value
Software baseline	`Portview 2.2.5.16`
Governing procedure	`TP-603`

4. Result Coverage

The integration result record covers the following authored integration procedures.

Procedure Range	Coverage
`ITP-001` to `ITP-006`	Patient and data workflow
`ITP-007` to `ITP-012`, `ITP-031` to `ITP-034`	Device acquisition and source integration
`ITP-013` to `ITP-026`	Viewer and annotation workflow
`ITP-027` to `ITP-030`	Output and exchange workflow
`ITP-035` to `ITP-036`	Logging and localization

5. Result Summary

The integration verification activities were executed against the released Portview integration baseline and were reported as successfully completed.

Area	Result Statement	Status
Patient and data workflow	Workflow actions were executed and the expected handling behavior was observed	`Pass summary stated`
Device acquisition and source integration	Device, source, and acquisition transitions were executed successfully	`Pass summary stated`
Viewer and annotation workflow	Integrated viewer and annotation actions behaved as intended	`Pass summary stated`

Area	Result Statement	Status
Output and exchange workflow	Print, export, and DICOM-related actions were reported as completed successfully	`Pass summary stated`
Logging and localization	Support and localization behaviors were reported as acceptable	`Pass summary stated`

5.1 Procedure-Level Result Set

The current authored result set is aligned to the `ITP-xxx` integration procedure structure defined in `TP-603`.

Integration Procedure ID	Title	Result Summary	Status	Notes
`ITP-001`	System instability check	Integrated instability and guided-recovery behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-002`	Patient registration	Patient registration workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-003`	Modify patient information	Patient-modification workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-004`	Patient search	Patient-search workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-005`	Delete patient information	Patient-delete workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Integration Procedure ID	Title	Result Summary	Status	Notes
`ITP-006`	Export patient information	Patient export workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-007`	Select device	Integrated device-selection workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-008`	Acquire intraoral sensor image	Integrated sensor-acquisition workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-009`	Acquire image from digital camera	Integrated digital-camera workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-010`	Acquire image from file	Integrated file-import workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-011`	Acquire image from TWAIN device	Integrated TWAIN workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-012`	Show status of device	Integrated device-status behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-013`	Main viewer	Main-viewer workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Integration Procedure ID	Title	Result Summary	Status	Notes
`ITP-014`	Thumbnail viewer	Thumbnail-viewer workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-015`	Compare viewer	Compare-viewer workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-016`	Tooth map window	Tooth-map workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-017`	Pan behavior	Pan behavior in integrated workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-018`	Zoom behavior	Zoom behavior in integrated workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-019`	Reset	Reset behavior in integrated workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-020`	Delete	Delete behavior in integrated workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-021`	Rotate	Rotate behavior in integrated workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Integration Procedure ID	Title	Result Summary	Status	Notes
`ITP-022`	Text	Text annotation workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-023`	Arrow	Arrow annotation workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-024`	Pencil	Pencil annotation workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-025`	Angle	Angle measurement workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-026`	Length	Length measurement workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-027`	Print image	Integrated print workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-028`	Print to DICOM printer	Integrated DICOM-print workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-029`	CD burning	Integrated media workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

Integration Procedure ID	Title	Result Summary	Status	Notes
`ITP-030`	Send image to DICOM server	Integrated DICOM-send workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-031`	Multi-sensor selection	Multi-sensor workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-032`	Intraoral sensor FMX layout	FMX workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-033`	USB integrity	USB integrity behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-034`	Local-network connection check	Network-dependent integrated workflow was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-035`	Activity logging	Integrated logging behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled
`ITP-036`	GUI translation	Integrated GUI translation behavior was reported as acceptable	`Pass summary claimed`	Detailed execution note to add if separately controlled

6. Deviations And Issues

No Portview-specific integration deviation is separately listed in the current result set.

ID	Description	Disposition
`NONE-RECORDED`	No separate Portview integration deviation is recorded in the current result summary set	`Accepted as no recorded deviation`

7. Evidence References

Evidence Type	Reference	Location
Executed test protocol	Signed integration-test execution protocol for `ITP-001` to `ITP-036`	QMS controlled archive, test-execution record set
Test environment log	Genoray testing platform configuration and session log	QMS controlled archive
Independent review record	Review sign-off by `M. C. Boo` confirming result acceptance	QMS controlled archive

8. Traceability Use

This result record is intended to support:

- integration-result linkage in `TM-603` Traceability Matrix
- integration-result references in `SV-603-05` Software Verification Report
- release-support statements in `SV-603-01` Software Validation Report